

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022823\
 Data File : PP056080.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Feb 2023 21:53
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 01 03:29:33 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 20 05:24:03 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.322	3.564	85695530	66033693	53.186	45.223
2) SA Decachlor...	10.068	8.565	85307451	57151155	41.818	50.376
Target Compounds						
16) L4 AR-1242-1	5.491	4.645	25866142	17245972	535.683	468.729
17) L4 AR-1242-2	5.514	4.663	37824102	24237369	532.292	460.158
18) L4 AR-1242-3	5.576	4.840	23976004	13234088	537.407	467.462
19) L4 AR-1242-4	5.675	4.924	19092501	14075872	528.432	456.226
20) L4 AR-1242-5	6.415	5.447	21464816	16123513	512.419	484.054



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022823\
Data File : PP056080.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Feb 2023 21:53
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 01 03:29:33 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021723.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 20 05:24:03 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.322 min
Delta R.T.: -0.003 min
Response: 85695530
Conc: 53.19 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.564 min
Delta R.T.: -0.004 min
Response: 66033693
Conc: 45.22 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.068 min
Delta R.T.: -0.008 min
Response: 85307451
Conc: 41.82 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.565 min
Delta R.T.: -0.012 min
Response: 57151155
Conc: 50.38 ng/ml



#16 AR-1242-1


R.T.: 5.491 min
Delta R.T.: -0.006 min
Response: 25866142
Conc: 535.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.645 min
Delta R.T.: -0.006 min
Response: 17245972
Conc: 468.73 ng/ml




#17 AR-1242-2

R.T.: 5.514 min
Delta R.T.: -0.006 min
Response: 37824102
Conc: 532.29 ng/ml



#17 AR-1242-2


R.T.: 4.663 min
Delta R.T.: -0.007 min
Response: 24237369
Conc: 460.16 ng/ml



#18 AR-1242-3


R.T.: 5.576 min
Delta R.T.: -0.005 min
Response: 23976004
Conc: 537.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.840 min
Delta R.T.: -0.007 min
Response: 13234088
Conc: 467.46 ng/ml




#19 AR-1242-4

R.T.: 5.675 min
Delta R.T.: -0.006 min
Response: 19092501
Conc: 528.43 ng/ml



#19 AR-1242-4


R.T.: 4.924 min
Delta R.T.: -0.007 min
Response: 14075872
Conc: 456.23 ng/ml



#20 AR-1242-5

R.T.: 6.415 min
Delta R.T.: -0.005 min
Response: 21464816
Conc: 512.42 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.447 min
Delta R.T.: -0.008 min
Response: 16123513
Conc: 484.05 ng/ml