

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030122\
 Data File : PP044143.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Mar 2022 21:12
 Operator : YP\AJ
 Sample : N1715-11 10X
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 02 00:29:39 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 11 04:48:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.912	3.170	1906616	2255164	0.862	1.201 #
2)	SA Decachlor...	10.052	8.539	1557019	1947899	1.311	1.201
Target Compounds							
3)	L1 AR-1016-1	5.171	4.327	186312	21411	2.568	0.349 #
4)	L1 AR-1016-2	5.189	4.341	104727	43873	0.993	0.519 #
5)	L1 AR-1016-3	5.262	4.521	9779	43517	0.149	0.932 #
6)	L1 AR-1016-4	5.374	4.574	35688	37838	0.658	1.012 #
7)	L1 AR-1016-5	5.682	4.797	20888	43775	0.398	0.958 #
8)	L2 AR-1221-1	4.135	3.405	136275	63319	5.353	2.655 #
9)	L2 AR-1221-2	4.235	3.497	78456	44797	4.087	2.480 #
10)	L2 AR-1221-3	4.320	3.569	31645	19950	0.551	0.387 #
11)	L3 AR-1232-1	4.320	3.569	31645	19950	0.652	0.464 #
12)	L3 AR-1232-2	4.878	4.341	32648	43873	1.378	1.229
13)	L3 AR-1232-3	5.189	4.521	104727	43517	2.432	2.225
14)	L3 AR-1232-4	5.374	4.616	35688	37784	1.635	2.185 #
15)	L3 AR-1232-5	5.469	4.797	105730	43775	6.555	2.330 #
16)	L4 AR-1242-1	5.171	4.327	186312	21411	3.316	0.439 #
17)	L4 AR-1242-2	5.189	4.341	104727	43873	1.266	0.652 #
18)	L4 AR-1242-3	5.262	4.521	9779	43517	0.189	1.159 #
19)	L4 AR-1242-4	5.374	4.616	35688	37784	0.836	1.046 #
20)	L4 AR-1242-5	6.167	5.170	387773	47710	8.980	1.108 #
21)	L5 AR-1248-1	5.171	4.327	186312	21411	4.544	0.575 #
22)	L5 AR-1248-2	5.469	4.574	105730	37838	1.861	0.773 #
23)	L5 AR-1248-3	5.682	4.616	20888	37784	0.313	0.736 #
24)	L5 AR-1248-4	6.128	4.797	169995	43775	2.433	0.747 #
25)	L5 AR-1248-5	6.167	5.222	387773	29257	5.609	0.511 #
26)	L6 AR-1254-1	6.094	5.170	368510	47710	4.877	0.546 #
27)	L6 AR-1254-2	6.341	5.346	72333	45842	0.645	0.601
28)	L6 AR-1254-3	6.745	5.786	98132	56197	0.848	0.459 #
29)	L6 AR-1254-4	7.027f	6.017	73195	110091	0.917	1.395 #
30)	L6 AR-1254-5	7.527	6.471	196268	78663	2.288	0.680 #
31)	L7 AR-1260-1	6.933	5.910	247829	39380	2.959	0.492 #
32)	L7 AR-1260-2	7.218	6.123	11765	32202	0.130	0.330 #
33)	L7 AR-1260-3	7.608	6.281	69947	7873	0.965	0.086 #
34)	L7 AR-1260-4	7.848	6.798	115138	94463	1.341	1.168
35)	L7 AR-1260-5	8.189	7.066	82296	226746	0.531	1.227 #
36)	L8 AR-1262-1	7.608	6.519	69947	35393	0.708	0.322 #
37)	L8 AR-1262-2	8.189	6.798	82296	94463	0.510	0.926 #
38)	L8 AR-1262-3	8.528	7.374	55896	10506	0.511	0.123 #
39)	L8 AR-1262-4	8.626	7.437	77615	13160	1.618	0.085 #
40)	L8 AR-1262-5	9.320f	7.987	25336	64518	0.455	0.829 #
41)	L9 AR-1268-1	8.528	7.374	55896	10506	0.287	0.041 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030122\
 Data File : PP044143.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Mar 2022 21:12
 Operator : YP\AJ
 Sample : N1715-11 10X
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 02 00:29:39 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 11 04:48:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.626	7.437	77615	13160	0.443	0.058 #
43) L9	AR-1268-3	8.845	7.655	32927	72526	0.213	0.372 #
44) L9	AR-1268-4	9.320f	7.987	25336	64518	0.405	0.717 #
45) L9	AR-1268-5	9.728	8.287	9027	26460	0.020	0.044 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030122\
Data File : PP044143.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Mar 2022 21:12
Operator : YP\AJ
Sample : N1715-11 10X
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 02 00:29:39 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 11 04:48:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.912 min
Delta R.T.: -0.007 min
Response: 1906616
Conc: 0.86 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.170 min
Delta R.T.: -0.005 min
Response: 2255164
Conc: 1.20 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.052 min
Delta R.T.: -0.008 min
Response: 1557019
Conc: 1.31 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.539 min
Delta R.T.: -0.011 min
Response: 1947899
Conc: 1.20 ng/ml



#3 AR-1016-1


R.T.: 5.171 min
Delta R.T.: -0.003 min
Response: 186312
Conc: 2.57 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.327 min
Delta R.T.: 0.004 min
Response: 21411
Conc: 0.35 ng/ml




#4 AR-1016-2

R.T.: 5.189 min
Delta R.T.: -0.008 min
Response: 104727
Conc: 0.99 ng/ml



#4 AR-1016-2


R.T.: 4.341 min
Delta R.T.: 0.000 min
Response: 43873
Conc: 0.52 ng/ml



#5 AR-1016-3


R.T.: 5.262 min
Delta R.T.: 0.000 min
Response: 9779
Conc: 0.15 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.521 min
Delta R.T.: -0.006 min
Response: 43517
Conc: 0.93 ng/ml




#6 AR-1016-4

R.T.: 5.374 min
Delta R.T.: 0.005 min
Response: 35688
Conc: 0.66 ng/ml



#6 AR-1016-4


R.T.: 4.574 min
Delta R.T.: -0.002 min
Response: 37838
Conc: 1.01 ng/ml



#7 AR-1016-5


R.T.: 5.682 min
Delta R.T.: -0.009 min
Response: 20888
Conc: 0.40 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 4.797 min
Delta R.T.: -0.004 min
Response: 43775
Conc: 0.96 ng/ml




#8 AR-1221-1

R.T.: 4.135 min
Delta R.T.: -0.013 min
Response: 136275
Conc: 5.35 ng/ml



#8 AR-1221-1


R.T.: 3.405 min
Delta R.T.: 0.002 min
Response: 63319
Conc: 2.65 ng/ml



#9 AR-1221-2


R.T.: 4.235 min
Delta R.T.: -0.003 min
Response: 78456
Conc: 4.09 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.497 min
Delta R.T.: 0.005 min
Response: 44797
Conc: 2.48 ng/ml




#10 AR-1221-3

R.T.: 4.320 min
Delta R.T.: 0.002 min
Response: 31645
Conc: 0.55 ng/ml



#10 AR-1221-3


R.T.: 3.569 min
Delta R.T.: -0.003 min
Response: 19950
Conc: 0.39 ng/ml



#11 AR-1232-1


R.T.: 4.320 min
Delta R.T.: 0.002 min
Response: 31645
Conc: 0.65 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 3.569 min
Delta R.T.: -0.003 min
Response: 19950
Conc: 0.46 ng/ml




#12 AR-1232-2

R.T.: 4.878 min
Delta R.T.: -0.005 min
Response: 32648
Conc: 1.38 ng/ml



#12 AR-1232-2


R.T.: 4.341 min
Delta R.T.: 0.000 min
Response: 43873
Conc: 1.23 ng/ml



#13 AR-1232-3


R.T.: 5.189 min
Delta R.T.: -0.008 min
Response: 104727
Conc: 2.43 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.521 min
Delta R.T.: -0.006 min
Response: 43517
Conc: 2.23 ng/ml




#14 AR-1232-4

R.T.: 5.374 min
Delta R.T.: 0.005 min
Response: 35688
Conc: 1.63 ng/ml



#14 AR-1232-4


R.T.: 4.616 min
Delta R.T.: -0.003 min
Response: 37784
Conc: 2.19 ng/ml



#15 AR-1232-5


R.T.: 5.469 min
Delta R.T.: -0.003 min
Response: 105730
Conc: 6.56 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 4.797 min
Delta R.T.: -0.004 min
Response: 43775
Conc: 2.33 ng/ml




#16 AR-1242-1

R.T.: 5.171 min
Delta R.T.: -0.003 min
Response: 186312
Conc: 3.32 ng/ml



#16 AR-1242-1


R.T.: 4.327 min
Delta R.T.: 0.004 min
Response: 21411
Conc: 0.44 ng/ml



#17 AR-1242-2


R.T.: 5.189 min
Delta R.T.: -0.008 min
Response: 104727
Conc: 1.27 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.341 min
Delta R.T.: -0.001 min
Response: 43873
Conc: 0.65 ng/ml




#18 AR-1242-3

R.T.: 5.262 min
Delta R.T.: -0.001 min
Response: 9779
Conc: 0.19 ng/ml



#18 AR-1242-3


R.T.: 4.521 min
Delta R.T.: -0.006 min
Response: 43517
Conc: 1.16 ng/ml



#19 AR-1242-4


R.T.: 5.374 min
Delta R.T.: 0.005 min
Response: 35688
Conc: 0.84 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.616 min
Delta R.T.: -0.003 min
Response: 37784
Conc: 1.05 ng/ml




#20 AR-1242-5

R.T.: 6.167 min
Delta R.T.: -0.003 min
Response: 387773
Conc: 8.98 ng/ml



#20 AR-1242-5


R.T.: 5.170 min
Delta R.T.: -0.009 min
Response: 47710
Conc: 1.11 ng/ml



#21 AR-1248-1


R.T.: 5.171 min
Delta R.T.: -0.003 min
Response: 186312
Conc: 4.54 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.327 min
Delta R.T.: 0.005 min
Response: 21411
Conc: 0.57 ng/ml




#22 AR-1248-2

R.T.: 5.469 min
Delta R.T.: -0.003 min
Response: 105730
Conc: 1.86 ng/ml



#22 AR-1248-2


R.T.: 4.574 min
Delta R.T.: -0.002 min
Response: 37838
Conc: 0.77 ng/ml



#23 AR-1248-3


R.T.: 5.682 min
Delta R.T.: -0.008 min
Response: 20888
Conc: 0.31 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.616 min
Delta R.T.: -0.003 min
Response: 37784
Conc: 0.74 ng/ml




#24 AR-1248-4

R.T.: 6.128 min
Delta R.T.: -0.003 min
Response: 169995
Conc: 2.43 ng/ml



#24 AR-1248-4


R.T.: 4.797 min
Delta R.T.: -0.004 min
Response: 43775
Conc: 0.75 ng/ml



#25 AR-1248-5


R.T.: 6.167 min
Delta R.T.: -0.004 min
Response: 387773
Conc: 5.61 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.222 min
Delta R.T.: 0.000 min
Response: 29257
Conc: 0.51 ng/ml




#26 AR-1254-1

R.T.: 6.094 min
Delta R.T.: -0.008 min
Response: 368510
Conc: 4.88 ng/ml



#26 AR-1254-1


R.T.: 5.170 min
Delta R.T.: -0.009 min
Response: 47710
Conc: 0.55 ng/ml



#27 AR-1254-2


R.T.: 6.341 min
Delta R.T.: 0.000 min
Response: 72333
Conc: 0.64 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.346 min
Delta R.T.: 0.007 min
Response: 45842
Conc: 0.60 ng/ml




#28 AR-1254-3

R.T.: 6.745 min
Delta R.T.: 0.004 min
Response: 98132
Conc: 0.85 ng/ml



#28 AR-1254-3


R.T.: 5.786 min
Delta R.T.: 0.014 min
Response: 56197
Conc: 0.46 ng/ml



#29 AR-1254-4


R.T.: 7.027 min
Delta R.T.: -0.029 min
Response: 73195
Conc: 0.92 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.017 min
Delta R.T.: -0.004 min
Response: 110091
Conc: 1.39 ng/ml




#30 AR-1254-5

R.T.: 7.527 min
Delta R.T.: 0.010 min
Response: 196268
Conc: 2.29 ng/ml



#30 AR-1254-5


R.T.: 6.471 min
Delta R.T.: -0.003 min
Response: 78663
Conc: 0.68 ng/ml



#31 AR-1260-1


R.T.: 6.933 min
Delta R.T.: 0.010 min
Response: 247829
Conc: 2.96 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 5.910 min
Delta R.T.: -0.003 min
Response: 39380
Conc: 0.49 ng/ml




#32 AR-1260-2

R.T.: 7.218 min
Delta R.T.: 0.013 min
Response: 11765
Conc: 0.13 ng/ml



#32 AR-1260-2


R.T.: 6.123 min
Delta R.T.: 0.003 min
Response: 32202
Conc: 0.33 ng/ml



#33 AR-1260-3


R.T.: 7.608 min
Delta R.T.: 0.007 min
Response: 69947
Conc: 0.96 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.281 min
Delta R.T.: -0.002 min
Response: 7873
Conc: 0.09 ng/ml




#34 AR-1260-4

R.T.: 7.848 min
Delta R.T.: 0.000 min
Response: 115138
Conc: 1.34 ng/ml



#34 AR-1260-4


R.T.: 6.798 min
Delta R.T.: 0.002 min
Response: 94463
Conc: 1.17 ng/ml



#35 AR-1260-5


R.T.: 8.189 min
Delta R.T.: -0.003 min
Response: 82296
Conc: 0.53 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.066 min
Delta R.T.: 0.002 min
Response: 226746
Conc: 1.23 ng/ml




#36 AR-1262-1

R.T.: 7.608 min
Delta R.T.: 0.006 min
Response: 69947
Conc: 0.71 ng/ml



#36 AR-1262-1


R.T.: 6.519 min
Delta R.T.: 0.003 min
Response: 35393
Conc: 0.32 ng/ml



#37 AR-1262-2


R.T.: 8.189 min
Delta R.T.: -0.004 min
Response: 82296
Conc: 0.51 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 6.798 min
Delta R.T.: 0.001 min
Response: 94463
Conc: 0.93 ng/ml




#38 AR-1262-3

R.T.: 8.528 min
Delta R.T.: 0.002 min
Response: 55896
Conc: 0.51 ng/ml



#38 AR-1262-3


R.T.: 7.374 min
Delta R.T.: 0.002 min
Response: 10506
Conc: 0.12 ng/ml



#39 AR-1262-4


R.T.: 8.626 min
Delta R.T.: 0.007 min
Response: 77615
Conc: 1.62 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.437 min
Delta R.T.: -0.004 min
Response: 13160
Conc: 0.08 ng/ml




#40 AR-1262-5

R.T.: 9.320 min
Delta R.T.: 0.017 min
Response: 25336
Conc: 0.45 ng/ml



#40 AR-1262-5


R.T.: 7.987 min
Delta R.T.: 0.003 min
Response: 64518
Conc: 0.83 ng/ml



#41 AR-1268-1


R.T.: 8.528 min
Delta R.T.: 0.005 min
Response: 55896
Conc: 0.29 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.374 min
Delta R.T.: 0.003 min
Response: 10506
Conc: 0.04 ng/ml




#42 AR-1268-2

R.T.: 8.626 min
Delta R.T.: 0.005 min
Response: 77615
Conc: 0.44 ng/ml



#42 AR-1268-2


R.T.: 7.437 min
Delta R.T.: -0.005 min
Response: 13160
Conc: 0.06 ng/ml



#43 AR-1268-3


R.T.: 8.845 min
Delta R.T.: -0.013 min
Response: 32927
Conc: 0.21 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.655 min
Delta R.T.: -0.010 min
Response: 72526
Conc: 0.37 ng/ml




#44 AR-1268-4

R.T.: 9.320 min
Delta R.T.: 0.018 min
Response: 25336
Conc: 0.40 ng/ml



#44 AR-1268-4


R.T.: 7.987 min
Delta R.T.: 0.003 min
Response: 64518
Conc: 0.72 ng/ml



#45 AR-1268-5

R.T.: 9.728 min
Delta R.T.: 0.000 min
Response: 9027
Conc: 0.02 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.287 min
Delta R.T.: -0.004 min
Response: 26460
Conc: 0.04 ng/ml