

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030123\
 Data File : PP056116.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Mar 2023 11:47
 Operator : YP\AJ
 Sample : 01723-15
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 HH-CON-8

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/02/2023
 Supervised By :Ankita Jodhani 03/02/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 01 20:34:32 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 20 05:24:03 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.321	3.563	43936314	33314764	27.269	22.816
2) SA Decachlor...	10.063	8.562	48749817	31504541	23.897	27.770m
Target Compounds						
41) L9 AR-1268-1	8.571	7.469	3209649	2426058	11.246	14.429 #
42) L9 AR-1268-2	8.665	7.534	4075180	2487252	15.284	16.261
43) L9 AR-1268-3	8.891	7.739	3340680	2184406	14.196	15.920
44) L9 AR-1268-4	9.315	8.028	1076822	820378	10.907	14.911 #
45) L9 AR-1268-5	9.727	8.314	11260057	7164252	14.957	17.761

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030123\
Data File : PP056116.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Mar 2023 11:47
Operator : YP\AJ
Sample : 01723-15
Misc :
ALS Vial : 12 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
HH-CON-8


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/02/2023
Supervised By :Ankita Jodhani 03/02/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 01 20:34:32 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021723.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 20 05:24:03 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.321 min
Delta R.T.: -0.004 min
Response: 43936314
Conc: 27.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
HH-CON-8


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/02/2023
Supervised By :Ankita Jodhani 03/02/2023




#1 Tetrachloro-m-xylene

R.T.: 3.563 min
Delta R.T.: -0.005 min
Response: 33314764
Conc: 22.82 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.063 min
Delta R.T.: -0.012 min
Response: 48749817
Conc: 23.90 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.562 min
Delta R.T.: -0.015 min
Response: 31504541
Conc: 27.77 ng/ml m




#41 AR-1268-1

R.T.: 8.571 min
Delta R.T.: -0.005 min
Response: 3209649
Conc: 11.25 ng/ml

Instrument :
ECD_P
ClientSampleId :
HH-CON-8


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/02/2023
Supervised By :Ankita Jodhani 03/02/2023




#41 AR-1268-1

R.T.: 7.469 min
Delta R.T.: -0.009 min
Response: 2426058
Conc: 14.43 ng/ml




#42 AR-1268-2

R.T.: 8.665 min
Delta R.T.: -0.004 min
Response: 4075180
Conc: 15.28 ng/ml



#42 AR-1268-2

R.T.: 7.534 min
Delta R.T.: -0.009 min
Response: 2487252
Conc: 16.26 ng/ml




#43 AR-1268-3

R.T.: 8.891 min
Delta R.T.: -0.005 min
Response: 3340680
Conc: 14.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
HH-CON-8


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/02/2023
Supervised By :Ankita Jodhani 03/02/2023




#43 AR-1268-3

R.T.: 7.739 min
Delta R.T.: -0.009 min
Response: 2184406
Conc: 15.92 ng/ml




#44 AR-1268-4

R.T.: 9.315 min
Delta R.T.: -0.005 min
Response: 1076822
Conc: 10.91 ng/ml



#44 AR-1268-4

R.T.: 8.028 min
Delta R.T.: -0.009 min
Response: 820378
Conc: 14.91 ng/ml




#45 AR-1268-5

R.T.: 9.727 min
Delta R.T.: -0.006 min
Response: 11260057
Conc: 14.96 ng/ml

Instrument :
ECD_P
ClientSampleId :
HH-CON-8

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/02/2023
Supervised By :Ankita Jodhani 03/02/2023



#45 AR-1268-5

R.T.: 8.314 min
Delta R.T.: -0.010 min
Response: 7164252
Conc: 17.76 ng/ml