

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030222\
 Data File : PP044185.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Mar 2022 17:27
 Operator : YP\AJ
 Sample : AR1254ICC750
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 03 01:31:20 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 03 01:26:10 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.912	3.166	158.3E6	136.0E6	74.892	75.160
2) SA Decachlor...	10.052	8.536	98938029	110.6E6	74.622	74.839
Target Compounds						
26) L6 AR-1254-1	6.092	5.166	53203622	56822176	747.959	746.219
27) L6 AR-1254-2	6.331	5.327	79413338	49162020	746.611	745.450
28) L6 AR-1254-3	6.732	5.759	82910011	80673416	747.861	748.074
29) L6 AR-1254-4	7.045	6.008	57196501	52546212	747.587	745.925
30) L6 AR-1254-5	7.506	6.460	62306401	77001063	749.020	746.790



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030222\
Data File : PP044185.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Mar 2022 17:27
Operator : YP\AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 03 01:31:20 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 03 01:26:10 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.912 min
Delta R.T.: 0.000 min
Response: 158296888
Conc: 74.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.166 min
Delta R.T.: 0.000 min
Response: 135970558
Conc: 75.16 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.052 min
Delta R.T.: 0.000 min
Response: 98938029
Conc: 74.62 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.536 min
Delta R.T.: 0.000 min
Response: 110588267
Conc: 74.84 ng/ml



#26 AR-1254-1


R.T.: 6.092 min
Delta R.T.: 0.000 min
Response: 53203622
Conc: 747.96 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750




#26 AR-1254-1

R.T.: 5.166 min
Delta R.T.: 0.000 min
Response: 56822176
Conc: 746.22 ng/ml




#27 AR-1254-2

R.T.: 6.331 min
Delta R.T.: 0.000 min
Response: 79413338
Conc: 746.61 ng/ml



#27 AR-1254-2


R.T.: 5.327 min
Delta R.T.: 0.000 min
Response: 49162020
Conc: 745.45 ng/ml



#28 AR-1254-3


R.T.: 6.732 min
Delta R.T.: 0.000 min
Response: 82910011
Conc: 747.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750




#28 AR-1254-3

R.T.: 5.759 min
Delta R.T.: 0.000 min
Response: 80673416
Conc: 748.07 ng/ml




#29 AR-1254-4

R.T.: 7.045 min
Delta R.T.: 0.000 min
Response: 57196501
Conc: 747.59 ng/ml



#29 AR-1254-4


R.T.: 6.008 min
Delta R.T.: 0.000 min
Response: 52546212
Conc: 745.93 ng/ml



#30 AR-1254-5

R.T.: 7.506 min
Delta R.T.: 0.000 min
Response: 62306401
Conc: 749.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750



#30 AR-1254-5

R.T.: 6.460 min
Delta R.T.: 0.000 min
Response: 77001063
Conc: 746.79 ng/ml