

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030222\
 Data File : PP044176.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Mar 2022 15:10
 Operator : YP\AJ
 Sample : AR1242ICC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 03 00:49:38 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 03 00:46:12 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.913	3.167	108.0E6	88646462	50.000	50.000
2) SA Decachlor...	10.053	8.537	68508559	75389015	50.000	50.000
Target Compounds						
16) L4 AR-1242-1	5.165	4.313	27236871	22013246	500.000	500.000
17) L4 AR-1242-2	5.189	4.332	39734920	30503901	500.000	500.000
18) L4 AR-1242-3	5.254	4.516	24001843	17305450	500.000	500.000
19) L4 AR-1242-4	5.360	4.608	19607096	16653222	500.000	500.000
20) L4 AR-1242-5	6.161	5.167	21180547	19955453	500.000	500.000


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030222\
Data File : PP044176.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Mar 2022 15:10
Operator : YP\AJ
Sample : AR1242ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 03 00:49:38 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 03 00:46:12 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.913 min
Delta R.T.: 0.000 min
Response: 108001772
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.167 min
Delta R.T.: 0.000 min
Response: 88646462
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.053 min
Delta R.T.: 0.000 min
Response: 68508559
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.537 min
Delta R.T.: 0.000 min
Response: 75389015
Conc: 50.00 ng/ml



#16 AR-1242-1


R.T.: 5.165 min
Delta R.T.: 0.000 min
Response: 27236871
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500




#16 AR-1242-1

R.T.: 4.313 min
Delta R.T.: 0.000 min
Response: 22013246
Conc: 500.00 ng/ml




#17 AR-1242-2

R.T.: 5.189 min
Delta R.T.: 0.000 min
Response: 39734920
Conc: 500.00 ng/ml



#17 AR-1242-2


R.T.: 4.332 min
Delta R.T.: 0.000 min
Response: 30503901
Conc: 500.00 ng/ml



#18 AR-1242-3


R.T.: 5.254 min
Delta R.T.: 0.000 min
Response: 24001843
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500




#18 AR-1242-3

R.T.: 4.516 min
Delta R.T.: 0.000 min
Response: 17305450
Conc: 500.00 ng/ml




#19 AR-1242-4

R.T.: 5.360 min
Delta R.T.: 0.000 min
Response: 19607096
Conc: 500.00 ng/ml



#19 AR-1242-4


R.T.: 4.608 min
Delta R.T.: 0.000 min
Response: 16653222
Conc: 500.00 ng/ml



#20 AR-1242-5

R.T.: 6.161 min
Delta R.T.: 0.000 min
Response: 21180547
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500



#20 AR-1242-5

R.T.: 5.167 min
Delta R.T.: 0.000 min
Response: 19955453
Conc: 500.00 ng/ml