Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030322\
Data File : PP044231.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Mar 2022 20:41
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/04/2022
Supervised By :Ankita Jodhani 03/04/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 04 01:09:20 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 03 02:21:53 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1) SA	Tetrachlo...	3.911	3.166	104.6E6	86592957	50.463	49.497
2) SA	Decachlor...	10.050	8.535	67497479	74555627	49.041	48.603
	Target Compounds						
3) L1	AR-1016-1	5.164	4.311	33281372	27165501	490.115	488.581
4) L1	AR-1016-2	5.188	4.329	48615844	37898848	480.059	489.122
5) L1	AR-1016-3	5.254	4.515	28905713	21598229	460.844	503.496
6) L1	AR-1016-4	5.359	4.564	24143753	17067362	465.074	506.364
7) L1	AR-1016-5	5.681	4.789	24072860	18848802	477.962	453.143
31) L7	AR-1260-1	6.909	5.897	44461604	37026663	505.935	461.579m
32) L7	AR-1260-2	7.193	6.104	46669476	46556221	485.838	472.631m
33) L7	AR-1260-3	7.588	6.267	32027625	42732122	493.644	460.729m
34) L7	AR-1260-4	7.835	6.780	39179764	34665213	502.330	486.163
35) L7	AR-1260-5	8.178	7.048	69249795	79578582	494.091	481.946

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030322\
Data File : PP044231.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Mar 2022 20:41
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 03/04/2022
Supervised By :Ankita Jodhani 03/04/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 04 01:09:20 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 03 02:21:53 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm