

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030325\
 Data File : PP070198.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Mar 2025 19:47
 Operator : YP\AJ
 Sample : Q1476-02
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 TRE-25-0016

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 04 02:19:59 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Feb 25 05:10:19 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.525	3.828	34424400	22251084	23.456	23.274
2)	SA Decachlor...	10.248	8.880	21408548	19303015	18.795	17.805
Target Compounds							
3)	L1 AR-1016-1	5.733f	4.917	27266151	-588579	547.226	N.D. #
4)	L1 AR-1016-2	5.733	4.943	27266151	1965911	385.235	42.192 #
5)	L1 AR-1016-3	5.777	5.141	19027437	1556342	433.205	62.173 #
6)	L1 AR-1016-4	5.884	5.141	31112945	1556342	858.019	77.543 #
7)	L1 AR-1016-5	6.206f	5.391	44457672	4659246	1325.569	179.571 #
8)	L2 AR-1221-1	4.729	4.039	40986059	744502	2166.751	55.472 #
9)	L2 AR-1221-2	4.837	4.119	4371163	2246216	299.567	220.950 #
10)	L2 AR-1221-3	4.908	4.199	15014988	1695064	347.905	55.833 #
11)	L3 AR-1232-1	4.908	4.199	15014988	1695064	456.606	74.534 #
12)	L3 AR-1232-2	5.415	4.943	24429621	1965911	1502.417	86.099 #
13)	L3 AR-1232-3	5.733	5.141	27266151	1556342	782.001	129.437 #
14)	L3 AR-1232-4	5.884	5.188	31112945	1424862	1747.972	131.138 #
15)	L3 AR-1232-5	5.949	5.391	14433517	4659246	1144.864	412.125 #
16)	L4 AR-1242-1	5.733	4.943	27266151	1965911	642.234	73.719 #
17)	L4 AR-1242-2	5.733	4.943	27266151	1965911	473.843	52.845 #
18)	L4 AR-1242-3	5.777	5.141	19027437	1556342	501.708	79.059 #
19)	L4 AR-1242-4	5.884	5.188	31112945	1424862	1043.541	72.494 #
20)	L4 AR-1242-5	6.658f	5.739	20877665	7963540	592.522	320.187 #
21)	L5 AR-1248-1	5.733f	4.943	27266151	1965911	833.568	92.864 #
22)	L5 AR-1248-2	5.949	5.141	14433517	1556342	342.414	55.087 #
23)	L5 AR-1248-3	6.206f	5.188	44457672	1424862	984.181	47.885 #
24)	L5 AR-1248-4	0.000	5.391	0	4659246	N.D.	132.571 #
25)	L5 AR-1248-5	6.658f	5.768	20877665	4361685	365.083	120.595 #
26)	L6 AR-1254-1	6.477f	5.739	207.6E6	7963540	3660.615	145.934 #
27)	L6 AR-1254-2	6.744	5.854	54749208	5501632	624.382	112.864 #
28)	L6 AR-1254-3	7.086	6.297	9969873	5630017	111.862	74.258 #
29)	L6 AR-1254-4	7.391	6.550	14775866	9922467	198.304	192.819
30)	L6 AR-1254-5	7.821	6.920	27256393	5924628	405.829	87.682 #
31)	L7 AR-1260-1	7.268	6.407	15145465	15207631	259.518	306.981
32)	L7 AR-1260-2	7.569f	6.590	24630025	18735157	301.376	286.382
33)	L7 AR-1260-3	7.886	6.752	10393365	2762800	165.602	45.801 #
34)	L7 AR-1260-4	8.131	7.223	42123088	4509566	664.392	92.282 #
35)	L7 AR-1260-5	8.471f	7.457	9147507	2621035	69.730	21.993 #
36)	L8 AR-1262-1	8.131	6.974	42123088	4045340	521.783	48.462 #
37)	L8 AR-1262-2	8.471f	7.223	9147507	4509566	56.786	68.027
38)	L8 AR-1262-3	8.785	7.779	7132083	7311989	64.479	120.699 #
39)	L8 AR-1262-4	8.831	7.836	13406255	1381709	159.762	13.100 #
40)	L8 AR-1262-5	9.504	8.313	2458157	1920484	41.802	36.982
41)	L9 AR-1268-1	8.714	7.779	9205121	7311989	47.920	44.115

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030325\
 Data File : PP070198.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Mar 2025 19:47
 Operator : YP\AJ
 Sample : Q1476-02
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 TRE-25-0016

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 04 02:19:59 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Feb 25 05:10:19 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.831	7.836	13406255	1381709	80.963	9.526 #
43) L9	AR-1268-3	0.000	8.020	0	1735299	N.D.	14.170 #
44) L9	AR-1268-4	9.504	8.313	2458157	1920484	38.988	36.109
45) L9	AR-1268-5	9.927	8.612	639435	1084850	1.537	3.054 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030325\
Data File : PP070198.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Mar 2025 19:47
Operator : YP\AJ
Sample : Q1476-02
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
TRE-25-0016

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 04 02:19:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 25 05:10:19 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.525 min
Delta R.T.: -0.002 min
Response: 34424400
Conc: 23.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
TRE-25-0016




#1 Tetrachloro-m-xylene

R.T.: 3.828 min
Delta R.T.: -0.003 min
Response: 22251084
Conc: 23.27 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.248 min
Delta R.T.: -0.008 min
Response: 21408548
Conc: 18.80 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.880 min
Delta R.T.: -0.008 min
Response: 19303015
Conc: 17.81 ng/ml



#3 AR-1016-1


R.T.: 5.733 min
Delta R.T.: 0.052 min
Response: 27266151
Conc: 547.23 ng/ml

Instrument :
ECD_P
ClientSampleId :
TRE-25-0016




#3 AR-1016-1

R.T.: 4.917 min
Delta R.T.: -0.003 min
Response: -588579
Conc: N.D.




#4 AR-1016-2

R.T.: 5.733 min
Delta R.T.: 0.031 min
Response: 27266151
Conc: 385.24 ng/ml



#4 AR-1016-2


R.T.: 4.943 min
Delta R.T.: 0.004 min
Response: 1965911
Conc: 42.19 ng/ml



#5 AR-1016-3


R.T.: 5.777 min
Delta R.T.: 0.012 min
Response: 19027437
Conc: 433.21 ng/ml

Instrument :
ECD_P
ClientSampleId :
TRE-25-0016




#5 AR-1016-3

R.T.: 5.141 min
Delta R.T.: 0.024 min
Response: 1556342
Conc: 62.17 ng/ml




#6 AR-1016-4

R.T.: 5.884 min
Delta R.T.: 0.021 min
Response: 31112945
Conc: 858.02 ng/ml



#6 AR-1016-4


R.T.: 5.141 min
Delta R.T.: -0.017 min
Response: 1556342
Conc: 77.54 ng/ml



#7 AR-1016-5


R.T.: 6.206 min
Delta R.T.: 0.050 min
Response: 44457672
Conc: 1325.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
TRE-25-0016




#7 AR-1016-5

R.T.: 5.391 min
Delta R.T.: 0.018 min
Response: 4659246
Conc: 179.57 ng/ml




#8 AR-1221-1

R.T.: 4.729 min
Delta R.T.: 0.004 min
Response: 40986059
Conc: 2166.75 ng/ml



#8 AR-1221-1


R.T.: 4.039 min
Delta R.T.: 0.000 min
Response: 744502
Conc: 55.47 ng/ml



#9 AR-1221-2


R.T.: 4.837 min
Delta R.T.: 0.022 min
Response: 4371163
Conc: 299.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
TRE-25-0016




#9 AR-1221-2

R.T.: 4.119 min
Delta R.T.: -0.010 min
Response: 2246216
Conc: 220.95 ng/ml




#10 AR-1221-3

R.T.: 4.908 min
Delta R.T.: 0.019 min
Response: 15014988
Conc: 347.91 ng/ml



#10 AR-1221-3


R.T.: 4.199 min
Delta R.T.: -0.006 min
Response: 1695064
Conc: 55.83 ng/ml



#11 AR-1232-1


R.T.: 4.908 min
Delta R.T.: 0.019 min
Response: 15014988
Conc: 456.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
TRE-25-0016




#11 AR-1232-1

R.T.: 4.199 min
Delta R.T.: -0.006 min
Response: 1695064
Conc: 74.53 ng/ml




#12 AR-1232-2

R.T.: 5.415 min
Delta R.T.: -0.001 min
Response: 24429621
Conc: 1502.42 ng/ml



#12 AR-1232-2


R.T.: 4.943 min
Delta R.T.: 0.004 min
Response: 1965911
Conc: 86.10 ng/ml



#13 AR-1232-3


R.T.: 5.733 min
Delta R.T.: 0.031 min
Response: 27266151
Conc: 782.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
TRE-25-0016




#13 AR-1232-3

R.T.: 5.141 min
Delta R.T.: 0.024 min
Response: 1556342
Conc: 129.44 ng/ml




#14 AR-1232-4

R.T.: 5.884 min
Delta R.T.: 0.021 min
Response: 31112945
Conc: 1747.97 ng/ml



#14 AR-1232-4


R.T.: 5.188 min
Delta R.T.: -0.013 min
Response: 1424862
Conc: 131.14 ng/ml



#15 AR-1232-5


R.T.: 5.949 min
Delta R.T.: -0.003 min
Response: 14433517
Conc: 1144.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
TRE-25-0016




#15 AR-1232-5

R.T.: 5.391 min
Delta R.T.: 0.018 min
Response: 4659246
Conc: 412.13 ng/ml




#16 AR-1242-1

R.T.: 5.733 min
Delta R.T.: 0.031 min
Response: 27266151
Conc: 642.23 ng/ml



#16 AR-1242-1


R.T.: 4.943 min
Delta R.T.: 0.004 min
Response: 1965911
Conc: 73.72 ng/ml



#17 AR-1242-2


R.T.: 5.733 min
Delta R.T.: 0.031 min
Response: 27266151
Conc: 473.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
TRE-25-0016




#17 AR-1242-2

R.T.: 4.943 min
Delta R.T.: 0.004 min
Response: 1965911
Conc: 52.85 ng/ml




#18 AR-1242-3

R.T.: 5.777 min
Delta R.T.: 0.012 min
Response: 19027437
Conc: 501.71 ng/ml



#18 AR-1242-3


R.T.: 5.141 min
Delta R.T.: 0.024 min
Response: 1556342
Conc: 79.06 ng/ml



#19 AR-1242-4


R.T.: 5.884 min
Delta R.T.: 0.021 min
Response: 31112945
Conc: 1043.54 ng/ml

Instrument :
ECD_P
ClientSampleId :
TRE-25-0016




#19 AR-1242-4

R.T.: 5.188 min
Delta R.T.: -0.013 min
Response: 1424862
Conc: 72.49 ng/ml




#20 AR-1242-5

R.T.: 6.658 min
Delta R.T.: 0.064 min
Response: 20877665
Conc: 592.52 ng/ml



#20 AR-1242-5


R.T.: 5.739 min
Delta R.T.: 0.013 min
Response: 7963540
Conc: 320.19 ng/ml



#21 AR-1248-1


R.T.: 5.733 min
Delta R.T.: 0.052 min
Response: 27266151
Conc: 833.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
TRE-25-0016




#21 AR-1248-1

R.T.: 4.943 min
Delta R.T.: 0.004 min
Response: 1965911
Conc: 92.86 ng/ml




#22 AR-1248-2

R.T.: 5.949 min
Delta R.T.: -0.003 min
Response: 14433517
Conc: 342.41 ng/ml



#22 AR-1248-2


R.T.: 5.141 min
Delta R.T.: -0.017 min
Response: 1556342
Conc: 55.09 ng/ml



#23 AR-1248-3


R.T.: 6.206 min
Delta R.T.: 0.050 min
Response: 44457672
Conc: 984.18 ng/ml

Instrument :
ECD_P
ClientSampleId :
TRE-25-0016




#23 AR-1248-3

R.T.: 5.188 min
Delta R.T.: -0.013 min
Response: 1424862
Conc: 47.89 ng/ml




#24 AR-1248-4

R.T.: 0.000 min
Exp R.T. : 6.552 min
Response: 0
Conc: N.D.



#24 AR-1248-4


R.T.: 5.391 min
Delta R.T.: 0.018 min
Response: 4659246
Conc: 132.57 ng/ml



#25 AR-1248-5


R.T.: 6.658 min
Delta R.T.: 0.064 min
Response: 20877665
Conc: 365.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
TRE-25-0016




#25 AR-1248-5

R.T.: 5.768 min
Delta R.T.: 0.000 min
Response: 4361685
Conc: 120.59 ng/ml




#26 AR-1254-1

R.T.: 6.477 min
Delta R.T.: -0.053 min
Response: 207558034
Conc: 3660.61 ng/ml



#26 AR-1254-1


R.T.: 5.739 min
Delta R.T.: 0.013 min
Response: 7963540
Conc: 145.93 ng/ml



#27 AR-1254-2


R.T.: 6.744 min
Delta R.T.: -0.003 min
Response: 54749208
Conc: 624.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
TRE-25-0016




#27 AR-1254-2

R.T.: 5.854 min
Delta R.T.: -0.021 min
Response: 5501632
Conc: 112.86 ng/ml




#28 AR-1254-3

R.T.: 7.086 min
Delta R.T.: -0.023 min
Response: 9969873
Conc: 111.86 ng/ml



#28 AR-1254-3


R.T.: 6.297 min
Delta R.T.: 0.000 min
Response: 5630017
Conc: 74.26 ng/ml



#29 AR-1254-4


R.T.: 7.391 min
Delta R.T.: -0.002 min
Response: 14775866
Conc: 198.30 ng/ml

Instrument :
ECD_P
ClientSampleId :
TRE-25-0016




#29 AR-1254-4

R.T.: 6.550 min
Delta R.T.: 0.011 min
Response: 9922467
Conc: 192.82 ng/ml




#30 AR-1254-5

R.T.: 7.821 min
Delta R.T.: 0.012 min
Response: 27256393
Conc: 405.83 ng/ml



#30 AR-1254-5


R.T.: 6.920 min
Delta R.T.: -0.007 min
Response: 5924628
Conc: 87.68 ng/ml



#31 AR-1260-1


R.T.: 7.268 min
Delta R.T.: -0.007 min
Response: 15145465
Conc: 259.52 ng/ml

Instrument :
ECD_P
ClientSampleId :
TRE-25-0016




#31 AR-1260-1

R.T.: 6.407 min
Delta R.T.: -0.005 min
Response: 15207631
Conc: 306.98 ng/ml




#32 AR-1260-2

R.T.: 7.569 min
Delta R.T.: 0.040 min
Response: 24630025
Conc: 301.38 ng/ml



#32 AR-1260-2


R.T.: 6.590 min
Delta R.T.: -0.010 min
Response: 18735157
Conc: 286.38 ng/ml



#33 AR-1260-3


R.T.: 7.886 min
Delta R.T.: -0.001 min
Response: 10393365
Conc: 165.60 ng/ml

Instrument :
ECD_P
ClientSampleId :
TRE-25-0016




#33 AR-1260-3

R.T.: 6.752 min
Delta R.T.: -0.001 min
Response: 2762800
Conc: 45.80 ng/ml




#34 AR-1260-4

R.T.: 8.131 min
Delta R.T.: 0.020 min
Response: 42123088
Conc: 664.39 ng/ml



#34 AR-1260-4


R.T.: 7.223 min
Delta R.T.: -0.003 min
Response: 4509566
Conc: 92.28 ng/ml



#35 AR-1260-5


R.T.: 8.471 min
Delta R.T.: 0.038 min
Response: 9147507
Conc: 69.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
TRE-25-0016




#35 AR-1260-5

R.T.: 7.457 min
Delta R.T.: -0.010 min
Response: 2621035
Conc: 21.99 ng/ml




#36 AR-1262-1

R.T.: 8.131 min
Delta R.T.: 0.020 min
Response: 42123088
Conc: 521.78 ng/ml



#36 AR-1262-1


R.T.: 6.974 min
Delta R.T.: 0.008 min
Response: 4045340
Conc: 48.46 ng/ml



#37 AR-1262-2


R.T.: 8.471 min
Delta R.T.: 0.038 min
Response: 9147507
Conc: 56.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
TRE-25-0016




#37 AR-1262-2

R.T.: 7.223 min
Delta R.T.: -0.003 min
Response: 4509566
Conc: 68.03 ng/ml




#38 AR-1262-3

R.T.: 8.785 min
Delta R.T.: 0.028 min
Response: 7132083
Conc: 64.48 ng/ml



#38 AR-1262-3


R.T.: 7.779 min
Delta R.T.: 0.028 min
Response: 7311989
Conc: 120.70 ng/ml



#39 AR-1262-4


R.T.: 8.831 min
Delta R.T.: 0.005 min
Response: 13406255
Conc: 159.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
TRE-25-0016




#39 AR-1262-4

R.T.: 7.836 min
Delta R.T.: 0.022 min
Response: 1381709
Conc: 13.10 ng/ml




#40 AR-1262-5

R.T.: 9.504 min
Delta R.T.: 0.010 min
Response: 2458157
Conc: 41.80 ng/ml



#40 AR-1262-5


R.T.: 8.313 min
Delta R.T.: -0.004 min
Response: 1920484
Conc: 36.98 ng/ml



#41 AR-1268-1


R.T.: 8.714 min
Delta R.T.: -0.032 min
Response: 9205121
Conc: 47.92 ng/ml

Instrument :
ECD_P
ClientSampleId :
TRE-25-0016




#41 AR-1268-1

R.T.: 7.779 min
Delta R.T.: 0.029 min
Response: 7311989
Conc: 44.11 ng/ml




#42 AR-1268-2

R.T.: 8.831 min
Delta R.T.: -0.008 min
Response: 13406255
Conc: 80.96 ng/ml



#42 AR-1268-2


R.T.: 7.836 min
Delta R.T.: 0.021 min
Response: 1381709
Conc: 9.53 ng/ml



#43 AR-1268-3


R.T.: 0.000 min
Exp R.T.: 9.073 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
TRE-25-0016




#43 AR-1268-3

R.T.: 8.020 min
Delta R.T.: -0.002 min
Response: 1735299
Conc: 14.17 ng/ml




#44 AR-1268-4

R.T.: 9.504 min
Delta R.T.: 0.012 min
Response: 2458157
Conc: 38.99 ng/ml



#44 AR-1268-4


R.T.: 8.313 min
Delta R.T.: -0.003 min
Response: 1920484
Conc: 36.11 ng/ml



#45 AR-1268-5

R.T.: 9.927 min
Delta R.T.: 0.016 min
Response: 639435
Conc: 1.54 ng/ml

Instrument :
ECD_P
ClientSampleId :
TRE-25-0016



#45 AR-1268-5

R.T.: 8.612 min
Delta R.T.: -0.009 min
Response: 1084850
Conc: 3.05 ng/ml