

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030325\
 Data File : PP070204.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Mar 2025 21:56
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/04/2025
 Supervised By :Ankita Jodhani 03/04/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 04 03:17:11 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Feb 25 05:10:19 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.522	3.827	74645653	47672512	50.862	49.865
2) SA Decachlor...	10.246	8.880	51753165	45770343	45.436	42.220
Target Compounds						
26) L6 AR-1254-1	6.524	5.722	26812854	25537326	472.887	467.979
27) L6 AR-1254-2	6.740	5.870	39715447	22326616	452.931	458.024
28) L6 AR-1254-3	7.103	6.275	39815406	34367133	446.730	453.293
29) L6 AR-1254-4	7.385	6.503f	33095483	24006893	444.168	466.515m
30) L6 AR-1254-5	7.802	6.922	30893323	30143822	459.980	446.118

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030325\
Data File : PP070204.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Mar 2025 21:56
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/04/2025
Supervised By :Ankita Jodhani 03/04/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 04 03:17:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 25 05:10:19 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.522 min
Delta R.T.: -0.005 min
Response: 74645653
Conc: 50.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/04/2025
Supervised By :Ankita Jodhani 03/04/2025




#1 Tetrachloro-m-xylene

R.T.: 3.827 min
Delta R.T.: -0.003 min
Response: 47672512
Conc: 49.86 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.246 min
Delta R.T.: -0.010 min
Response: 51753165
Conc: 45.44 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.880 min
Delta R.T.: -0.008 min
Response: 45770343
Conc: 42.22 ng/ml




#26 AR-1254-1

R.T.: 6.524 min
Delta R.T.: -0.006 min
Response: 26812854
Conc: 472.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/04/2025
Supervised By :Ankita Jodhani 03/04/2025




#26 AR-1254-1

R.T.: 5.722 min
Delta R.T.: -0.005 min
Response: 25537326
Conc: 467.98 ng/ml




#27 AR-1254-2

R.T.: 6.740 min
Delta R.T.: -0.007 min
Response: 39715447
Conc: 452.93 ng/ml



#27 AR-1254-2

R.T.: 5.870 min
Delta R.T.: -0.005 min
Response: 22326616
Conc: 458.02 ng/ml




#28 AR-1254-3

R.T.: 7.103 min
Delta R.T.: -0.007 min
Response: 39815406
Conc: 446.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/04/2025
Supervised By :Ankita Jodhani 03/04/2025




#28 AR-1254-3

R.T.: 6.275 min
Delta R.T.: -0.021 min
Response: 34367133
Conc: 453.29 ng/ml




#29 AR-1254-4

R.T.: 7.385 min
Delta R.T.: -0.008 min
Response: 33095483
Conc: 444.17 ng/ml



#29 AR-1254-4

R.T.: 6.503 min
Delta R.T.: -0.035 min
Response: 24006893
Conc: 466.52 ng/ml m




#30 AR-1254-5

R.T.: 7.802 min
Delta R.T.: -0.006 min
Response: 30893323
Conc: 459.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/04/2025
Supervised By :Ankita Jodhani 03/04/2025



#30 AR-1254-5

R.T.: 6.922 min
Delta R.T.: -0.005 min
Response: 30143822
Conc: 446.12 ng/ml