

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030421\
 Data File : PP033679.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 04 Mar 2021 19:42
 Operator : DD\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP030421

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 05 04:42:39 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Mar 05 04:37:34 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.861	4.274	5431887	2038133	47.287	51.192
2) SA Decachlor...	10.796	9.616	4429682	1210681	48.951	51.824
Target Compounds						
26) L6 AR-1254-1	7.079	6.408	1675903	785215	481.736	498.954
27) L6 AR-1254-2	7.309	6.565	2541926	677206	482.942	494.315
28) L6 AR-1254-3	7.688	6.987	2711697	1066361	485.877	497.337
29) L6 AR-1254-4	7.985	7.223	1816588	557516	473.480	504.558
30) L6 AR-1254-5	8.412	7.652	2076031	822046	495.196	479.791



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030421\
Data File : PP033679.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 Mar 2021 19:42
Operator : DD\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP030421

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 05 04:42:39 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030421.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 05 04:37:34 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.861 min
Delta R.T.: 0.002 min
Response: 5431887
Conc: 47.29 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP030421




#1 Tetrachloro-m-xylene

R.T.: 4.274 min
Delta R.T.: 0.002 min
Response: 2038133
Conc: 51.19 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.796 min
Delta R.T.: 0.004 min
Response: 4429682
Conc: 48.95 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.616 min
Delta R.T.: -0.002 min
Response: 1210681
Conc: 51.82 ng/ml



#26 AR-1254-1


R.T.: 7.079 min
Delta R.T.: 0.000 min
Response: 1675903
Conc: 481.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP030421




#26 AR-1254-1

R.T.: 6.408 min
Delta R.T.: 0.002 min
Response: 785215
Conc: 498.95 ng/ml




#27 AR-1254-2

R.T.: 7.309 min
Delta R.T.: 0.000 min
Response: 2541926
Conc: 482.94 ng/ml



#27 AR-1254-2


R.T.: 6.565 min
Delta R.T.: 0.001 min
Response: 677206
Conc: 494.31 ng/ml



#28 AR-1254-3


R.T.: 7.688 min
Delta R.T.: 0.000 min
Response: 2711697
Conc: 485.88 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP030421




#28 AR-1254-3

R.T.: 6.987 min
Delta R.T.: 0.000 min
Response: 1066361
Conc: 497.34 ng/ml




#29 AR-1254-4

R.T.: 7.985 min
Delta R.T.: 0.000 min
Response: 1816588
Conc: 473.48 ng/ml



#29 AR-1254-4


R.T.: 7.223 min
Delta R.T.: 0.001 min
Response: 557516
Conc: 504.56 ng/ml



#30 AR-1254-5

R.T.: 8.412 min
Delta R.T.: 0.000 min
Response: 2076031
Conc: 495.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP030421



#30 AR-1254-5

R.T.: 7.652 min
Delta R.T.: 0.001 min
Response: 822046
Conc: 479.79 ng/ml