

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030422\
 Data File : PP044257.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Mar 2022 12:51
 Operator : YP\AJ
 Sample : N1786-01 10X
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 05 00:54:35 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 03 02:21:53 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.912	3.166	3765463	4676160	1.817	2.673 #
2)	SA Decachlor...	10.049	8.535	3061493	3505366	2.224	2.285
Target Compounds							
3)	L1 AR-1016-1	5.157	4.313	49916	24345	0.735	0.438 #
4)	L1 AR-1016-2	5.184	4.332	118226	66783	1.167	0.862 #
5)	L1 AR-1016-3	5.242	4.514	25265	119439	0.403	2.784 #
6)	L1 AR-1016-4	5.351	4.565	119830	112072	2.308	3.325 #
7)	L1 AR-1016-5	5.686	4.793	75628	84058	1.502	2.021 #
8)	L2 AR-1221-1	4.140	3.388	149392	72117	6.046	3.141 #
9)	L2 AR-1221-2	4.230	3.478	76087	100687	4.055	5.829 #
10)	L2 AR-1221-3	4.311	3.561	61432	42384	1.098	0.852
11)	L3 AR-1232-1	4.311	3.561	61432	42384	1.339	1.020
12)	L3 AR-1232-2	4.880	4.332	96702	66783	4.297	1.934 #
13)	L3 AR-1232-3	5.184	4.514	118226	119439	2.732	6.402 #
14)	L3 AR-1232-4	5.351	4.605	119830	78207	5.968	4.743
15)	L3 AR-1232-5	5.464	4.793	22663	84058	1.494	4.643 #
16)	L4 AR-1242-1	5.157	4.313	49916	24345	0.917	0.531 #
17)	L4 AR-1242-2	5.184	4.332	118226	66783	1.485	1.057 #
18)	L4 AR-1242-3	5.242	4.514	25265	119439	0.535	3.344 #
19)	L4 AR-1242-4	5.351	4.605	119830	78207	3.096	2.261 #
20)	L4 AR-1242-5	6.150	5.168	240380	33942	5.638	0.804 #
21)	L5 AR-1248-1	5.157	4.313	49916	24345	1.207	0.686 #
22)	L5 AR-1248-2	5.464	4.565	22663	112072	0.404	2.381 #
23)	L5 AR-1248-3	5.686	4.605	75628	78207	1.107	1.587 #
24)	L5 AR-1248-4	6.123	4.793	48446	84058	0.664	1.487 #
25)	L5 AR-1248-5	6.150	5.211	240380	33010	3.273	0.589 #
26)	L6 AR-1254-1	6.108f	5.168	85033	33942	1.134	0.411 #
27)	L6 AR-1254-2	6.337	5.326	248256	178462	2.209	2.481
28)	L6 AR-1254-3	6.734	5.759	250836	58819	2.144	0.511 #
29)	L6 AR-1254-4	7.043	6.012	884136	64857	10.997	0.862 #
30)	L6 AR-1254-5	7.506	6.459	400224	150282	4.598	1.379 #
31)	L7 AR-1260-1	6.912	5.901	127383	74903	1.450	0.934 #
32)	L7 AR-1260-2	7.198	6.107	219959	189827	2.290	1.927
33)	L7 AR-1260-3	7.587	6.267	108798	146576	1.677	1.580
34)	L7 AR-1260-4	7.836	6.781	113550	54508	1.456	0.764 #
35)	L7 AR-1260-5	8.181	7.047	786187	314529	5.609	1.905 #
36)	L8 AR-1262-1	7.587	6.504	108798	44790	1.076	0.432 #
37)	L8 AR-1262-2	8.181	6.781	786187	54508	4.735	0.570 #
38)	L8 AR-1262-3	8.515	7.349	96941	54617	0.866	0.692
39)	L8 AR-1262-4	8.598	7.426	74290	159281	1.497	1.103 #
40)	L8 AR-1262-5	9.303f	7.975	21629	396042	0.379	5.532 #
41)	L9 AR-1268-1	8.507	7.363	30232	38981	0.155	0.171

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030422\
 Data File : PP044257.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Mar 2022 12:51
 Operator : YP\AJ
 Sample : N1786-01 10X
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 05 00:54:35 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 03 02:21:53 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.598	7.426	74290	159281	0.408	0.778 #
43) L9	AR-1268-3	8.835	7.652	91011	14798	0.590	0.085 #
44) L9	AR-1268-4	9.303f	7.975	21629	396042	0.346	4.880 #
45) L9	AR-1268-5	9.710	8.277	165895	28416	0.353	0.053 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030422\
Data File : PP044257.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Mar 2022 12:51
Operator : YP\AJ
Sample : N1786-01 10X
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 05 00:54:35 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 03 02:21:53 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.912 min
Delta R.T.: 0.000 min
Response: 3765463
Conc: 1.82 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.166 min
Delta R.T.: 0.000 min
Response: 4676160
Conc: 2.67 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.049 min
Delta R.T.: -0.003 min
Response: 3061493
Conc: 2.22 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.535 min
Delta R.T.: -0.002 min
Response: 3505366
Conc: 2.29 ng/ml



#3 AR-1016-1


R.T.: 5.157 min
Delta R.T.: -0.008 min
Response: 49916
Conc: 0.74 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.313 min
Delta R.T.: 0.000 min
Response: 24345
Conc: 0.44 ng/ml




#4 AR-1016-2

R.T.: 5.184 min
Delta R.T.: -0.004 min
Response: 118226
Conc: 1.17 ng/ml



#4 AR-1016-2


R.T.: 4.332 min
Delta R.T.: 0.000 min
Response: 66783
Conc: 0.86 ng/ml



#5 AR-1016-3


R.T.: 5.242 min
Delta R.T.: -0.013 min
Response: 25265
Conc: 0.40 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.514 min
Delta R.T.: -0.002 min
Response: 119439
Conc: 2.78 ng/ml




#6 AR-1016-4

R.T.: 5.351 min
Delta R.T.: -0.009 min
Response: 119830
Conc: 2.31 ng/ml



#6 AR-1016-4


R.T.: 4.565 min
Delta R.T.: 0.000 min
Response: 112072
Conc: 3.33 ng/ml



#7 AR-1016-5


R.T.: 5.686 min
Delta R.T.: 0.005 min
Response: 75628
Conc: 1.50 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 4.793 min
Delta R.T.: 0.003 min
Response: 84058
Conc: 2.02 ng/ml




#8 AR-1221-1

R.T.: 4.140 min
Delta R.T.: 0.001 min
Response: 149392
Conc: 6.05 ng/ml



#8 AR-1221-1


R.T.: 3.388 min
Delta R.T.: -0.006 min
Response: 72117
Conc: 3.14 ng/ml



#9 AR-1221-2


R.T.: 4.230 min
Delta R.T.: 0.000 min
Response: 76087
Conc: 4.06 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.478 min
Delta R.T.: -0.005 min
Response: 100687
Conc: 5.83 ng/ml




#10 AR-1221-3

R.T.: 4.311 min
Delta R.T.: 0.000 min
Response: 61432
Conc: 1.10 ng/ml



#10 AR-1221-3


R.T.: 3.561 min
Delta R.T.: -0.002 min
Response: 42384
Conc: 0.85 ng/ml



#11 AR-1232-1


R.T.: 4.311 min
Delta R.T.: 0.000 min
Response: 61432
Conc: 1.34 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 3.561 min
Delta R.T.: -0.002 min
Response: 42384
Conc: 1.02 ng/ml




#12 AR-1232-2

R.T.: 4.880 min
Delta R.T.: 0.004 min
Response: 96702
Conc: 4.30 ng/ml



#12 AR-1232-2


R.T.: 4.332 min
Delta R.T.: 0.001 min
Response: 66783
Conc: 1.93 ng/ml



#13 AR-1232-3


R.T.: 5.184 min
Delta R.T.: -0.005 min
Response: 118226
Conc: 2.73 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.514 min
Delta R.T.: -0.002 min
Response: 119439
Conc: 6.40 ng/ml




#14 AR-1232-4

R.T.: 5.351 min
Delta R.T.: -0.010 min
Response: 119830
Conc: 5.97 ng/ml



#14 AR-1232-4


R.T.: 4.605 min
Delta R.T.: -0.003 min
Response: 78207
Conc: 4.74 ng/ml



#15 AR-1232-5


R.T.: 5.464 min
Delta R.T.: 0.000 min
Response: 22663
Conc: 1.49 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 4.793 min
Delta R.T.: 0.003 min
Response: 84058
Conc: 4.64 ng/ml




#16 AR-1242-1

R.T.: 5.157 min
Delta R.T.: -0.008 min
Response: 49916
Conc: 0.92 ng/ml



#16 AR-1242-1


R.T.: 4.313 min
Delta R.T.: 0.001 min
Response: 24345
Conc: 0.53 ng/ml



#17 AR-1242-2


R.T.: 5.184 min
Delta R.T.: -0.004 min
Response: 118226
Conc: 1.48 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.332 min
Delta R.T.: 0.001 min
Response: 66783
Conc: 1.06 ng/ml




#18 AR-1242-3

R.T.: 5.242 min
Delta R.T.: -0.014 min
Response: 25265
Conc: 0.54 ng/ml



#18 AR-1242-3


R.T.: 4.514 min
Delta R.T.: -0.002 min
Response: 119439
Conc: 3.34 ng/ml



#19 AR-1242-4


R.T.: 5.351 min
Delta R.T.: -0.009 min
Response: 119830
Conc: 3.10 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.605 min
Delta R.T.: -0.003 min
Response: 78207
Conc: 2.26 ng/ml




#20 AR-1242-5

R.T.: 6.150 min
Delta R.T.: -0.011 min
Response: 240380
Conc: 5.64 ng/ml



#20 AR-1242-5


R.T.: 5.168 min
Delta R.T.: 0.000 min
Response: 33942
Conc: 0.80 ng/ml



#21 AR-1248-1


R.T.: 5.157 min
Delta R.T.: -0.008 min
Response: 49916
Conc: 1.21 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.313 min
Delta R.T.: 0.000 min
Response: 24345
Conc: 0.69 ng/ml




#22 AR-1248-2

R.T.: 5.464 min
Delta R.T.: 0.002 min
Response: 22663
Conc: 0.40 ng/ml



#22 AR-1248-2


R.T.: 4.565 min
Delta R.T.: 0.000 min
Response: 112072
Conc: 2.38 ng/ml



#23 AR-1248-3


R.T.: 5.686 min
Delta R.T.: 0.005 min
Response: 75628
Conc: 1.11 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.605 min
Delta R.T.: -0.003 min
Response: 78207
Conc: 1.59 ng/ml




#24 AR-1248-4

R.T.: 6.123 min
Delta R.T.: 0.003 min
Response: 48446
Conc: 0.66 ng/ml



#24 AR-1248-4


R.T.: 4.793 min
Delta R.T.: 0.003 min
Response: 84058
Conc: 1.49 ng/ml



#25 AR-1248-5


R.T.: 6.150 min
Delta R.T.: -0.011 min
Response: 240380
Conc: 3.27 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.211 min
Delta R.T.: 0.000 min
Response: 33010
Conc: 0.59 ng/ml




#26 AR-1254-1

R.T.: 6.108 min
Delta R.T.: 0.017 min
Response: 85033
Conc: 1.13 ng/ml



#26 AR-1254-1


R.T.: 5.168 min
Delta R.T.: 0.000 min
Response: 33942
Conc: 0.41 ng/ml



#27 AR-1254-2


R.T.: 6.337 min
Delta R.T.: 0.006 min
Response: 248256
Conc: 2.21 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.326 min
Delta R.T.: 0.000 min
Response: 178462
Conc: 2.48 ng/ml




#28 AR-1254-3

R.T.: 6.734 min
Delta R.T.: 0.003 min
Response: 250836
Conc: 2.14 ng/ml



#28 AR-1254-3


R.T.: 5.759 min
Delta R.T.: 0.000 min
Response: 58819
Conc: 0.51 ng/ml



#29 AR-1254-4


R.T.: 7.043 min
Delta R.T.: -0.002 min
Response: 884136
Conc: 11.00 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.012 min
Delta R.T.: 0.003 min
Response: 64857
Conc: 0.86 ng/ml




#30 AR-1254-5

R.T.: 7.506 min
Delta R.T.: 0.000 min
Response: 400224
Conc: 4.60 ng/ml



#30 AR-1254-5


R.T.: 6.459 min
Delta R.T.: -0.001 min
Response: 150282
Conc: 1.38 ng/ml



#31 AR-1260-1


R.T.: 6.912 min
Delta R.T.: 0.000 min
Response: 127383
Conc: 1.45 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 5.901 min
Delta R.T.: 0.000 min
Response: 74903
Conc: 0.93 ng/ml




#32 AR-1260-2

R.T.: 7.198 min
Delta R.T.: 0.003 min
Response: 219959
Conc: 2.29 ng/ml



#32 AR-1260-2


R.T.: 6.107 min
Delta R.T.: 0.000 min
Response: 189827
Conc: 1.93 ng/ml



#33 AR-1260-3


R.T.: 7.587 min
Delta R.T.: -0.002 min
Response: 108798
Conc: 1.68 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.267 min
Delta R.T.: -0.002 min
Response: 146576
Conc: 1.58 ng/ml




#34 AR-1260-4

R.T.: 7.836 min
Delta R.T.: 0.000 min
Response: 113550
Conc: 1.46 ng/ml



#34 AR-1260-4


R.T.: 6.781 min
Delta R.T.: -0.001 min
Response: 54508
Conc: 0.76 ng/ml



#35 AR-1260-5


R.T.: 8.181 min
Delta R.T.: 0.000 min
Response: 786187
Conc: 5.61 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.047 min
Delta R.T.: -0.003 min
Response: 314529
Conc: 1.90 ng/ml




#36 AR-1262-1

R.T.: 7.587 min
Delta R.T.: -0.003 min
Response: 108798
Conc: 1.08 ng/ml



#36 AR-1262-1


R.T.: 6.504 min
Delta R.T.: 0.002 min
Response: 44790
Conc: 0.43 ng/ml



#37 AR-1262-2


R.T.: 8.181 min
Delta R.T.: 0.001 min
Response: 786187
Conc: 4.73 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 6.781 min
Delta R.T.: 0.000 min
Response: 54508
Conc: 0.57 ng/ml




#38 AR-1262-3

R.T.: 8.515 min
Delta R.T.: 0.003 min
Response: 96941
Conc: 0.87 ng/ml



#38 AR-1262-3


R.T.: 7.349 min
Delta R.T.: -0.007 min
Response: 54617
Conc: 0.69 ng/ml



#39 AR-1262-4


R.T.: 8.598 min
Delta R.T.: -0.007 min
Response: 74290
Conc: 1.50 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.426 min
Delta R.T.: 0.000 min
Response: 159281
Conc: 1.10 ng/ml




#40 AR-1262-5

R.T.: 9.303 min
Delta R.T.: 0.016 min
Response: 21629
Conc: 0.38 ng/ml



#40 AR-1262-5


R.T.: 7.975 min
Delta R.T.: 0.007 min
Response: 396042
Conc: 5.53 ng/ml



#41 AR-1268-1


R.T.: 8.507 min
Delta R.T.: -0.002 min
Response: 30232
Conc: 0.15 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.363 min
Delta R.T.: 0.007 min
Response: 38981
Conc: 0.17 ng/ml




#42 AR-1268-2

R.T.: 8.598 min
Delta R.T.: -0.011 min
Response: 74290
Conc: 0.41 ng/ml



#42 AR-1268-2


R.T.: 7.426 min
Delta R.T.: -0.002 min
Response: 159281
Conc: 0.78 ng/ml



#43 AR-1268-3


R.T.: 8.835 min
Delta R.T.: -0.009 min
Response: 91011
Conc: 0.59 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.652 min
Delta R.T.: 0.003 min
Response: 14798
Conc: 0.08 ng/ml




#44 AR-1268-4

R.T.: 9.303 min
Delta R.T.: 0.016 min
Response: 21629
Conc: 0.35 ng/ml



#44 AR-1268-4


R.T.: 7.975 min
Delta R.T.: 0.006 min
Response: 396042
Conc: 4.88 ng/ml



#45 AR-1268-5

R.T.: 9.710 min
Delta R.T.: 0.000 min
Response: 165895
Conc: 0.35 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.277 min
Delta R.T.: 0.003 min
Response: 28416
Conc: 0.05 ng/ml