

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030422\
 Data File : PP044247.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Mar 2022 10:03
 Operator : YP\AJ
 Sample : N1748-03 10X
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 31462-63

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/07/2022
 Supervised By :Ankita Jodhani 03/07/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 04 16:02:32 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 03 02:21:53 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.912	3.167	3257446	3593864	1.572m	2.054 #
2) SA Decachlor...	10.050	8.537	3626934	4639052	2.635m	3.024m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030422\
Data File : PP044247.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Mar 2022 10:03
Operator : YP\AJ
Sample : N1748-03 10X
Misc :
ALS Vial : 8 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
31462-63


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/07/2022
Supervised By :Ankita Jodhani 03/07/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 04 16:02:32 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 03 02:21:53 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.912 min
Delta R.T.: 0.000 min
Response: 3257446
Conc: 1.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
31462-63


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/07/2022
Supervised By :Ankita Jodhani 03/07/2022




#1 Tetrachloro-m-xylene

R.T.: 3.167 min
Delta R.T.: 0.000 min
Response: 3593864
Conc: 2.05 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.050 min
Delta R.T.: -0.002 min
Response: 3626934
Conc: 2.64 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.537 min
Delta R.T.: 0.000 min
Response: 4639052
Conc: 3.02 ng/ml m