Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030521\
Data File : PP033704.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 05 Mar 2021 15:17
Operator : DD\AJ
Sample : M1566-01MSD
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
RO-282MSD

Manual Integrations APPROVED
Ankita
3/8/2021 12:10:44 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 05 15:49:58 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030421.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 05 04:37:34 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.860	4.274	1923439	710931	16.745	17.857m
2)	SA Decachlor...	10.795	9.615	1399967	399166	15.471	17.087
	Target Compounds						
3)	L1 AR-1016-1	6.173	5.534	1415492	526638	425.965	491.072
4)	L1 AR-1016-2	6.197	5.555	2082613	742250	418.774	479.395
5)	L1 AR-1016-3	6.263	5.748	1274843	413425	420.033	500.185
6)	L1 AR-1016-4	6.369	5.798	1040455	315635	414.848	485.580
7)	L1 AR-1016-5	6.683	6.029	1021936	406294	413.472	481.314
31)	L7 AR-1260-1	7.855	7.122	1895368	699750	458.274	516.642
32)	L7 AR-1260-2	8.121	7.316	2267216	769950	476.390	497.478
33)	L7 AR-1260-3	8.485	7.473	1405444	751620	368.100	507.603 #
34)	L7 AR-1260-4	8.715	7.955	1768363	539560	399.711	440.410
35)	L7 AR-1260-5	9.035	8.197	3384088	1160412	398.534	441.823

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030521\
Data File : PP033704.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 05 Mar 2021 15:17
Operator : DD\AJ
Sample : M1566-01MSD
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
RO-282MSD

Manual Integrations APPROVED
Ankita
3/8/2021 12:10:44 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 05 15:49:58 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030421.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 05 04:37:34 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm