

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030521\
 Data File : PP033709.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 05 Mar 2021 16:42
 Operator : DD\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 05 17:59:44 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Mar 05 04:37:34 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.860	4.274	4985379	1935316	43.400	48.610
2) SA Decachlor...	10.796	9.615	4477267	1249660	49.477	53.493
Target Compounds						
26) L6 AR-1254-1	7.080	6.408	1642409	784430	472.108	498.455
27) L6 AR-1254-2	7.309	6.565	2406994	679627	457.306	496.082
28) L6 AR-1254-3	7.688	6.986	2756800	1144705	493.958	533.876
29) L6 AR-1254-4	7.984	7.222	1848097	590410	481.692	534.327
30) L6 AR-1254-5	8.411	7.651	2051677	865450	489.387	505.124



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030521\
Data File : PP033709.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 05 Mar 2021 16:42
Operator : DD\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 05 17:59:44 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030421.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 05 04:37:34 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.860 min
Delta R.T.: 0.001 min
Response: 4985379
Conc: 43.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.274 min
Delta R.T.: 0.003 min
Response: 1935316
Conc: 48.61 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.796 min
Delta R.T.: 0.004 min
Response: 4477267
Conc: 49.48 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.615 min
Delta R.T.: -0.003 min
Response: 1249660
Conc: 53.49 ng/ml



#26 AR-1254-1


R.T.: 7.080 min
Delta R.T.: 0.000 min
Response: 1642409
Conc: 472.11 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 6.408 min
Delta R.T.: 0.001 min
Response: 784430
Conc: 498.46 ng/ml




#27 AR-1254-2

R.T.: 7.309 min
Delta R.T.: 0.000 min
Response: 2406994
Conc: 457.31 ng/ml



#27 AR-1254-2


R.T.: 6.565 min
Delta R.T.: 0.000 min
Response: 679627
Conc: 496.08 ng/ml



#28 AR-1254-3


R.T.: 7.688 min
Delta R.T.: 0.000 min
Response: 2756800
Conc: 493.96 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.986 min
Delta R.T.: 0.000 min
Response: 1144705
Conc: 533.88 ng/ml




#29 AR-1254-4

R.T.: 7.984 min
Delta R.T.: 0.000 min
Response: 1848097
Conc: 481.69 ng/ml



#29 AR-1254-4


R.T.: 7.222 min
Delta R.T.: 0.000 min
Response: 590410
Conc: 534.33 ng/ml



#30 AR-1254-5

R.T.: 8.411 min
Delta R.T.: 0.000 min
Response: 2051677
Conc: 489.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 7.651 min
Delta R.T.: 0.000 min
Response: 865450
Conc: 505.12 ng/ml