

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030521\
 Data File : PP033715.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 05 Mar 2021 19:31
 Operator : DD\AJ
 Sample : M1573-01 10X
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 JEH475K-1-1

Manual Integrations
 APPROVED

Ankita
 3/8/2021 12:10:58 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 06 00:04:31 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Mar 05 04:37:34 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.860	4.273	158115	79228	1.376	1.990 #
2) SA Decachlor...	10.795	9.616	203195	49979	2.245m	2.139
Target Compounds						
26) L6 AR-1254-1	7.079	6.407	317092	195137	91.148m	123.997m#
27) L6 AR-1254-2	7.309	6.564	468893	147710	89.085	107.818
28) L6 AR-1254-3	7.688	6.986	433518	199850	77.677	93.207
29) L6 AR-1254-4	7.984	7.222	365967	135540	95.386	122.665 #
30) L6 AR-1254-5	8.412	7.651	306667	153907	73.149	89.829

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030521\
Data File : PP033715.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 05 Mar 2021 19:31
Operator : DD\AJ
Sample : M1573-01 10X
Misc :
ALS Vial : 26 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
JEH475K-1-1


Manual Integrations
APPROVED

Ankita
3/8/2021 12:10:58 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 06 00:04:31 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030421.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 05 04:37:34 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.860 min
Delta R.T.: 0.000 min
Response: 158115
Conc: 1.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
JEH475K-1-1


Manual Integrations
APPROVED

Ankita
3/8/2021 12:10:58 PM




#1 Tetrachloro-m-xylene

R.T.: 4.273 min
Delta R.T.: 0.002 min
Response: 79228
Conc: 1.99 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.795 min
Delta R.T.: 0.003 min
Response: 203195
Conc: 2.25 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.616 min
Delta R.T.: -0.002 min
Response: 49979
Conc: 2.14 ng/ml




#26 AR-1254-1

R.T.: 7.079 min
Delta R.T.: -0.001 min
Response: 317092
Conc: 91.15 ng/ml m

Instrument :
ECD_P
ClientSampleId :
JEH475K-1-1


Manual Integrations
APPROVED

Ankita
3/8/2021 12:10:58 PM




#26 AR-1254-1

R.T.: 6.407 min
Delta R.T.: 0.000 min
Response: 195137
Conc: 124.00 ng/ml m




#27 AR-1254-2

R.T.: 7.309 min
Delta R.T.: 0.000 min
Response: 468893
Conc: 89.09 ng/ml



#27 AR-1254-2

R.T.: 6.564 min
Delta R.T.: 0.000 min
Response: 147710
Conc: 107.82 ng/ml




#28 AR-1254-3

R.T.: 7.688 min
Delta R.T.: -0.001 min
Response: 433518
Conc: 77.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
JEH475K-1-1


Manual Integrations
APPROVED

Ankita
3/8/2021 12:10:58 PM




#28 AR-1254-3

R.T.: 6.986 min
Delta R.T.: 0.000 min
Response: 199850
Conc: 93.21 ng/ml




#29 AR-1254-4

R.T.: 7.984 min
Delta R.T.: 0.000 min
Response: 365967
Conc: 95.39 ng/ml



#29 AR-1254-4

R.T.: 7.222 min
Delta R.T.: 0.000 min
Response: 135540
Conc: 122.67 ng/ml




#30 AR-1254-5

R.T.: 8.412 min
Delta R.T.: 0.000 min
Response: 306667
Conc: 73.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
JEH475K-1-1

Manual Integrations
APPROVED

Ankita
3/8/2021 12:10:58 PM



#30 AR-1254-5

R.T.: 7.651 min
Delta R.T.: 0.000 min
Response: 153907
Conc: 89.83 ng/ml