

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030521\
 Data File : PP033716.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 05 Mar 2021 20:05
 Operator : DD\AJ
 Sample : M1573-02 10X
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 JEH475K-1-2

Manual Integrations
 APPROVED

Ankita
 3/8/2021 12:11:00 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 06 00:04:56 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Mar 05 04:37:34 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.860	4.273	168246	87254	1.465	2.192 #
2) SA Decachlor...	10.796	9.617	197668	54381	2.184m	2.328
Target Compounds						
26) L6 AR-1254-1	7.079	6.406	531182	342338	152.687	217.534 #
27) L6 AR-1254-2	7.309	6.564	798988	261727	151.800	191.043 #
28) L6 AR-1254-3	7.688	6.985	748375	361750	134.092	168.716 #
29) L6 AR-1254-4	7.984	7.221	611903	241625	159.488	218.673m#
30) L6 AR-1254-5	8.411	7.650	558958	280537	133.328	163.737

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030521\
Data File : PP033716.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 05 Mar 2021 20:05
Operator : DD\AJ
Sample : M1573-02 10X
Misc :
ALS Vial : 27 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampled :
JEH475K-1-2


Manual Integrations
APPROVED

Ankita
3/8/2021 12:11:00 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 06 00:04:56 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030421.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 05 04:37:34 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.860 min
Delta R.T.: 0.001 min
Response: 168246
Conc: 1.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
JEH475K-1-2


Manual Integrations
APPROVED

Ankita
3/8/2021 12:11:00 PM




#1 Tetrachloro-m-xylene

R.T.: 4.273 min
Delta R.T.: 0.002 min
Response: 87254
Conc: 2.19 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.796 min
Delta R.T.: 0.005 min
Response: 197668
Conc: 2.18 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.617 min
Delta R.T.: -0.001 min
Response: 54381
Conc: 2.33 ng/ml




#26 AR-1254-1

R.T.: 7.079 min
Delta R.T.: 0.000 min
Response: 531182
Conc: 152.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
JEH475K-1-2


Manual Integrations
APPROVED

Ankita
3/8/2021 12:11:00 PM




#26 AR-1254-1

R.T.: 6.406 min
Delta R.T.: 0.000 min
Response: 342338
Conc: 217.53 ng/ml




#27 AR-1254-2

R.T.: 7.309 min
Delta R.T.: 0.000 min
Response: 798988
Conc: 151.80 ng/ml



#27 AR-1254-2

R.T.: 6.564 min
Delta R.T.: 0.000 min
Response: 261727
Conc: 191.04 ng/ml




#28 AR-1254-3

R.T.: 7.688 min
Delta R.T.: -0.001 min
Response: 748375
Conc: 134.09 ng/ml

Instrument :
ECD_P
ClientSampleId :
JEH475K-1-2


Manual Integrations
APPROVED

Ankita
3/8/2021 12:11:00 PM




#28 AR-1254-3

R.T.: 6.985 min
Delta R.T.: 0.000 min
Response: 361750
Conc: 168.72 ng/ml




#29 AR-1254-4

R.T.: 7.984 min
Delta R.T.: 0.000 min
Response: 611903
Conc: 159.49 ng/ml



#29 AR-1254-4

R.T.: 7.221 min
Delta R.T.: -0.001 min
Response: 241625
Conc: 218.67 ng/ml m




#30 AR-1254-5

R.T.: 8.411 min
Delta R.T.: 0.000 min
Response: 558958
Conc: 133.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
JEH475K-1-2

Manual Integrations
APPROVED

Ankita
3/8/2021 12:11:00 PM



#30 AR-1254-5

R.T.: 7.650 min
Delta R.T.: 0.000 min
Response: 280537
Conc: 163.74 ng/ml