

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030623\
 Data File : PP056230.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Mar 2023 16:23
 Operator : YP\AJ
 Sample : 01787-02
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 06 17:13:38 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 20 05:24:03 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.322	3.564	35140000	26009909	21.809	17.813
2)	SA Decachlor...	10.067	8.564	28034930	19498088	13.743	17.187 #
Target Compounds							
3)	L1 AR-1016-1	5.497	4.660	90533	170742	1.517	3.792 #
4)	L1 AR-1016-2	5.519	4.660	35042	170742	0.401	2.629 #
5)	L1 AR-1016-3	5.573	4.839	167764	49503	3.020	1.424 #
6)	L1 AR-1016-4	5.678	4.889	30151	72597	0.677	2.429 #
7)	L1 AR-1016-5	5.977	5.081f	40367	114897	0.836	2.906 #
8)	L2 AR-1221-1	4.527	3.787	96625	21811	4.642	1.214 #
9)	L2 AR-1221-2	4.628	3.863	573905	542013	36.079	41.270
10)	L2 AR-1221-3	4.697	3.940	46399	70864	0.956	1.698 #
11)	L3 AR-1232-1	4.697	3.940	46399	70864	1.171	2.058 #
12)	L3 AR-1232-2	5.226	4.660	74825	170742	3.602	5.817 #
13)	L3 AR-1232-3	5.519	4.839	35042	49503	0.889	3.257 #
14)	L3 AR-1232-4	5.678	4.917	30151	30403	1.539	2.006 #
15)	L3 AR-1232-5	5.771	5.081f	20029	114897	1.252	7.010 #
16)	L4 AR-1242-1	5.497	4.660	90533	170742	1.875	4.641 #
17)	L4 AR-1242-2	5.519	4.660	35042	170742	0.493	3.242 #
18)	L4 AR-1242-3	5.573	4.839	167764	49503	3.760	1.749 #
19)	L4 AR-1242-4	5.678	4.917	30151	30403	0.835	0.985
20)	L4 AR-1242-5	6.423	5.441	80301	231919	1.917	6.963 #
21)	L5 AR-1248-1	5.497	4.660	90533	170742	2.488	6.114 #
22)	L5 AR-1248-2	5.771	4.889	20029	72597	0.348	1.707 #
23)	L5 AR-1248-3	5.977	4.917	40367	30403	0.621	0.683
24)	L5 AR-1248-4	6.392	5.081f	243348	114897	3.354	2.244 #
25)	L5 AR-1248-5	6.423	5.497	80301	74639	1.133	1.693 #
26)	L6 AR-1254-1	6.358	5.441	105827	231919	1.439	3.152 #
27)	L6 AR-1254-2	6.569	5.622f	8634	340031	0.075	5.404 #
28)	L6 AR-1254-3	6.938	0.000	66713	0	0.542	N.D. #
29)	L6 AR-1254-4	7.226	0.000	117352	0	1.278	N.D. #
30)	L6 AR-1254-5	7.662	0.000	573461	0	5.456	N.D. #
31)	L7 AR-1260-1	7.105	0.000	76002	0	0.734	N.D. #
32)	L7 AR-1260-2	7.365	0.000	175302	0	1.394	N.D. #
33)	L7 AR-1260-3	7.708f	6.471	39471	48952	0.449	0.579 #
34)	L7 AR-1260-4	7.959	0.000	288852	0	2.688	N.D. #
35)	L7 AR-1260-5	8.270	7.207	70490	469460	0.341	3.468 #
36)	L8 AR-1262-1	7.708f	0.000	39471	0	0.335	N.D. #
37)	L8 AR-1262-2	8.270	0.000	70490	0	0.321	N.D. #
38)	L8 AR-1262-3	8.574	7.470	31280	74881	0.205	1.350 #
39)	L8 AR-1262-4	8.665	7.539	265117	158643	3.522	1.563 #
40)	L8 AR-1262-5	9.324	0.000	50061	0	0.607	N.D. #
41)	L9 AR-1268-1	8.574	7.470	31280	74881	0.110	0.445 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030623\
Data File : PP056230.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Mar 2023 16:23
Operator : YP\AJ
Sample : 01787-02
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 06 17:13:38 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021723.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 20 05:24:03 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	8.665	7.539	265117	158643	0.994	1.037
43)	L9 AR-1268-3	8.896	7.771f	82899	116270	0.352	0.847 #
44)	L9 AR-1268-4	9.316	0.000	74778	0	0.757	N.D. #
45)	L9 AR-1268-5	9.732	8.317	79813	108464	0.106	0.269 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030623\
Data File : PP056230.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Mar 2023 16:23
Operator : YP\AJ
Sample : 01787-02
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 06 17:13:38 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021723.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 20 05:24:03 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.322 min
Delta R.T.: -0.003 min
Response: 35140000
Conc: 21.81 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.564 min
Delta R.T.: -0.004 min
Response: 26009909
Conc: 17.81 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.067 min
Delta R.T.: -0.009 min
Response: 28034930
Conc: 13.74 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.564 min
Delta R.T.: -0.013 min
Response: 19498088
Conc: 17.19 ng/ml



#3 AR-1016-1


R.T.: 5.497 min
Delta R.T.: 0.000 min
Response: 90533
Conc: 1.52 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.660 min
Delta R.T.: 0.007 min
Response: 170742
Conc: 3.79 ng/ml




#4 AR-1016-2

R.T.: 5.519 min
Delta R.T.: 0.000 min
Response: 35042
Conc: 0.40 ng/ml



#4 AR-1016-2


R.T.: 4.660 min
Delta R.T.: -0.011 min
Response: 170742
Conc: 2.63 ng/ml



#5 AR-1016-3


R.T.: 5.573 min
Delta R.T.: -0.009 min
Response: 167764
Conc: 3.02 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.839 min
Delta R.T.: -0.007 min
Response: 49503
Conc: 1.42 ng/ml




#6 AR-1016-4

R.T.: 5.678 min
Delta R.T.: -0.002 min
Response: 30151
Conc: 0.68 ng/ml



#6 AR-1016-4


R.T.: 4.889 min
Delta R.T.: 0.000 min
Response: 72597
Conc: 2.43 ng/ml



#7 AR-1016-5


R.T.: 5.977 min
Delta R.T.: 0.000 min
Response: 40367
Conc: 0.84 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.081 min
Delta R.T.: -0.021 min
Response: 114897
Conc: 2.91 ng/ml




#8 AR-1221-1

R.T.: 4.527 min
Delta R.T.: -0.007 min
Response: 96625
Conc: 4.64 ng/ml



#8 AR-1221-1


R.T.: 3.787 min
Delta R.T.: 0.005 min
Response: 21811
Conc: 1.21 ng/ml



#9 AR-1221-2


R.T.: 4.628 min
Delta R.T.: 0.007 min
Response: 573905
Conc: 36.08 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.863 min
Delta R.T.: -0.005 min
Response: 542013
Conc: 41.27 ng/ml




#10 AR-1221-3

R.T.: 4.697 min
Delta R.T.: 0.000 min
Response: 46399
Conc: 0.96 ng/ml



#10 AR-1221-3


R.T.: 3.940 min
Delta R.T.: -0.003 min
Response: 70864
Conc: 1.70 ng/ml



#11 AR-1232-1


R.T.: 4.697 min
Delta R.T.: 0.000 min
Response: 46399
Conc: 1.17 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 3.940 min
Delta R.T.: -0.003 min
Response: 70864
Conc: 2.06 ng/ml




#12 AR-1232-2

R.T.: 5.226 min
Delta R.T.: -0.003 min
Response: 74825
Conc: 3.60 ng/ml



#12 AR-1232-2


R.T.: 4.660 min
Delta R.T.: -0.011 min
Response: 170742
Conc: 5.82 ng/ml



#13 AR-1232-3


R.T.: 5.519 min
Delta R.T.: -0.001 min
Response: 35042
Conc: 0.89 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.839 min
Delta R.T.: -0.007 min
Response: 49503
Conc: 3.26 ng/ml




#14 AR-1232-4

R.T.: 5.678 min
Delta R.T.: -0.003 min
Response: 30151
Conc: 1.54 ng/ml



#14 AR-1232-4


R.T.: 4.917 min
Delta R.T.: -0.014 min
Response: 30403
Conc: 2.01 ng/ml



#15 AR-1232-5


R.T.: 5.771 min
Delta R.T.: -0.001 min
Response: 20029
Conc: 1.25 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.081 min
Delta R.T.: -0.022 min
Response: 114897
Conc: 7.01 ng/ml




#16 AR-1242-1

R.T.: 5.497 min
Delta R.T.: 0.000 min
Response: 90533
Conc: 1.87 ng/ml



#16 AR-1242-1


R.T.: 4.660 min
Delta R.T.: 0.008 min
Response: 170742
Conc: 4.64 ng/ml



#17 AR-1242-2


R.T.: 5.519 min
Delta R.T.: -0.001 min
Response: 35042
Conc: 0.49 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.660 min
Delta R.T.: -0.011 min
Response: 170742
Conc: 3.24 ng/ml




#18 AR-1242-3

R.T.: 5.573 min
Delta R.T.: -0.009 min
Response: 167764
Conc: 3.76 ng/ml



#18 AR-1242-3


R.T.: 4.839 min
Delta R.T.: -0.007 min
Response: 49503
Conc: 1.75 ng/ml



#19 AR-1242-4


R.T.: 5.678 min
Delta R.T.: -0.002 min
Response: 30151
Conc: 0.83 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.917 min
Delta R.T.: -0.014 min
Response: 30403
Conc: 0.99 ng/ml




#20 AR-1242-5

R.T.: 6.423 min
Delta R.T.: 0.003 min
Response: 80301
Conc: 1.92 ng/ml



#20 AR-1242-5


R.T.: 5.441 min
Delta R.T.: -0.013 min
Response: 231919
Conc: 6.96 ng/ml



#21 AR-1248-1


R.T.: 5.497 min
Delta R.T.: 0.000 min
Response: 90533
Conc: 2.49 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.660 min
Delta R.T.: 0.008 min
Response: 170742
Conc: 6.11 ng/ml




#22 AR-1248-2

R.T.: 5.771 min
Delta R.T.: 0.000 min
Response: 20029
Conc: 0.35 ng/ml



#22 AR-1248-2


R.T.: 4.889 min
Delta R.T.: 0.000 min
Response: 72597
Conc: 1.71 ng/ml



#23 AR-1248-3


R.T.: 5.977 min
Delta R.T.: 0.001 min
Response: 40367
Conc: 0.62 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.917 min
Delta R.T.: -0.013 min
Response: 30403
Conc: 0.68 ng/ml




#24 AR-1248-4

R.T.: 6.392 min
Delta R.T.: 0.011 min
Response: 243348
Conc: 3.35 ng/ml



#24 AR-1248-4


R.T.: 5.081 min
Delta R.T.: -0.021 min
Response: 114897
Conc: 2.24 ng/ml



#25 AR-1248-5


R.T.: 6.423 min
Delta R.T.: 0.003 min
Response: 80301
Conc: 1.13 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.497 min
Delta R.T.: 0.003 min
Response: 74639
Conc: 1.69 ng/ml




#26 AR-1254-1

R.T.: 6.358 min
Delta R.T.: 0.004 min
Response: 105827
Conc: 1.44 ng/ml



#26 AR-1254-1


R.T.: 5.441 min
Delta R.T.: -0.012 min
Response: 231919
Conc: 3.15 ng/ml



#27 AR-1254-2


R.T.: 6.569 min
Delta R.T.: -0.006 min
Response: 8634
Conc: 0.07 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.622 min
Delta R.T.: 0.020 min
Response: 340031
Conc: 5.40 ng/ml




#28 AR-1254-3

R.T.: 6.938 min
Delta R.T.: -0.003 min
Response: 66713
Conc: 0.54 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T.: 6.006 min
Response: 0
Conc: N.D.



#29 AR-1254-4


R.T.: 7.226 min
Delta R.T.: -0.003 min
Response: 117352
Conc: 1.28 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 0.000 min
Exp R.T. : 6.236 min
Response: 0
Conc: N.D.




#30 AR-1254-5

R.T.: 7.662 min
Delta R.T.: 0.013 min
Response: 573461
Conc: 5.46 ng/ml



#30 AR-1254-5


R.T.: 0.000 min
Exp R.T. : 6.655 min
Response: 0
Conc: N.D.



#31 AR-1260-1


R.T.: 7.105 min
Delta R.T.: -0.003 min
Response: 76002
Conc: 0.73 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 0.000 min
Exp R.T. : 6.138 min
Response: 0
Conc: N.D.




#32 AR-1260-2

R.T.: 7.365 min
Delta R.T.: 0.000 min
Response: 175302
Conc: 1.39 ng/ml



#32 AR-1260-2


R.T.: 0.000 min
Exp R.T. : 6.327 min
Response: 0
Conc: N.D.



#33 AR-1260-3


R.T.: 7.708 min
Delta R.T.: -0.019 min
Response: 39471
Conc: 0.45 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.471 min
Delta R.T.: -0.009 min
Response: 48952
Conc: 0.58 ng/ml




#34 AR-1260-4

R.T.: 7.959 min
Delta R.T.: 0.006 min
Response: 288852
Conc: 2.69 ng/ml



#34 AR-1260-4


R.T.: 0.000 min
Exp R.T.: 6.954 min
Response: 0
Conc: N.D.



#35 AR-1260-5


R.T.: 8.270 min
Delta R.T.: 0.000 min
Response: 70490
Conc: 0.34 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.207 min
Delta R.T.: 0.011 min
Response: 469460
Conc: 3.47 ng/ml




#36 AR-1262-1

R.T.: 7.708 min
Delta R.T.: -0.016 min
Response: 39471
Conc: 0.33 ng/ml



#36 AR-1262-1


R.T.: 0.000 min
Exp R.T.: 6.745 min
Response: 0
Conc: N.D.



#37 AR-1262-2


R.T.: 8.270 min
Delta R.T.: 0.002 min
Response: 70490
Conc: 0.32 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 0.000 min
Exp R.T. : 6.950 min
Response: 0
Conc: N.D.




#38 AR-1262-3

R.T.: 8.574 min
Delta R.T.: -0.004 min
Response: 31280
Conc: 0.20 ng/ml



#38 AR-1262-3


R.T.: 7.470 min
Delta R.T.: -0.007 min
Response: 74881
Conc: 1.35 ng/ml



#39 AR-1262-4


R.T.: 8.665 min
Delta R.T.: 0.000 min
Response: 265117
Conc: 3.52 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.539 min
Delta R.T.: -0.003 min
Response: 158643
Conc: 1.56 ng/ml




#40 AR-1262-5

R.T.: 9.324 min
Delta R.T.: 0.002 min
Response: 50061
Conc: 0.61 ng/ml



#40 AR-1262-5


R.T.: 0.000 min
Exp R.T. : 8.038 min
Response: 0
Conc: N.D.



#41 AR-1268-1


R.T.: 8.574 min
Delta R.T.: -0.002 min
Response: 31280
Conc: 0.11 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.470 min
Delta R.T.: -0.007 min
Response: 74881
Conc: 0.45 ng/ml




#42 AR-1268-2

R.T.: 8.665 min
Delta R.T.: -0.004 min
Response: 265117
Conc: 0.99 ng/ml



#42 AR-1268-2


R.T.: 7.539 min
Delta R.T.: -0.004 min
Response: 158643
Conc: 1.04 ng/ml



#43 AR-1268-3


R.T.: 8.896 min
Delta R.T.: 0.000 min
Response: 82899
Conc: 0.35 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.771 min
Delta R.T.: 0.023 min
Response: 116270
Conc: 0.85 ng/ml




#44 AR-1268-4

R.T.: 9.316 min
Delta R.T.: -0.003 min
Response: 74778
Conc: 0.76 ng/ml



#44 AR-1268-4


R.T.: 0.000 min
Exp R.T. : 8.037 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 9.732 min
Delta R.T.: -0.001 min
Response: 79813
Conc: 0.11 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.317 min
Delta R.T.: -0.007 min
Response: 108464
Conc: 0.27 ng/ml