

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030722\
 Data File : PP044298.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Mar 2022 18:42
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 08 01:30:24 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 03 02:21:53 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.912	3.167	112.6E6	97220068	54.321	55.572
2) SA Decachlor...	10.050	8.534	66778967	77745308	48.519	50.682
Target Compounds						
26) L6 AR-1254-1	6.090	5.164	35284191	42124852	470.469	509.532
27) L6 AR-1254-2	6.330	5.325	52555882	36529273	467.730	507.784
28) L6 AR-1254-3	6.729	5.757	54570320	58429083	466.348	507.141
29) L6 AR-1254-4	7.044	6.007	38176237	38186267	474.822	507.235
30) L6 AR-1254-5	7.505	6.458	40345325	56089805	463.497	514.678



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030722\
Data File : PP044298.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Mar 2022 18:42
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 08 01:30:24 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 03 02:21:53 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.912 min
Delta R.T.: 0.000 min
Response: 112551839
Conc: 54.32 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.167 min
Delta R.T.: 0.000 min
Response: 97220068
Conc: 55.57 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.050 min
Delta R.T.: -0.002 min
Response: 66778967
Conc: 48.52 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.534 min
Delta R.T.: -0.003 min
Response: 77745308
Conc: 50.68 ng/ml



#26 AR-1254-1


R.T.: 6.090 min
Delta R.T.: -0.001 min
Response: 35284191
Conc: 470.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.164 min
Delta R.T.: -0.002 min
Response: 42124852
Conc: 509.53 ng/ml




#27 AR-1254-2

R.T.: 6.330 min
Delta R.T.: -0.001 min
Response: 52555882
Conc: 467.73 ng/ml



#27 AR-1254-2


R.T.: 5.325 min
Delta R.T.: -0.002 min
Response: 36529273
Conc: 507.78 ng/ml



#28 AR-1254-3


R.T.: 6.729 min
Delta R.T.: -0.001 min
Response: 54570320
Conc: 466.35 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 5.757 min
Delta R.T.: -0.002 min
Response: 58429083
Conc: 507.14 ng/ml




#29 AR-1254-4

R.T.: 7.044 min
Delta R.T.: -0.001 min
Response: 38176237
Conc: 474.82 ng/ml



#29 AR-1254-4


R.T.: 6.007 min
Delta R.T.: -0.002 min
Response: 38186267
Conc: 507.23 ng/ml



#30 AR-1254-5

R.T.: 7.505 min
Delta R.T.: 0.000 min
Response: 40345325
Conc: 463.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.458 min
Delta R.T.: -0.002 min
Response: 56089805
Conc: 514.68 ng/ml