

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_P\Data\PP030722\  
 Data File : PP044304.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 07 Mar 2022 20:13  
 Operator : YP\AJ  
 Sample : N1804-05 10X  
 Misc :  
 ALS Vial : 21 Sample Multiplier: 1

Instrument :  
 ECD\_P  
 ClientSampleId :  
 SOM-2022-9A

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Mar 08 01:33:16 2022  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_P\methods\PP030222.M  
 Quant Title : GC EXTRACTABLES  
 QLast Update : Thu Mar 03 02:21:53 2022  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 2 µl  
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

| Compound                    | RT#1   | RT#2  | Resp#1  | Resp#2  | ng/ml  | ng/ml   |
|-----------------------------|--------|-------|---------|---------|--------|---------|
| -----                       |        |       |         |         |        |         |
| System Monitoring Compounds |        |       |         |         |        |         |
| 1) SA Tetrachlo...          | 3.910  | 3.167 | 2268230 | 3355542 | 1.095m | 1.918 # |
| 2) SA Decachlor...          | 10.051 | 8.538 | 2335190 | 2661310 | 1.697  | 1.735   |

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_P\Data\PP030722\  
Data File : PP044304.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 07 Mar 2022 20:13  
Operator : YP\AJ  
Sample : N1804-05 10X  
Misc :  
ALS Vial : 21 Sample Multiplier: 1

Instrument :  
ECD\_P  
ClientSampleId :  
SOM-2022-9A

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Mar 08 01:33:16 2022  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_P\methods\PP030222.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Thu Mar 03 02:21:53 2022  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.910 min  
Delta R.T.: -0.002 min  
Response: 2268230  
Conc: 1.09 ng/ml

Instrument :  
ECD\_P  
ClientSampleId :  
SOM-2022-9A




#1 Tetrachloro-m-xylene

R.T.: 3.167 min  
Delta R.T.: 0.000 min  
Response: 3355542  
Conc: 1.92 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.051 min  
Delta R.T.: 0.000 min  
Response: 2335190  
Conc: 1.70 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.538 min  
Delta R.T.: 0.000 min  
Response: 2661310  
Conc: 1.73 ng/ml