

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030821\
 Data File : PP033725.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 08 Mar 2021 13:34
 Operator : DD\AJ
 Sample : M1580-02
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 03-05B

Manual Integrations
 APPROVED

Ankita
 3/9/2021 1:05:23 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 09 03:54:14 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Mar 05 04:37:34 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.858	4.271	1804226	706616	15.707	17.748
2) SA Decachlor...	10.784	9.604	1742197	515180	19.253	22.053m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030821\
Data File : PP033725.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 08 Mar 2021 13:34
Operator : DD\AJ
Sample : M1580-02
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
03-05B


Manual Integrations
APPROVED

Ankita
3/9/2021 1:05:23 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 09 03:54:14 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030421.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 05 04:37:34 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.858 min
Delta R.T.: -0.001 min
Response: 1804226
Conc: 15.71 ng/ml

Instrument :
ECD_P
ClientSampleId :
03-05B


Manual Integrations
APPROVED

Ankita
3/9/2021 1:05:23 PM




#1 Tetrachloro-m-xylene

R.T.: 4.271 min
Delta R.T.: 0.000 min
Response: 706616
Conc: 17.75 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.784 min
Delta R.T.: -0.008 min
Response: 1742197
Conc: 19.25 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.604 min
Delta R.T.: -0.014 min
Response: 515180
Conc: 22.05 ng/ml m