

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_P\Data\PP030921\  
 Data File : PP033793.D  
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
 Acq On : 10 Mar 2021 5:12  
 Operator : DD\AJ  
 Sample : M1458-07  
 Misc : AR1262 LOD 40 PPB  
 ALS Vial : 49 Sample Multiplier: 1

Instrument :  
 ECD\_P  
 ClientSampleId :  
 LOD-MDL-WATER-01-QT1-2021

Manual Integrations  
 APPROVED

Ankita  
 3/10/2021 2:01:25 PM

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Mar 10 06:03:20 2021  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_P\methods\PP030921.M  
 Quant Title : GC EXTRACTABLES  
 QLast Update : Wed Mar 10 03:01:34 2021  
 Response via : Initial Calibration  
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl  
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

| Compound                    | RT#1   | RT#2  | Resp#1  | Resp#2 | ng/ml   | ng/ml    |
|-----------------------------|--------|-------|---------|--------|---------|----------|
| -----                       |        |       |         |        |         |          |
| System Monitoring Compounds |        |       |         |        |         |          |
| 1) SA Tetrachlo...          | 4.856  | 4.269 | 1768848 | 832456 | 20.443  | 20.491   |
| 2) SA Decachlor...          | 10.781 | 9.597 | 1778780 | 591703 | 21.062  | 22.173   |
| Target Compounds            |        |       |         |        |         |          |
| 36) L8 AR-1262-1            | 8.478  | 7.639 | 189599  | 40712  | 39.847m | 38.298   |
| 37) L8 AR-1262-2            | 9.027  | 8.186 | 338992  | 149915 | 39.220  | 44.990   |
| 38) L8 AR-1262-3            | 9.329  | 8.473 | 208768  | 68611  | 34.400  | 48.043 # |
| 39) L8 AR-1262-4            | 9.421  | 8.536 | 168788  | 100477 | 34.052  | 39.853   |
| 40) L8 AR-1262-5            | 10.063 | 9.034 | 129677  | 43302  | 39.399m | 37.989   |
| -----                       |        |       |         |        |         |          |

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_P\Data\PP030921\  
Data File : PP033793.D  
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
Acq On : 10 Mar 2021 5:12  
Operator : DD\AJ  
Sample : M1458-07  
Misc : AR1262 LOD 40 PPB  
ALS Vial : 49 Sample Multiplier: 1

Instrument :  
ECD\_P  
ClientSampleId :  
LOD-MDL-WATER-01-QT1-2021

Manual Integrations  
APPROVED

Ankita  
3/10/2021 2:01:25 PM

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Mar 10 06:03:20 2021  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_P\methods\PP030921.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Wed Mar 10 03:01:34 2021  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

