

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030921\
 Data File : PP033745.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 09 Mar 2021 15:42
 Operator : DD\AJ
 Sample : AR1221ICC500
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1221ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 10 01:40:24 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030921.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 10 01:40:06 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.860	4.271	4174824	1932707	50.000	50.000
2) SA Decachlor...	10.788	9.604	3981895	1240466	50.000	50.000
Target Compounds						
8) L2 AR-1221-1	5.109	4.529	461398	212908	500.000	500.000
9) L2 AR-1221-2	5.207	4.629	351339	161789	500.000	500.000
10) L2 AR-1221-3	5.293	4.718	1052409	490107	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030921\
Data File : PP033745.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 09 Mar 2021 15:42
Operator : DD\AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 10 01:40:24 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030921.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 10 01:40:06 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.860 min
Delta R.T.: 0.000 min
Response: 4174824
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500




#1 Tetrachloro-m-xylene

R.T.: 4.271 min
Delta R.T.: 0.000 min
Response: 1932707
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.788 min
Delta R.T.: 0.000 min
Response: 3981895
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.604 min
Delta R.T.: 0.000 min
Response: 1240466
Conc: 50.00 ng/ml



#8 AR-1221-1


R.T.: 5.109 min
Delta R.T.: 0.000 min
Response: 461398
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500




#8 AR-1221-1

R.T.: 4.529 min
Delta R.T.: 0.000 min
Response: 212908
Conc: 500.00 ng/ml




#9 AR-1221-2

R.T.: 5.207 min
Delta R.T.: 0.000 min
Response: 351339
Conc: 500.00 ng/ml



#9 AR-1221-2


R.T.: 4.629 min
Delta R.T.: 0.000 min
Response: 161789
Conc: 500.00 ng/ml



#10 AR-1221-3

R.T.: 5.293 min
Delta R.T.: 0.000 min
Response: 1052409
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500



#10 AR-1221-3

R.T.: 4.718 min
Delta R.T.: 0.000 min
Response: 490107
Conc: 500.00 ng/ml