

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030921\
 Data File : PP033752.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 09 Mar 2021 17:40
 Operator : DD\AJ
 Sample : PP030921ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP030921

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 10 03:08:07 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030921.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 10 03:01:34 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.858	4.272	4364741	2005716	50.445	49.371
2) SA Decachlor...	10.787	9.604	4156699	1284990	49.218	48.153
Target Compounds						
3) L1 AR-1016-1	6.171	5.530	1316993	559992	494.341	492.020
4) L1 AR-1016-2	6.194	5.551	1992808	811174	493.045	490.716
5) L1 AR-1016-3	6.260	5.744	1260511	454700	509.383	496.229
6) L1 AR-1016-4	6.366	5.793	1035096	364877	507.783	505.292
7) L1 AR-1016-5	6.680	6.024	1028515	467290	507.403	505.400
31) L7 AR-1260-1	7.852	7.116	1788533	748303	520.818	489.640
32) L7 AR-1260-2	8.118	7.310	2095203	857465	508.885	474.064
33) L7 AR-1260-3	8.481	7.467	1661452	833563	513.305	492.189
34) L7 AR-1260-4	8.711	7.947	1889570	686757	510.119	466.959
35) L7 AR-1260-5	9.031	8.190	3760016	1558933	486.028	479.480



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030921\
Data File : PP033752.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 09 Mar 2021 17:40
Operator : DD\AJ
Sample : PP030921ICV500
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :
ICVPP030921

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 10 03:08:07 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030921.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 10 03:01:34 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.858 min
Delta R.T.: 0.000 min
Response: 4364741
Conc: 50.44 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP030921




#1 Tetrachloro-m-xylene

R.T.: 4.272 min
Delta R.T.: 0.000 min
Response: 2005716
Conc: 49.37 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.787 min
Delta R.T.: 0.000 min
Response: 4156699
Conc: 49.22 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.604 min
Delta R.T.: 0.000 min
Response: 1284990
Conc: 48.15 ng/ml



#3 AR-1016-1


R.T.: 6.171 min
Delta R.T.: 0.000 min
Response: 1316993
Conc: 494.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP030921




#3 AR-1016-1

R.T.: 5.530 min
Delta R.T.: 0.000 min
Response: 559992
Conc: 492.02 ng/ml




#4 AR-1016-2

R.T.: 6.194 min
Delta R.T.: 0.000 min
Response: 1992808
Conc: 493.05 ng/ml



#4 AR-1016-2


R.T.: 5.551 min
Delta R.T.: 0.000 min
Response: 811174
Conc: 490.72 ng/ml



#5 AR-1016-3


R.T.: 6.260 min
Delta R.T.: 0.000 min
Response: 1260511
Conc: 509.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP030921




#5 AR-1016-3

R.T.: 5.744 min
Delta R.T.: 0.000 min
Response: 454700
Conc: 496.23 ng/ml




#6 AR-1016-4

R.T.: 6.366 min
Delta R.T.: 0.000 min
Response: 1035096
Conc: 507.78 ng/ml



#6 AR-1016-4


R.T.: 5.793 min
Delta R.T.: 0.000 min
Response: 364877
Conc: 505.29 ng/ml



#7 AR-1016-5


R.T.: 6.680 min
Delta R.T.: 0.000 min
Response: 1028515
Conc: 507.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP030921




#7 AR-1016-5

R.T.: 6.024 min
Delta R.T.: 0.000 min
Response: 467290
Conc: 505.40 ng/ml




#31 AR-1260-1

R.T.: 7.852 min
Delta R.T.: 0.000 min
Response: 1788533
Conc: 520.82 ng/ml



#31 AR-1260-1


R.T.: 7.116 min
Delta R.T.: 0.000 min
Response: 748303
Conc: 489.64 ng/ml



#32 AR-1260-2


R.T.: 8.118 min
Delta R.T.: 0.000 min
Response: 2095203
Conc: 508.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP030921




#32 AR-1260-2

R.T.: 7.310 min
Delta R.T.: 0.000 min
Response: 857465
Conc: 474.06 ng/ml




#33 AR-1260-3

R.T.: 8.481 min
Delta R.T.: -0.002 min
Response: 1661452
Conc: 513.31 ng/ml



#33 AR-1260-3


R.T.: 7.467 min
Delta R.T.: 0.000 min
Response: 833563
Conc: 492.19 ng/ml



#34 AR-1260-4


R.T.: 8.711 min
Delta R.T.: 0.000 min
Response: 1889570
Conc: 510.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP030921




#34 AR-1260-4

R.T.: 7.947 min
Delta R.T.: 0.000 min
Response: 686757
Conc: 466.96 ng/ml



#35 AR-1260-5

R.T.: 9.031 min
Delta R.T.: 0.000 min
Response: 3760016
Conc: 486.03 ng/ml



#35 AR-1260-5

R.T.: 8.190 min
Delta R.T.: 0.000 min
Response: 1558933
Conc: 479.48 ng/ml