

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030922\
 Data File : PP044382.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Mar 2022 10:07
 Operator : YP\AJ
 Sample : N1827-08 10X
 Misc :
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 09 13:46:48 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 03 02:21:53 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.910	3.166	2830474	3014191	1.366	1.723 #
2)	SA Decachlor...	10.047	8.531	2649452	3094275	1.925	2.017
Target Compounds							
3)	L1 AR-1016-1	5.158	4.309	59250	19273	0.873	0.347 #
4)	L1 AR-1016-2	5.195	4.333	41396	36989	0.409	0.477
5)	L1 AR-1016-3	5.264	4.513	124557	5173	1.986	0.121 #
6)	L1 AR-1016-4	5.360	4.557	36695	27048	0.707	0.802
7)	L1 AR-1016-5	5.670	4.792	73028	26863	1.450	0.646 #
8)	L2 AR-1221-1	4.132	3.387	87846	16914	3.555	0.737 #
9)	L2 AR-1221-2	4.236	3.479	86168	36945	4.593	2.139 #
10)	L2 AR-1221-3	4.304	3.561	226244	17989	4.043	0.362 #
11)	L3 AR-1232-1	4.304	3.561	226244	17989	4.930	0.433 #
12)	L3 AR-1232-2	4.880	4.333	66237	36989	2.943	1.071 #
13)	L3 AR-1232-3	5.195	4.513	41396	5173	0.957	0.277 #
14)	L3 AR-1232-4	5.360	4.610	36695	15833	1.828	0.960 #
15)	L3 AR-1232-5	5.466	4.787	175518	24225	11.569	1.338 #
16)	L4 AR-1242-1	5.158	4.309	59250	19273	1.088	0.420 #
17)	L4 AR-1242-2	5.195	4.333	41396	36989	0.520	0.585
18)	L4 AR-1242-3	5.264	4.513	124557	5173	2.638	0.145 #
19)	L4 AR-1242-4	5.360	4.610	36695	15833	0.948	0.458 #
20)	L4 AR-1242-5	6.160	5.169	348875	22947	8.183	0.543 #
21)	L5 AR-1248-1	5.158	4.309	59250	19273	1.433	0.543 #
22)	L5 AR-1248-2	5.466	4.557	175518	27048	3.129	0.575 #
23)	L5 AR-1248-3	5.670	4.610	73028	15833	1.069	0.321 #
24)	L5 AR-1248-4	6.124	4.787	96300	24225	1.321	0.429 #
25)	L5 AR-1248-5	6.160	5.226f	348875	51273	4.750	0.914 #
26)	L6 AR-1254-1	6.089	5.169	85894	22947	1.145	0.278 #
27)	L6 AR-1254-2	6.329	5.334	132282	30372	1.177	0.422 #
28)	L6 AR-1254-3	6.728	5.758	303697	19432	2.595	0.169 #
29)	L6 AR-1254-4	7.027f	6.006	72890	27811	0.907	0.369 #
30)	L6 AR-1254-5	7.508	6.458	319176	38808	3.667	0.356 #
31)	L7 AR-1260-1	6.917	5.904	118941	26864	1.353	0.335 #
32)	L7 AR-1260-2	7.185	6.108	232954	22712	2.425	0.231 #
33)	L7 AR-1260-3	7.586	6.263	301818	94071	4.652	1.014 #
34)	L7 AR-1260-4	7.844	6.780	618597	17763	7.931	0.249 #
35)	L7 AR-1260-5	8.191	7.047	666380	10621	4.755	0.064 #
36)	L8 AR-1262-1	7.586	6.498	301818	29393	2.986	0.283 #
37)	L8 AR-1262-2	8.191	6.780	666380	17763	4.013	0.186 #
38)	L8 AR-1262-3	8.509	7.348	94584	37018	0.845	0.469 #
39)	L8 AR-1262-4	8.602	7.425	45230	40592	0.911	0.281 #
40)	L8 AR-1262-5	9.293	7.970	79146	30105	1.387	0.420 #
41)	L9 AR-1268-1	8.509	7.365	94584	11450	0.484	0.050 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030922\
Data File : PP044382.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Mar 2022 10:07
Operator : YP\AJ
Sample : N1827-08 10X
Misc :
ALS Vial : 46 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 09 13:46:48 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 03 02:21:53 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	8.602	7.425	45230	40592	0.248	0.198
43)	L9 AR-1268-3	8.847	7.645	18359	14725	0.119	0.084 #
44)	L9 AR-1268-4	9.293	7.970	79146	30105	1.265	0.371 #
45)	L9 AR-1268-5	9.714	8.279	86824	12297	0.185	0.023 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030922\
Data File : PP044382.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Mar 2022 10:07
Operator : YP\AJ
Sample : N1827-08 10X
Misc :
ALS Vial : 46 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 09 13:46:48 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 03 02:21:53 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.910 min
Delta R.T.: -0.002 min
Response: 2830474
Conc: 1.37 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.166 min
Delta R.T.: -0.001 min
Response: 3014191
Conc: 1.72 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.047 min
Delta R.T.: -0.005 min
Response: 2649452
Conc: 1.92 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.531 min
Delta R.T.: -0.006 min
Response: 3094275
Conc: 2.02 ng/ml



#3 AR-1016-1


R.T.: 5.158 min
Delta R.T.: -0.007 min
Response: 59250
Conc: 0.87 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.309 min
Delta R.T.: -0.003 min
Response: 19273
Conc: 0.35 ng/ml




#4 AR-1016-2

R.T.: 5.195 min
Delta R.T.: 0.007 min
Response: 41396
Conc: 0.41 ng/ml



#4 AR-1016-2


R.T.: 4.333 min
Delta R.T.: 0.002 min
Response: 36989
Conc: 0.48 ng/ml



#5 AR-1016-3


R.T.: 5.264 min
Delta R.T.: 0.009 min
Response: 124557
Conc: 1.99 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.513 min
Delta R.T.: -0.003 min
Response: 5173
Conc: 0.12 ng/ml




#6 AR-1016-4

R.T.: 5.360 min
Delta R.T.: 0.000 min
Response: 36695
Conc: 0.71 ng/ml



#6 AR-1016-4


R.T.: 4.557 min
Delta R.T.: -0.008 min
Response: 27048
Conc: 0.80 ng/ml



#7 AR-1016-5


R.T.: 5.670 min
Delta R.T.: -0.011 min
Response: 73028
Conc: 1.45 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 4.792 min
Delta R.T.: 0.003 min
Response: 26863
Conc: 0.65 ng/ml




#8 AR-1221-1

R.T.: 4.132 min
Delta R.T.: -0.006 min
Response: 87846
Conc: 3.56 ng/ml



#8 AR-1221-1


R.T.: 3.387 min
Delta R.T.: -0.007 min
Response: 16914
Conc: 0.74 ng/ml



#9 AR-1221-2


R.T.: 4.236 min
Delta R.T.: 0.005 min
Response: 86168
Conc: 4.59 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.479 min
Delta R.T.: -0.004 min
Response: 36945
Conc: 2.14 ng/ml




#10 AR-1221-3

R.T.: 4.304 min
Delta R.T.: -0.007 min
Response: 226244
Conc: 4.04 ng/ml



#10 AR-1221-3


R.T.: 3.561 min
Delta R.T.: -0.002 min
Response: 17989
Conc: 0.36 ng/ml



#11 AR-1232-1


R.T.: 4.304 min
Delta R.T.: -0.007 min
Response: 226244
Conc: 4.93 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 3.561 min
Delta R.T.: -0.002 min
Response: 17989
Conc: 0.43 ng/ml




#12 AR-1232-2

R.T.: 4.880 min
Delta R.T.: 0.004 min
Response: 66237
Conc: 2.94 ng/ml



#12 AR-1232-2


R.T.: 4.333 min
Delta R.T.: 0.003 min
Response: 36989
Conc: 1.07 ng/ml



#13 AR-1232-3


R.T.: 5.195 min
Delta R.T.: 0.007 min
Response: 41396
Conc: 0.96 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.513 min
Delta R.T.: -0.002 min
Response: 5173
Conc: 0.28 ng/ml




#14 AR-1232-4

R.T.: 5.360 min
Delta R.T.: 0.000 min
Response: 36695
Conc: 1.83 ng/ml



#14 AR-1232-4


R.T.: 4.610 min
Delta R.T.: 0.002 min
Response: 15833
Conc: 0.96 ng/ml



#15 AR-1232-5


R.T.: 5.466 min
Delta R.T.: 0.003 min
Response: 175518
Conc: 11.57 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 4.787 min
Delta R.T.: -0.003 min
Response: 24225
Conc: 1.34 ng/ml




#16 AR-1242-1

R.T.: 5.158 min
Delta R.T.: -0.008 min
Response: 59250
Conc: 1.09 ng/ml



#16 AR-1242-1


R.T.: 4.309 min
Delta R.T.: -0.003 min
Response: 19273
Conc: 0.42 ng/ml



#17 AR-1242-2


R.T.: 5.195 min
Delta R.T.: 0.007 min
Response: 41396
Conc: 0.52 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.333 min
Delta R.T.: 0.002 min
Response: 36989
Conc: 0.59 ng/ml




#18 AR-1242-3

R.T.: 5.264 min
Delta R.T.: 0.008 min
Response: 124557
Conc: 2.64 ng/ml



#18 AR-1242-3


R.T.: 4.513 min
Delta R.T.: -0.002 min
Response: 5173
Conc: 0.14 ng/ml



#19 AR-1242-4


R.T.: 5.360 min
Delta R.T.: 0.000 min
Response: 36695
Conc: 0.95 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.610 min
Delta R.T.: 0.002 min
Response: 15833
Conc: 0.46 ng/ml




#20 AR-1242-5

R.T.: 6.160 min
Delta R.T.: -0.001 min
Response: 348875
Conc: 8.18 ng/ml



#20 AR-1242-5


R.T.: 5.169 min
Delta R.T.: 0.002 min
Response: 22947
Conc: 0.54 ng/ml



#21 AR-1248-1


R.T.: 5.158 min
Delta R.T.: -0.008 min
Response: 59250
Conc: 1.43 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.309 min
Delta R.T.: -0.003 min
Response: 19273
Conc: 0.54 ng/ml




#22 AR-1248-2

R.T.: 5.466 min
Delta R.T.: 0.004 min
Response: 175518
Conc: 3.13 ng/ml



#22 AR-1248-2


R.T.: 4.557 min
Delta R.T.: -0.008 min
Response: 27048
Conc: 0.57 ng/ml



#23 AR-1248-3


R.T.: 5.670 min
Delta R.T.: -0.011 min
Response: 73028
Conc: 1.07 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.610 min
Delta R.T.: 0.002 min
Response: 15833
Conc: 0.32 ng/ml




#24 AR-1248-4

R.T.: 6.124 min
Delta R.T.: 0.004 min
Response: 96300
Conc: 1.32 ng/ml



#24 AR-1248-4


R.T.: 4.787 min
Delta R.T.: -0.003 min
Response: 24225
Conc: 0.43 ng/ml



#25 AR-1248-5


R.T.: 6.160 min
Delta R.T.: 0.000 min
Response: 348875
Conc: 4.75 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.226 min
Delta R.T.: 0.016 min
Response: 51273
Conc: 0.91 ng/ml




#26 AR-1254-1

R.T.: 6.089 min
Delta R.T.: -0.002 min
Response: 85894
Conc: 1.15 ng/ml



#26 AR-1254-1


R.T.: 5.169 min
Delta R.T.: 0.002 min
Response: 22947
Conc: 0.28 ng/ml



#27 AR-1254-2


R.T.: 6.329 min
Delta R.T.: -0.001 min
Response: 132282
Conc: 1.18 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.334 min
Delta R.T.: 0.006 min
Response: 30372
Conc: 0.42 ng/ml




#28 AR-1254-3

R.T.: 6.728 min
Delta R.T.: -0.003 min
Response: 303697
Conc: 2.60 ng/ml



#28 AR-1254-3


R.T.: 5.758 min
Delta R.T.: -0.001 min
Response: 19432
Conc: 0.17 ng/ml



#29 AR-1254-4


R.T.: 7.027 min
Delta R.T.: -0.018 min
Response: 72890
Conc: 0.91 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.006 min
Delta R.T.: -0.003 min
Response: 27811
Conc: 0.37 ng/ml




#30 AR-1254-5

R.T.: 7.508 min
Delta R.T.: 0.003 min
Response: 319176
Conc: 3.67 ng/ml



#30 AR-1254-5


R.T.: 6.458 min
Delta R.T.: -0.002 min
Response: 38808
Conc: 0.36 ng/ml



#31 AR-1260-1


R.T.: 6.917 min
Delta R.T.: 0.006 min
Response: 118941
Conc: 1.35 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 5.904 min
Delta R.T.: 0.004 min
Response: 26864
Conc: 0.33 ng/ml




#32 AR-1260-2

R.T.: 7.185 min
Delta R.T.: -0.010 min
Response: 232954
Conc: 2.43 ng/ml



#32 AR-1260-2


R.T.: 6.108 min
Delta R.T.: 0.001 min
Response: 22712
Conc: 0.23 ng/ml



#33 AR-1260-3


R.T.: 7.586 min
Delta R.T.: -0.004 min
Response: 301818
Conc: 4.65 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.263 min
Delta R.T.: -0.007 min
Response: 94071
Conc: 1.01 ng/ml




#34 AR-1260-4

R.T.: 7.844 min
Delta R.T.: 0.007 min
Response: 618597
Conc: 7.93 ng/ml



#34 AR-1260-4


R.T.: 6.780 min
Delta R.T.: -0.003 min
Response: 17763
Conc: 0.25 ng/ml



#35 AR-1260-5


R.T.: 8.191 min
Delta R.T.: 0.011 min
Response: 666380
Conc: 4.75 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.047 min
Delta R.T.: -0.003 min
Response: 10621
Conc: 0.06 ng/ml




#36 AR-1262-1

R.T.: 7.586 min
Delta R.T.: -0.005 min
Response: 301818
Conc: 2.99 ng/ml



#36 AR-1262-1


R.T.: 6.498 min
Delta R.T.: -0.003 min
Response: 29393
Conc: 0.28 ng/ml



#37 AR-1262-2


R.T.: 8.191 min
Delta R.T.: 0.011 min
Response: 666380
Conc: 4.01 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 6.780 min
Delta R.T.: -0.002 min
Response: 17763
Conc: 0.19 ng/ml




#38 AR-1262-3

R.T.: 8.509 min
Delta R.T.: -0.003 min
Response: 94584
Conc: 0.85 ng/ml



#38 AR-1262-3


R.T.: 7.348 min
Delta R.T.: -0.008 min
Response: 37018
Conc: 0.47 ng/ml



#39 AR-1262-4


R.T.: 8.602 min
Delta R.T.: -0.003 min
Response: 45230
Conc: 0.91 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.425 min
Delta R.T.: -0.001 min
Response: 40592
Conc: 0.28 ng/ml




#40 AR-1262-5

R.T.: 9.293 min
Delta R.T.: 0.006 min
Response: 79146
Conc: 1.39 ng/ml



#40 AR-1262-5


R.T.: 7.970 min
Delta R.T.: 0.001 min
Response: 30105
Conc: 0.42 ng/ml



#41 AR-1268-1


R.T.: 8.509 min
Delta R.T.: 0.000 min
Response: 94584
Conc: 0.48 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.365 min
Delta R.T.: 0.008 min
Response: 11450
Conc: 0.05 ng/ml




#42 AR-1268-2

R.T.: 8.602 min
Delta R.T.: -0.006 min
Response: 45230
Conc: 0.25 ng/ml



#42 AR-1268-2


R.T.: 7.425 min
Delta R.T.: -0.002 min
Response: 40592
Conc: 0.20 ng/ml



#43 AR-1268-3


R.T.: 8.847 min
Delta R.T.: 0.003 min
Response: 18359
Conc: 0.12 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.645 min
Delta R.T.: -0.003 min
Response: 14725
Conc: 0.08 ng/ml




#44 AR-1268-4

R.T.: 9.293 min
Delta R.T.: 0.006 min
Response: 79146
Conc: 1.26 ng/ml



#44 AR-1268-4


R.T.: 7.970 min
Delta R.T.: 0.000 min
Response: 30105
Conc: 0.37 ng/ml



#45 AR-1268-5

R.T.: 9.714 min
Delta R.T.: 0.004 min
Response: 86824
Conc: 0.18 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.279 min
Delta R.T.: 0.005 min
Response: 12297
Conc: 0.02 ng/ml