

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030922\
 Data File : PP044396.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Mar 2022 14:09
 Operator : YP\AJ
 Sample : PB143197BL
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 10 00:38:35 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 03 02:21:53 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.910	3.166	41706797	36664298	20.129	20.958
2)	SA Decachlor...	10.046	8.531	26933273	30916443	19.568	20.154
Target Compounds							
3)	L1 AR-1016-1	5.172	4.320	112568	23403	1.658	0.421 #
4)	L1 AR-1016-2	5.189	4.329	19920	11504	0.197	0.148
5)	L1 AR-1016-3	5.250	4.517	19841	12249	0.316	0.286
6)	L1 AR-1016-4	5.370	4.560	56432	14837	1.087	0.440 #
7)	L1 AR-1016-5	5.680	4.791	149767	51347	2.974	1.234 #
8)	L2 AR-1221-1	4.129	3.393	3272024	7880	132.419	0.343 #
9)	L2 AR-1221-2	4.235	3.503f	999532	76192	53.274	4.411 #
10)	L2 AR-1221-3	4.308	3.565	230989	8154	4.128	0.164 #
11)	L3 AR-1232-1	4.308	3.565	230989	8154	5.034	0.196 #
12)	L3 AR-1232-2	4.881	4.329	224141	11504	9.960	0.333 #
13)	L3 AR-1232-3	5.189	4.517	19920	12249	0.460	0.657 #
14)	L3 AR-1232-4	5.370	4.601	56432	31916	2.811	1.936 #
15)	L3 AR-1232-5	5.464	4.791	53210	51347	3.507	2.836
16)	L4 AR-1242-1	5.172	4.320	112568	23403	2.068	0.510 #
17)	L4 AR-1242-2	5.189	4.329	19920	11504	0.250	0.182 #
18)	L4 AR-1242-3	5.260	4.517	13851	12249	0.293	0.343
19)	L4 AR-1242-4	5.370	4.601	56432	31916	1.458	0.923 #
20)	L4 AR-1242-5	6.148	5.163	260782	29507	6.117	0.699 #
21)	L5 AR-1248-1	5.172	4.320	112568	23403	2.723	0.660 #
22)	L5 AR-1248-2	5.464	4.560	53210	14837	0.949	0.315 #
23)	L5 AR-1248-3	5.680	4.601	149767	31916	2.193	0.648 #
24)	L5 AR-1248-4	6.124	4.791	150464	51347	2.063	0.908 #
25)	L5 AR-1248-5	6.148	5.216	260782	46933	3.551	0.837 #
26)	L6 AR-1254-1	6.087	5.163	79377	29507	1.058	0.357 #
27)	L6 AR-1254-2	6.330	5.324	62619	34160	0.557	0.475
28)	L6 AR-1254-3	6.730	5.764	363162	38158	3.104	0.331 #
29)	L6 AR-1254-4	7.058	6.005	175960	21298	2.189	0.283 #
30)	L6 AR-1254-5	7.505	6.461	98516	22943	1.132	0.211 #
31)	L7 AR-1260-1	6.914	5.897	62259	17062	0.708	0.213 #
32)	L7 AR-1260-2	7.198	6.100	40658	27888	0.423	0.283 #
33)	L7 AR-1260-3	7.608f	6.271	185552	14421	2.860	0.155 #
34)	L7 AR-1260-4	7.838	6.783	87654	6066	1.124	0.085 #
35)	L7 AR-1260-5	8.179	7.049	104822	46197	0.748	0.280 #
36)	L8 AR-1262-1	7.608f	6.502	185552	13293	1.836	0.128 #
37)	L8 AR-1262-2	8.179	6.783	104822	6066	0.631	0.063 #
38)	L8 AR-1262-3	8.512	7.359	57267	8886	0.512	0.113 #
39)	L8 AR-1262-4	8.608	7.422	71095	32833	1.432	0.227 #
40)	L8 AR-1262-5	9.306f	7.966	60433	33022	1.059	0.461 #
41)	L9 AR-1268-1	8.512	7.359	57267	8886	0.293	0.039 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030922\
Data File : PP044396.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Mar 2022 14:09
Operator : YP\AJ
Sample : PB143197BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 10 00:38:35 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 03 02:21:53 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	8.608	7.422	71095	32833	0.390	0.160 #
43)	L9 AR-1268-3	8.854	7.644	24554	44107	0.159	0.252 #
44)	L9 AR-1268-4	9.268f	7.971	23158	23621	0.370	0.291
45)	L9 AR-1268-5	9.710	8.270	130241	188292	0.277	0.349 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030922\
Data File : PP044396.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Mar 2022 14:09
Operator : YP\AJ
Sample : PB143197BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 10 00:38:35 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 03 02:21:53 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.910 min
Delta R.T.: -0.002 min
Response: 41706797
Conc: 20.13 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.166 min
Delta R.T.: -0.001 min
Response: 36664298
Conc: 20.96 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.046 min
Delta R.T.: -0.005 min
Response: 26933273
Conc: 19.57 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.531 min
Delta R.T.: -0.006 min
Response: 30916443
Conc: 20.15 ng/ml



#3 AR-1016-1


R.T.: 5.172 min
Delta R.T.: 0.007 min
Response: 112568
Conc: 1.66 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.320 min
Delta R.T.: 0.007 min
Response: 23403
Conc: 0.42 ng/ml




#4 AR-1016-2

R.T.: 5.189 min
Delta R.T.: 0.001 min
Response: 19920
Conc: 0.20 ng/ml



#4 AR-1016-2


R.T.: 4.329 min
Delta R.T.: -0.002 min
Response: 11504
Conc: 0.15 ng/ml



#5 AR-1016-3


R.T.: 5.250 min
Delta R.T.: -0.005 min
Response: 19841
Conc: 0.32 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.517 min
Delta R.T.: 0.001 min
Response: 12249
Conc: 0.29 ng/ml




#6 AR-1016-4

R.T.: 5.370 min
Delta R.T.: 0.010 min
Response: 56432
Conc: 1.09 ng/ml



#6 AR-1016-4


R.T.: 4.560 min
Delta R.T.: -0.005 min
Response: 14837
Conc: 0.44 ng/ml



#7 AR-1016-5


R.T.: 5.680 min
Delta R.T.: -0.001 min
Response: 149767
Conc: 2.97 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 4.791 min
Delta R.T.: 0.002 min
Response: 51347
Conc: 1.23 ng/ml




#8 AR-1221-1

R.T.: 4.129 min
Delta R.T.: -0.009 min
Response: 3272024
Conc: 132.42 ng/ml



#8 AR-1221-1


R.T.: 3.393 min
Delta R.T.: -0.001 min
Response: 7880
Conc: 0.34 ng/ml



#9 AR-1221-2


R.T.: 4.235 min
Delta R.T.: 0.004 min
Response: 999532
Conc: 53.27 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.503 min
Delta R.T.: 0.019 min
Response: 76192
Conc: 4.41 ng/ml




#10 AR-1221-3

R.T.: 4.308 min
Delta R.T.: -0.002 min
Response: 230989
Conc: 4.13 ng/ml



#10 AR-1221-3


R.T.: 3.565 min
Delta R.T.: 0.002 min
Response: 8154
Conc: 0.16 ng/ml



#11 AR-1232-1


R.T.: 4.308 min
Delta R.T.: -0.003 min
Response: 230989
Conc: 5.03 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 3.565 min
Delta R.T.: 0.002 min
Response: 8154
Conc: 0.20 ng/ml




#12 AR-1232-2

R.T.: 4.881 min
Delta R.T.: 0.005 min
Response: 224141
Conc: 9.96 ng/ml



#12 AR-1232-2


R.T.: 4.329 min
Delta R.T.: -0.002 min
Response: 11504
Conc: 0.33 ng/ml



#13 AR-1232-3


R.T.: 5.189 min
Delta R.T.: 0.000 min
Response: 19920
Conc: 0.46 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.517 min
Delta R.T.: 0.001 min
Response: 12249
Conc: 0.66 ng/ml




#14 AR-1232-4

R.T.: 5.370 min
Delta R.T.: 0.009 min
Response: 56432
Conc: 2.81 ng/ml



#14 AR-1232-4


R.T.: 4.601 min
Delta R.T.: -0.007 min
Response: 31916
Conc: 1.94 ng/ml



#15 AR-1232-5


R.T.: 5.464 min
Delta R.T.: 0.001 min
Response: 53210
Conc: 3.51 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 4.791 min
Delta R.T.: 0.002 min
Response: 51347
Conc: 2.84 ng/ml




#16 AR-1242-1

R.T.: 5.172 min
Delta R.T.: 0.007 min
Response: 112568
Conc: 2.07 ng/ml



#16 AR-1242-1


R.T.: 4.320 min
Delta R.T.: 0.008 min
Response: 23403
Conc: 0.51 ng/ml



#17 AR-1242-2


R.T.: 5.189 min
Delta R.T.: 0.000 min
Response: 19920
Conc: 0.25 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.329 min
Delta R.T.: -0.002 min
Response: 11504
Conc: 0.18 ng/ml




#18 AR-1242-3

R.T.: 5.260 min
Delta R.T.: 0.004 min
Response: 13851
Conc: 0.29 ng/ml



#18 AR-1242-3


R.T.: 4.517 min
Delta R.T.: 0.001 min
Response: 12249
Conc: 0.34 ng/ml



#19 AR-1242-4


R.T.: 5.370 min
Delta R.T.: 0.009 min
Response: 56432
Conc: 1.46 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.601 min
Delta R.T.: -0.007 min
Response: 31916
Conc: 0.92 ng/ml




#20 AR-1242-5

R.T.: 6.148 min
Delta R.T.: -0.014 min
Response: 260782
Conc: 6.12 ng/ml



#20 AR-1242-5


R.T.: 5.163 min
Delta R.T.: -0.004 min
Response: 29507
Conc: 0.70 ng/ml



#21 AR-1248-1


R.T.: 5.172 min
Delta R.T.: 0.007 min
Response: 112568
Conc: 2.72 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.320 min
Delta R.T.: 0.007 min
Response: 23403
Conc: 0.66 ng/ml




#22 AR-1248-2

R.T.: 5.464 min
Delta R.T.: 0.002 min
Response: 53210
Conc: 0.95 ng/ml



#22 AR-1248-2


R.T.: 4.560 min
Delta R.T.: -0.005 min
Response: 14837
Conc: 0.32 ng/ml



#23 AR-1248-3


R.T.: 5.680 min
Delta R.T.: 0.000 min
Response: 149767
Conc: 2.19 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.601 min
Delta R.T.: -0.007 min
Response: 31916
Conc: 0.65 ng/ml




#24 AR-1248-4

R.T.: 6.124 min
Delta R.T.: 0.004 min
Response: 150464
Conc: 2.06 ng/ml



#24 AR-1248-4


R.T.: 4.791 min
Delta R.T.: 0.002 min
Response: 51347
Conc: 0.91 ng/ml



#25 AR-1248-5


R.T.: 6.148 min
Delta R.T.: -0.014 min
Response: 260782
Conc: 3.55 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.216 min
Delta R.T.: 0.006 min
Response: 46933
Conc: 0.84 ng/ml




#26 AR-1254-1

R.T.: 6.087 min
Delta R.T.: -0.004 min
Response: 79377
Conc: 1.06 ng/ml



#26 AR-1254-1


R.T.: 5.163 min
Delta R.T.: -0.003 min
Response: 29507
Conc: 0.36 ng/ml



#27 AR-1254-2


R.T.: 6.330 min
Delta R.T.: 0.000 min
Response: 62619
Conc: 0.56 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.324 min
Delta R.T.: -0.003 min
Response: 34160
Conc: 0.47 ng/ml




#28 AR-1254-3

R.T.: 6.730 min
Delta R.T.: 0.000 min
Response: 363162
Conc: 3.10 ng/ml



#28 AR-1254-3


R.T.: 5.764 min
Delta R.T.: 0.005 min
Response: 38158
Conc: 0.33 ng/ml



#29 AR-1254-4


R.T.: 7.058 min
Delta R.T.: 0.013 min
Response: 175960
Conc: 2.19 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.005 min
Delta R.T.: -0.003 min
Response: 21298
Conc: 0.28 ng/ml




#30 AR-1254-5

R.T.: 7.505 min
Delta R.T.: 0.000 min
Response: 98516
Conc: 1.13 ng/ml



#30 AR-1254-5


R.T.: 6.461 min
Delta R.T.: 0.000 min
Response: 22943
Conc: 0.21 ng/ml



#31 AR-1260-1


R.T.: 6.914 min
Delta R.T.: 0.003 min
Response: 62259
Conc: 0.71 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 5.897 min
Delta R.T.: -0.003 min
Response: 17062
Conc: 0.21 ng/ml




#32 AR-1260-2

R.T.: 7.198 min
Delta R.T.: 0.003 min
Response: 40658
Conc: 0.42 ng/ml



#32 AR-1260-2


R.T.: 6.100 min
Delta R.T.: -0.006 min
Response: 27888
Conc: 0.28 ng/ml



#33 AR-1260-3


R.T.: 7.608 min
Delta R.T.: 0.019 min
Response: 185552
Conc: 2.86 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.271 min
Delta R.T.: 0.002 min
Response: 14421
Conc: 0.16 ng/ml




#34 AR-1260-4

R.T.: 7.838 min
Delta R.T.: 0.002 min
Response: 87654
Conc: 1.12 ng/ml



#34 AR-1260-4


R.T.: 6.783 min
Delta R.T.: 0.000 min
Response: 6066
Conc: 0.09 ng/ml



#35 AR-1260-5


R.T.: 8.179 min
Delta R.T.: -0.002 min
Response: 104822
Conc: 0.75 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.049 min
Delta R.T.: 0.000 min
Response: 46197
Conc: 0.28 ng/ml




#36 AR-1262-1

R.T.: 7.608 min
Delta R.T.: 0.018 min
Response: 185552
Conc: 1.84 ng/ml



#36 AR-1262-1


R.T.: 6.502 min
Delta R.T.: 0.000 min
Response: 13293
Conc: 0.13 ng/ml



#37 AR-1262-2


R.T.: 8.179 min
Delta R.T.: -0.002 min
Response: 104822
Conc: 0.63 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 6.783 min
Delta R.T.: 0.002 min
Response: 6066
Conc: 0.06 ng/ml




#38 AR-1262-3

R.T.: 8.512 min
Delta R.T.: 0.000 min
Response: 57267
Conc: 0.51 ng/ml



#38 AR-1262-3


R.T.: 7.359 min
Delta R.T.: 0.003 min
Response: 8886
Conc: 0.11 ng/ml



#39 AR-1262-4


R.T.: 8.608 min
Delta R.T.: 0.003 min
Response: 71095
Conc: 1.43 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.422 min
Delta R.T.: -0.005 min
Response: 32833
Conc: 0.23 ng/ml




#40 AR-1262-5

R.T.: 9.306 min
Delta R.T.: 0.019 min
Response: 60433
Conc: 1.06 ng/ml



#40 AR-1262-5


R.T.: 7.966 min
Delta R.T.: -0.003 min
Response: 33022
Conc: 0.46 ng/ml



#41 AR-1268-1


R.T.: 8.512 min
Delta R.T.: 0.003 min
Response: 57267
Conc: 0.29 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.359 min
Delta R.T.: 0.003 min
Response: 8886
Conc: 0.04 ng/ml




#42 AR-1268-2

R.T.: 8.608 min
Delta R.T.: 0.000 min
Response: 71095
Conc: 0.39 ng/ml



#42 AR-1268-2


R.T.: 7.422 min
Delta R.T.: -0.006 min
Response: 32833
Conc: 0.16 ng/ml



#43 AR-1268-3


R.T.: 8.854 min
Delta R.T.: 0.009 min
Response: 24554
Conc: 0.16 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.644 min
Delta R.T.: -0.005 min
Response: 44107
Conc: 0.25 ng/ml




#44 AR-1268-4

R.T.: 9.268 min
Delta R.T.: -0.019 min
Response: 23158
Conc: 0.37 ng/ml



#44 AR-1268-4


R.T.: 7.971 min
Delta R.T.: 0.002 min
Response: 23621
Conc: 0.29 ng/ml



#45 AR-1268-5

R.T.: 9.710 min
Delta R.T.: 0.000 min
Response: 130241
Conc: 0.28 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.270 min
Delta R.T.: -0.004 min
Response: 188292
Conc: 0.35 ng/ml