

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030922\
 Data File : PP044453.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Mar 2022 05:35
 Operator : YP\AJ
 Sample : N1845-14
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 SAMPLE-14

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 10 06:07:30 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 03 02:21:53 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.910	3.166	42074858	38532100	20.307	22.025
2) SA Decachlor...	10.046	8.531	27143685	27591719	19.721	17.987
Target Compounds						
26) L6 AR-1254-1	6.088	5.163	23189157	22646203	309.197	273.923
27) L6 AR-1254-2	6.327	5.323	35142712	23598528	312.758	328.037
28) L6 AR-1254-3	6.727	5.755	36007922	38088200	307.717	330.591
29) L6 AR-1254-4	7.040	6.004	26413127	26177030	328.517	347.714
30) L6 AR-1254-5	7.502	6.455	27950624	48514493	321.104	445.167 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030922\
Data File : PP044453.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Mar 2022 05:35
Operator : YP\AJ
Sample : N1845-14
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
SAMPLE-14

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 10 06:07:30 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 03 02:21:53 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.910 min
Delta R.T.: -0.002 min
Response: 42074858
Conc: 20.31 ng/ml

Instrument :
ECD_P
ClientSampleId :
SAMPLE-14




#1 Tetrachloro-m-xylene

R.T.: 3.166 min
Delta R.T.: 0.000 min
Response: 38532100
Conc: 22.03 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.046 min
Delta R.T.: -0.006 min
Response: 27143685
Conc: 19.72 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.531 min
Delta R.T.: -0.006 min
Response: 27591719
Conc: 17.99 ng/ml



#26 AR-1254-1


R.T.: 6.088 min
Delta R.T.: -0.003 min
Response: 23189157
Conc: 309.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
SAMPLE-14




#26 AR-1254-1

R.T.: 5.163 min
Delta R.T.: -0.004 min
Response: 22646203
Conc: 273.92 ng/ml




#27 AR-1254-2

R.T.: 6.327 min
Delta R.T.: -0.003 min
Response: 35142712
Conc: 312.76 ng/ml



#27 AR-1254-2


R.T.: 5.323 min
Delta R.T.: -0.004 min
Response: 23598528
Conc: 328.04 ng/ml



#28 AR-1254-3


R.T.: 6.727 min
Delta R.T.: -0.003 min
Response: 36007922
Conc: 307.72 ng/ml

Instrument :
ECD_P
ClientSampleId :
SAMPLE-14




#28 AR-1254-3

R.T.: 5.755 min
Delta R.T.: -0.004 min
Response: 38088200
Conc: 330.59 ng/ml




#29 AR-1254-4

R.T.: 7.040 min
Delta R.T.: -0.005 min
Response: 26413127
Conc: 328.52 ng/ml



#29 AR-1254-4


R.T.: 6.004 min
Delta R.T.: -0.004 min
Response: 26177030
Conc: 347.71 ng/ml



#30 AR-1254-5

R.T.: 7.502 min
Delta R.T.: -0.004 min
Response: 27950624
Conc: 321.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
SAMPLE-14



#30 AR-1254-5

R.T.: 6.455 min
Delta R.T.: -0.005 min
Response: 48514493
Conc: 445.17 ng/ml