

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030922\
 Data File : PP044455.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Mar 2022 06:35
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 10 07:50:26 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 03 02:21:53 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.912	3.166	112.4E6	94329731	54.230	53.920
2) SA Decachlor...	10.046	8.531	63704608	65895206	46.285	42.957
Target Compounds						
3) L1 AR-1016-1	5.163	4.309	33034376	28881816	486.477	519.449
4) L1 AR-1016-2	5.186	4.328	47976003	40503426	473.741	522.736
5) L1 AR-1016-3	5.253	4.513	28968998	22640966	461.853	527.804
6) L1 AR-1016-4	5.358	4.562	23658615	17900614	455.729	531.086
7) L1 AR-1016-5	5.678	4.787	23062441	20902833	457.900	502.524
31) L7 AR-1260-1	6.907	5.895	37630526	39540431	428.203	492.916
32) L7 AR-1260-2	7.191	6.102	41547338	47953067	432.516	486.812
33) L7 AR-1260-3	7.585	6.265	28207360	43863593	434.762	472.928
34) L7 AR-1260-4	7.833	6.777	34317372	33500436	439.988	469.828
35) L7 AR-1260-5	8.176	7.045	63136319	75824382	450.472	459.210



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030922\
Data File : PP044455.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Mar 2022 06:35
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 10 07:50:26 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 03 02:21:53 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.912 min
Delta R.T.: 0.000 min
Response: 112363766
Conc: 54.23 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.166 min
Delta R.T.: 0.000 min
Response: 94329731
Conc: 53.92 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.046 min
Delta R.T.: -0.005 min
Response: 63704608
Conc: 46.28 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.531 min
Delta R.T.: -0.005 min
Response: 65895206
Conc: 42.96 ng/ml



#3 AR-1016-1


R.T.: 5.163 min
Delta R.T.: -0.002 min
Response: 33034376
Conc: 486.48 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.309 min
Delta R.T.: -0.003 min
Response: 28881816
Conc: 519.45 ng/ml




#4 AR-1016-2

R.T.: 5.186 min
Delta R.T.: -0.002 min
Response: 47976003
Conc: 473.74 ng/ml



#4 AR-1016-2


R.T.: 4.328 min
Delta R.T.: -0.003 min
Response: 40503426
Conc: 522.74 ng/ml



#5 AR-1016-3


R.T.: 5.253 min
Delta R.T.: -0.002 min
Response: 28968998
Conc: 461.85 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.513 min
Delta R.T.: -0.003 min
Response: 22640966
Conc: 527.80 ng/ml




#6 AR-1016-4

R.T.: 5.358 min
Delta R.T.: -0.002 min
Response: 23658615
Conc: 455.73 ng/ml



#6 AR-1016-4


R.T.: 4.562 min
Delta R.T.: -0.004 min
Response: 17900614
Conc: 531.09 ng/ml



#7 AR-1016-5


R.T.: 5.678 min
Delta R.T.: -0.002 min
Response: 23062441
Conc: 457.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 4.787 min
Delta R.T.: -0.003 min
Response: 20902833
Conc: 502.52 ng/ml




#31 AR-1260-1

R.T.: 6.907 min
Delta R.T.: -0.004 min
Response: 37630526
Conc: 428.20 ng/ml



#31 AR-1260-1


R.T.: 5.895 min
Delta R.T.: -0.005 min
Response: 39540431
Conc: 492.92 ng/ml



#32 AR-1260-2


R.T.: 7.191 min
Delta R.T.: -0.003 min
Response: 41547338
Conc: 432.52 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.102 min
Delta R.T.: -0.004 min
Response: 47953067
Conc: 486.81 ng/ml




#33 AR-1260-3

R.T.: 7.585 min
Delta R.T.: -0.004 min
Response: 28207360
Conc: 434.76 ng/ml



#33 AR-1260-3


R.T.: 6.265 min
Delta R.T.: -0.004 min
Response: 43863593
Conc: 472.93 ng/ml



#34 AR-1260-4


R.T.: 7.833 min
Delta R.T.: -0.003 min
Response: 34317372
Conc: 439.99 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 6.777 min
Delta R.T.: -0.005 min
Response: 33500436
Conc: 469.83 ng/ml



#35 AR-1260-5

R.T.: 8.176 min
Delta R.T.: -0.004 min
Response: 63136319
Conc: 450.47 ng/ml



#35 AR-1260-5

R.T.: 7.045 min
Delta R.T.: -0.005 min
Response: 75824382
Conc: 459.21 ng/ml