Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030923\
Data File : PP056309.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Mar 2023 09:52
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 08 20:30:16 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030823.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 08 02:03:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml	
	System Monitoring Compounds							
1)	SA Tetrachlo...	4.318	3.560	81508576	71144671	52.494	50.419	
2)	SA Decachlor...	10.059	8.557	106.1E6	56807590	53.417	50.818	
	Target Compounds							
3)	L1 AR-1016-1	5.488	4.641	870757	621372	15.041	14.749	
4)	L1 AR-1016-2	5.513	4.659	1194442	568724	13.943	9.383	#
5)	L1 AR-1016-3	5.575	4.836	563302	655253	10.328	20.245	#
6)	L1 AR-1016-4	5.669	4.877	446296	16063711	10.278	572.367	#
7)	L1 AR-1016-5	5.967	5.089	13331849	13105741	290.778	366.315	#
8)	L2 AR-1221-1	4.526	3.768	79469	45385	3.967	2.610	#
9)	L2 AR-1221-2	4.625	3.858	1096433	978397	72.161	75.715	
10)	L2 AR-1221-3	4.690	3.933	323348	237745	6.951	5.860	
11)	L3 AR-1232-1	4.690	3.933	323348	237745	8.340	7.224	
12)	L3 AR-1232-2	5.221	4.659	459740	568724	22.725	20.512	
13)	L3 AR-1232-3	5.513	4.836	1194442	655253	30.664	46.266	#
14)	L3 AR-1232-4	5.669	4.917	446296	6289739	22.822	442.848	#
15)	L3 AR-1232-5	5.763	5.089	21682119	13105741	1368.822	854.527	#
16)	L4 AR-1242-1	5.488	4.641	870757	621372	17.867	17.571	
17)	L4 AR-1242-2	5.513	4.659	1194442	568724	16.741	11.193	#
18)	L4 AR-1242-3	5.575	4.836	563302	655253	12.374	24.282	#
19)	L4 AR-1242-4	5.669	4.917	446296	6289739	12.486	214.671	#
20)	L4 AR-1242-5	6.414	5.441	13896738	39031980	325.069	1209.779	#
21)	L5 AR-1248-1	5.488	4.641	870757	621372	24.453	23.932	
22)	L5 AR-1248-2	5.763	4.877	21682119	16063711	388.351	412.960	
23)	L5 AR-1248-3	5.967	4.917	13331849	6289739	219.093	154.238	#
24)	L5 AR-1248-4	6.373	5.089	20679290	13105741	299.372	279.860	
25)	L5 AR-1248-5	6.414	5.482	13896738	5762434	204.529	139.536	#
26)	L6 AR-1254-1	6.347	5.441	35836020	39031980	474.769	543.886	
27)	L6 AR-1254-2	6.566	5.590	60007346	32745961	517.174	536.986	
28)	L6 AR-1254-3	6.933	5.993	68002687	47842302	554.556	516.796	
29)	L6 AR-1254-4	7.220	6.222	54435465	27624108	599.438	512.433	
30)	L6 AR-1254-5	7.641	6.642	55514597	39917618	536.986	510.756	
31)	L7 AR-1260-1	7.098	6.124	33770738	24746828	342.682	363.011	
32)	L7 AR-1260-2	7.356	6.313	36726833	18940410	316.625	247.594	
33)	L7 AR-1260-3	7.701f	6.464	8130469	18854749	94.980	260.749	#
34)	L7 AR-1260-4	7.929f	6.939	21527808	1960633	217.749	32.695	#
35)	L7 AR-1260-5	8.259	7.181	5881766	4641491	29.432	36.324	
36)	L8 AR-1262-1	7.701f	6.731	8130469	1159249	68.119	35.271	#
37)	L8 AR-1262-2	8.259	6.939	5881766	1960633	26.948	27.504	
38)	L8 AR-1262-3	8.583	7.467	6360055	423620	41.628	7.779	#
39)	L8 AR-1262-4	8.632f	7.527	2088758	4561508	25.983	47.180	#
40)	L8 AR-1262-5	9.307	8.028	126426	296673	1.526	6.602	#
41)	L9 AR-1268-1	8.583f	7.467	6360055	423620	23.077	2.511	#

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030923\
Data File : PP056309.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Mar 2023 09:52
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 08 20:30:16 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030823.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 08 02:03:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml	
42) L9	AR-1268-2	8.632f	7.527	2088758	4561508	8.100	30.503	#
43) L9	AR-1268-3	8.890	7.738	711034	87805	3.102	0.646	#
44) L9	AR-1268-4	9.307	8.028	126426	296673	1.354	5.813	#
45) L9	AR-1268-5	9.717	8.314	338376	417827	0.466	1.074	#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030923\
Data File : PP056309.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Mar 2023 09:52
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 08 20:30:16 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030823.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 08 02:03:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.318 min
Delta R.T.: 0.000 min
Response: 81508576
Conc: 52.49 ng/ml

Instrument :
ECD_P
ClientSampleId :



#1 Tetrachloro-m-xylene

R.T.: 3.560 min
Delta R.T.: 0.000 min
Response: 71144671
Conc: 50.42 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.059 min
Delta R.T.: -0.004 min
Response: 106106608
Conc: 53.42 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.557 min
Delta R.T.: -0.004 min
Response: 56807590
Conc: 50.82 ng/ml



#3 AR-1016-1

R.T.: 5.488 min
Delta R.T.: 0.000 min
Response: 870757
Conc: 15.04 ng/ml

Instrument :
ECD_P
ClientSampleId :



#3 AR-1016-1

R.T.: 4.641 min
Delta R.T.: 0.000 min
Response: 621372
Conc: 14.75 ng/ml



#4 AR-1016-2

R.T.: 5.513 min
Delta R.T.: 0.002 min
Response: 1194442
Conc: 13.94 ng/ml



#4 AR-1016-2

R.T.: 4.659 min
Delta R.T.: -0.001 min
Response: 568724
Conc: 9.38 ng/ml



#5 AR-1016-3

R.T.: 5.575 min
Delta R.T.: 0.001 min
Response: 563302
Conc: 10.33 ng/ml

Instrument :
ECD_P
ClientSampleId :



#5 AR-1016-3

R.T.: 4.836 min
Delta R.T.: 0.000 min
Response: 655253
Conc: 20.24 ng/ml



#6 AR-1016-4

R.T.: 5.669 min
Delta R.T.: -0.003 min
Response: 446296
Conc: 10.28 ng/ml



#6 AR-1016-4

R.T.: 4.877 min
Delta R.T.: 0.000 min
Response: 16063711
Conc: 572.37 ng/ml



#7 AR-1016-5

R.T.: 5.967 min
Delta R.T.: -0.001 min
Response: 13331849
Conc: 290.78 ng/ml

Instrument :
ECD_P
ClientSampleId :



#7 AR-1016-5

R.T.: 5.089 min
Delta R.T.: -0.001 min
Response: 13105741
Conc: 366.32 ng/ml



#8 AR-1221-1

R.T.: 4.526 min
Delta R.T.: 0.000 min
Response: 79469
Conc: 3.97 ng/ml



#8 AR-1221-1

R.T.: 3.768 min
Delta R.T.: -0.004 min
Response: 45385
Conc: 2.61 ng/ml



#9 AR-1221-2

R.T.: 4.625 min
Delta R.T.: 0.011 min
Response: 1096433
Conc: 72.16 ng/ml

Instrument :
ECD_P
ClientSampleId :



#9 AR-1221-2

R.T.: 3.858 min
Delta R.T.: 0.000 min
Response: 978397
Conc: 75.72 ng/ml



#10 AR-1221-3

R.T.: 4.690 min
Delta R.T.: 0.000 min
Response: 323348
Conc: 6.95 ng/ml



#10 AR-1221-3

R.T.: 3.933 min
Delta R.T.: 0.000 min
Response: 237745
Conc: 5.86 ng/ml



#11 AR-1232-1

R.T.: 4.690 min
Delta R.T.: 0.000 min
Response: 323348
Conc: 8.34 ng/ml

Instrument :
ECD_P
ClientSampleId :



#11 AR-1232-1

R.T.: 3.933 min
Delta R.T.: 0.000 min
Response: 237745
Conc: 7.22 ng/ml



#12 AR-1232-2

R.T.: 5.221 min
Delta R.T.: 0.000 min
Response: 459740
Conc: 22.72 ng/ml



#12 AR-1232-2

R.T.: 4.659 min
Delta R.T.: 0.000 min
Response: 568724
Conc: 20.51 ng/ml



#13 AR-1232-3

R.T.: 5.513 min
Delta R.T.: 0.002 min
Response: 1194442
Conc: 30.66 ng/ml

Instrument :
ECD_P
ClientSampleId :



#13 AR-1232-3

R.T.: 4.836 min
Delta R.T.: 0.001 min
Response: 655253
Conc: 46.27 ng/ml



#14 AR-1232-4

R.T.: 5.669 min
Delta R.T.: -0.002 min
Response: 446296
Conc: 22.82 ng/ml



#14 AR-1232-4

R.T.: 4.917 min
Delta R.T.: -0.002 min
Response: 6289739
Conc: 442.85 ng/ml



#15 AR-1232-5

R.T.: 5.763 min
Delta R.T.: 0.000 min
Response: 21682119
Conc: 1368.82 ng/m

Instrument :
ECD_P
ClientSampleId :



#15 AR-1232-5

R.T.: 5.089 min
Delta R.T.: -0.001 min
Response: 13105741
Conc: 854.53 ng/ml



#16 AR-1242-1

R.T.: 5.488 min
Delta R.T.: 0.000 min
Response: 870757
Conc: 17.87 ng/ml



#16 AR-1242-1

R.T.: 4.641 min
Delta R.T.: 0.000 min
Response: 621372
Conc: 17.57 ng/ml



#17 AR-1242-2

R.T.: 5.513 min
Delta R.T.: 0.003 min
Response: 1194442
Conc: 16.74 ng/ml

Instrument :
ECD_P
ClientSampleId :



#17 AR-1242-2

R.T.: 4.659 min
Delta R.T.: 0.000 min
Response: 568724
Conc: 11.19 ng/ml



#18 AR-1242-3

R.T.: 5.575 min
Delta R.T.: 0.002 min
Response: 563302
Conc: 12.37 ng/ml



#18 AR-1242-3

R.T.: 4.836 min
Delta R.T.: 0.002 min
Response: 655253
Conc: 24.28 ng/ml



#19 AR-1242-4

R.T.: 5.669 min
Delta R.T.: -0.002 min
Response: 446296
Conc: 12.49 ng/ml

Instrument :
ECD_P
ClientSampleId :



#19 AR-1242-4

R.T.: 4.917 min
Delta R.T.: -0.001 min
Response: 6289739
Conc: 214.67 ng/ml



#20 AR-1242-5

R.T.: 6.414 min
Delta R.T.: 0.003 min
Response: 13896738
Conc: 325.07 ng/ml



#20 AR-1242-5

R.T.: 5.441 min
Delta R.T.: 0.000 min
Response: 39031980
Conc: 1209.78 ng/ml



#21 AR-1248-1

R.T.: 5.488 min
Delta R.T.: 0.000 min
Response: 870757
Conc: 24.45 ng/ml

Instrument :
ECD_P
ClientSampleId :



#21 AR-1248-1

R.T.: 4.641 min
Delta R.T.: 0.000 min
Response: 621372
Conc: 23.93 ng/ml



#22 AR-1248-2

R.T.: 5.763 min
Delta R.T.: 0.000 min
Response: 21682119
Conc: 388.35 ng/ml



#22 AR-1248-2

R.T.: 4.877 min
Delta R.T.: 0.000 min
Response: 16063711
Conc: 412.96 ng/ml



#23 AR-1248-3

R.T.: 5.967 min
Delta R.T.: 0.000 min
Response: 13331849
Conc: 219.09 ng/ml

Instrument :
ECD_P
ClientSampleId :



#23 AR-1248-3

R.T.: 4.917 min
Delta R.T.: -0.001 min
Response: 6289739
Conc: 154.24 ng/ml



#24 AR-1248-4

R.T.: 6.373 min
Delta R.T.: 0.000 min
Response: 20679290
Conc: 299.37 ng/ml



#24 AR-1248-4

R.T.: 5.089 min
Delta R.T.: 0.000 min
Response: 13105741
Conc: 279.86 ng/ml



#25 AR-1248-5

R.T.: 6.414 min
Delta R.T.: 0.002 min
Response: 13896738
Conc: 204.53 ng/ml

Instrument :
ECD_P
ClientSampleId :



#25 AR-1248-5

R.T.: 5.482 min
Delta R.T.: 0.000 min
Response: 5762434
Conc: 139.54 ng/ml



#26 AR-1254-1

R.T.: 6.347 min
Delta R.T.: 0.000 min
Response: 35836020
Conc: 474.77 ng/ml



#26 AR-1254-1

R.T.: 5.441 min
Delta R.T.: 0.000 min
Response: 39031980
Conc: 543.89 ng/ml



#27 AR-1254-2

R.T.: 6.566 min
Delta R.T.: -0.001 min
Response: 60007346
Conc: 517.17 ng/ml

Instrument :
ECD_P
ClientSampleId :



#27 AR-1254-2

R.T.: 5.590 min
Delta R.T.: 0.000 min
Response: 32745961
Conc: 536.99 ng/ml



#28 AR-1254-3

R.T.: 6.933 min
Delta R.T.: 0.000 min
Response: 68002687
Conc: 554.56 ng/ml



#28 AR-1254-3

R.T.: 5.993 min
Delta R.T.: 0.000 min
Response: 47842302
Conc: 516.80 ng/ml



#29 AR-1254-4

R.T.: 7.220 min
Delta R.T.: 0.000 min
Response: 54435465
Conc: 599.44 ng/ml

Instrument :
ECD_P
ClientSampleId :



#29 AR-1254-4

R.T.: 6.222 min
Delta R.T.: -0.001 min
Response: 27624108
Conc: 512.43 ng/ml



#30 AR-1254-5

R.T.: 7.641 min
Delta R.T.: 0.000 min
Response: 55514597
Conc: 536.99 ng/ml



#30 AR-1254-5

R.T.: 6.642 min
Delta R.T.: 0.000 min
Response: 39917618
Conc: 510.76 ng/ml



#31 AR-1260-1

R.T.: 7.098 min
Delta R.T.: -0.002 min
Response: 33770738
Conc: 342.68 ng/ml

Instrument :
ECD_P
ClientSampleId :



#31 AR-1260-1

R.T.: 6.124 min
Delta R.T.: -0.002 min
Response: 24746828
Conc: 363.01 ng/ml



#32 AR-1260-2

R.T.: 7.356 min
Delta R.T.: -0.002 min
Response: 36726833
Conc: 316.63 ng/ml



#32 AR-1260-2

R.T.: 6.313 min
Delta R.T.: -0.002 min
Response: 18940410
Conc: 247.59 ng/ml



#33 AR-1260-3

R.T.: 7.701 min
Delta R.T.: -0.019 min
Response: 8130469
Conc: 94.98 ng/ml

Instrument :
ECD_P
ClientSampleId :



#33 AR-1260-3

R.T.: 6.464 min
Delta R.T.: -0.003 min
Response: 18854749
Conc: 260.75 ng/ml



#34 AR-1260-4

R.T.: 7.929 min
Delta R.T.: -0.017 min
Response: 21527808
Conc: 217.75 ng/ml



#34 AR-1260-4

R.T.: 6.939 min
Delta R.T.: -0.001 min
Response: 1960633
Conc: 32.70 ng/ml



#35 AR-1260-5

R.T.: 8.259 min
Delta R.T.: -0.003 min
Response: 5881766
Conc: 29.43 ng/ml

Instrument :
ECD_P
ClientSampleId :



#35 AR-1260-5

R.T.: 7.181 min
Delta R.T.: -0.003 min
Response: 4641491
Conc: 36.32 ng/ml



#36 AR-1262-1

R.T.: 7.701 min
Delta R.T.: -0.017 min
Response: 8130469
Conc: 68.12 ng/ml



#36 AR-1262-1

R.T.: 6.731 min
Delta R.T.: -0.001 min
Response: 1159249
Conc: 35.27 ng/ml



#37 AR-1262-2

R.T.: 8.259 min
Delta R.T.: 0.000 min
Response: 5881766
Conc: 26.95 ng/ml

Instrument :
ECD_P
ClientSampleId :



#37 AR-1262-2

R.T.: 6.939 min
Delta R.T.: 0.000 min
Response: 1960633
Conc: 27.50 ng/ml



#38 AR-1262-3

R.T.: 8.583 min
Delta R.T.: 0.012 min
Response: 6360055
Conc: 41.63 ng/ml



#38 AR-1262-3

R.T.: 7.467 min
Delta R.T.: 0.002 min
Response: 423620
Conc: 7.78 ng/ml



#39 AR-1262-4

R.T.: 8.632 min
Delta R.T.: -0.026 min
Response: 2088758
Conc: 25.98 ng/ml

Instrument :
ECD_P
ClientSampleId :



#39 AR-1262-4

R.T.: 7.527 min
Delta R.T.: -0.002 min
Response: 4561508
Conc: 47.18 ng/ml



#40 AR-1262-5

R.T.: 9.307 min
Delta R.T.: -0.005 min
Response: 126426
Conc: 1.53 ng/ml



#40 AR-1262-5

R.T.: 8.028 min
Delta R.T.: 0.003 min
Response: 296673
Conc: 6.60 ng/ml



#41 AR-1268-1

R.T.: 8.583 min
Delta R.T.: 0.016 min
Response: 6360055
Conc: 23.08 ng/ml

Instrument :
ECD_P
ClientSampleId :



#41 AR-1268-1

R.T.: 7.467 min
Delta R.T.: 0.002 min
Response: 423620
Conc: 2.51 ng/ml



#42 AR-1268-2

R.T.: 8.632 min
Delta R.T.: -0.029 min
Response: 2088758
Conc: 8.10 ng/ml



#42 AR-1268-2

R.T.: 7.527 min
Delta R.T.: -0.003 min
Response: 4561508
Conc: 30.50 ng/ml



#43 AR-1268-3

R.T.: 8.890 min
Delta R.T.: 0.002 min
Response: 711034
Conc: 3.10 ng/ml

Instrument :
ECD_P
ClientSampleId :



#43 AR-1268-3

R.T.: 7.738 min
Delta R.T.: 0.003 min
Response: 87805
Conc: 0.65 ng/ml



#44 AR-1268-4

R.T.: 9.307 min
Delta R.T.: -0.004 min
Response: 126426
Conc: 1.35 ng/ml



#44 AR-1268-4

R.T.: 8.028 min
Delta R.T.: 0.003 min
Response: 296673
Conc: 5.81 ng/ml



#45 AR-1268-5

R.T.: 9.717 min
Delta R.T.: -0.007 min
Response: 338376
Conc: 0.47 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.314 min
Delta R.T.: 0.004 min
Response: 417827
Conc: 1.07 ng/ml