

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031021\
 Data File : PP033864.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 11 Mar 2021 5:47
 Operator : DD\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 11 06:59:30 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030921.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 10 03:01:34 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.856	4.268	2118656	1005304	24.486	24.745
2) SA Decachlor...	10.784	9.594	2114382	683515	25.036	25.614
Target Compounds						
9) L2 AR-1221-2	5.206	4.623	28699	15623	40.843	48.283
24) L5 AR-1248-4	7.082f	0.000	132900	0	48.038	N.D. #
26) L6 AR-1254-1	7.082	0.000	132900	0	46.610	N.D. #
28) L6 AR-1254-3	7.699	0.000	7522	0	1.585	N.D. #
32) L7 AR-1260-2	0.000	7.280f	0	534998	N.D.	295.782 #
45) L9 AR-1268-5	0.000	9.332	0	98383	N.D.	10.761 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031021\
Data File : PP033864.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 11 Mar 2021 5:47
Operator : DD\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 11 06:59:30 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030921.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 10 03:01:34 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.856 min
Delta R.T.: -0.002 min
Response: 2118656
Conc: 24.49 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.268 min
Delta R.T.: -0.003 min
Response: 1005304
Conc: 24.75 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.784 min
Delta R.T.: -0.003 min
Response: 2114382
Conc: 25.04 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.594 min
Delta R.T.: -0.010 min
Response: 683515
Conc: 25.61 ng/ml



#9 AR-1221-2


R.T.: 5.206 min
Delta R.T.: 0.000 min
Response: 28699
Conc: 40.84 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 4.623 min
Delta R.T.: -0.006 min
Response: 15623
Conc: 48.28 ng/ml




#24 AR-1248-4

R.T.: 7.082 min
Delta R.T.: -0.025 min
Response: 132900
Conc: 48.04 ng/ml



#24 AR-1248-4


R.T.: 0.000 min
Exp R.T.: 6.024 min
Response: 0
Conc: N.D.



#26 AR-1254-1


R.T.: 7.082 min
Delta R.T.: 0.005 min
Response: 132900
Conc: 46.61 ng/ml

Instrument :
ECD_P
ClientSampleId :




#26 AR-1254-1

R.T.: 0.000 min
Exp R.T. : 6.403 min
Response: 0
Conc: N.D.




#28 AR-1254-3

R.T.: 7.699 min
Delta R.T.: 0.013 min
Response: 7522
Conc: 1.59 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T. : 6.981 min
Response: 0
Conc: N.D.



#32 AR-1260-2


R.T.: 0.000 min
Exp R.T.: 8.117 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#32 AR-1260-2

R.T.: 7.280 min
Delta R.T.: -0.030 min
Response: 534998
Conc: 295.78 ng/ml



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T.: 10.466 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 9.332 min
Delta R.T.: -0.006 min
Response: 98383
Conc: 10.76 ng/ml