

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031021\
 Data File : PP033807.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 10 Mar 2021 12:21
 Operator : DD\AJ
 Sample : M1458-01
 Misc : AR1262 LOD 25 PPB
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 LOD-MDL-SOIL-01-QT1-2021

Manual Integrations
 APPROVED

Ankita
 3/11/2021 3:47:26 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 11 01:13:44 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030921.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 10 03:01:34 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.857	4.270	1546768	725115	17.877	17.849
2)	SA Decachlor...	10.784	9.597	1596910	504315	18.909	18.899
Target Compounds							
36)	L8 AR-1262-1	8.480	7.639	126442	32575	26.574	30.644m
37)	L8 AR-1262-2	9.028	8.186	212321	77624	24.565	23.295m
38)	L8 AR-1262-3	9.329	8.473	144629	45529	23.831	31.881 #
39)	L8 AR-1262-4	9.420	8.536	128243	64676	25.872	25.653
40)	L8 AR-1262-5	10.064	9.033	79349	25933	24.108	22.751

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031021\
Data File : PP033807.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 10 Mar 2021 12:21
Operator : DD\AJ
Sample : M1458-01
Misc : AR1262 LOD 25 PPB
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
LOD-MDL-SOIL-01-QT1-2021

Manual Integrations
APPROVED

Ankita
3/11/2021 3:47:26 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 11 01:13:44 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030921.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 10 03:01:34 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

