

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031021\
 Data File : PP033824.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 10 Mar 2021 17:08
 Operator : DD\AJ
 Sample : M1458-01
 Misc : AR1232 LOD 40 PPB
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 LOD-MDL-SOIL-01-QT1-2021

Manual Integrations
 APPROVED

Ankita
 3/11/2021 3:47:43 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 11 01:18:17 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030921.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 10 03:01:34 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.856	4.268	1887004	886325	21.809	21.817
2) SA Decachlor...	10.781	9.595	1857731	594778	21.997	22.289
Target Compounds						
11) L3 AR-1232-1	5.289	4.715	139213	58491	80.465	71.997
12) L3 AR-1232-2	5.874	5.548	72461	47678	79.119	65.790
13) L3 AR-1232-3	6.192	5.739	127245	26920	75.281	70.003
14) L3 AR-1232-4	6.362	5.835	69823	20670	81.736m	61.975
15) L3 AR-1232-5	6.461	6.020	45196	25174	76.186m	69.465

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031021\
Data File : PP033824.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 10 Mar 2021 17:08
Operator : DD\AJ
Sample : M1458-01
Misc : AR1232 LOD 40 PPB
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
LOD-MDL-SOIL-01-QT1-2021

Manual Integrations
APPROVED

Ankita
3/11/2021 3:47:43 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 11 01:18:17 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030921.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 10 03:01:34 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

