

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031022\
 Data File : PP044474.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Mar 2022 14:13
 Operator : YP\AJ
 Sample : N1645-08
 Misc : AR1242 LOQ 50 PPB
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 LOQ-WATER-02-QT1-2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 10 22:55:57 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 03 02:21:53 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.910	3.166	41253134	37062895	19.910	21.186
2) SA Decachlor...	10.045	8.529	27535280	30575804	20.006	19.932
Target Compounds						
16) L4 AR-1242-1	5.162	4.308	2311253	2329041	42.461	50.789
17) L4 AR-1242-2	5.186	4.327	3470934	3274990	43.585	51.823
18) L4 AR-1242-3	5.252	4.511	2076133	1766866	43.964	49.466
19) L4 AR-1242-4	5.356	4.603	1532567	1757830	39.597	50.810 #
20) L4 AR-1242-5	6.159	5.162	2041590	2106336	47.885	49.868

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031022\
Data File : PP044474.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Mar 2022 14:13
Operator : YP\AJ
Sample : N1645-08
Misc : AR1242 LOQ 50 PPB
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
LOQ-WATER-02-QT1-2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 10 22:55:57 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 03 02:21:53 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

