

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031022\
 Data File : PP044469.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Mar 2022 12:41
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 10 22:50:09 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 03 02:21:53 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.910	3.166	102.5E6	84647792	49.476	48.386
2) SA Decachlor...	10.042	8.528	62265930	66217036	45.240	43.167
Target Compounds						
3) L1 AR-1016-1	5.162	4.308	32730052	26671749	481.996	479.701
4) L1 AR-1016-2	5.184	4.327	47416937	37542752	468.221	484.526
5) L1 AR-1016-3	5.250	4.511	28674321	20795709	457.155	484.787
6) L1 AR-1016-4	5.356	4.560	23412895	16481990	450.996	488.997
7) L1 AR-1016-5	5.677	4.785	23606125	20224620	468.695	486.219
31) L7 AR-1260-1	6.906	5.894	38778272	38526153	441.264	480.272
32) L7 AR-1260-2	7.189	6.101	43157491	46684558	449.278	473.934
33) L7 AR-1260-3	7.584	6.263	29111599	43983523	448.699	474.221
34) L7 AR-1260-4	7.831	6.775	35116361	34316987	450.232	481.280
35) L7 AR-1260-5	8.174	7.043	63406988	77577248	452.403	469.826



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031022\
Data File : PP044469.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Mar 2022 12:41
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 10 22:50:09 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 03 02:21:53 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.910 min
Delta R.T.: -0.002 min
Response: 102514195
Conc: 49.48 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.166 min
Delta R.T.: 0.000 min
Response: 84647792
Conc: 48.39 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.042 min
Delta R.T.: -0.009 min
Response: 62265930
Conc: 45.24 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.528 min
Delta R.T.: -0.009 min
Response: 66217036
Conc: 43.17 ng/ml



#3 AR-1016-1


R.T.: 5.162 min
Delta R.T.: -0.004 min
Response: 32730052
Conc: 482.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.308 min
Delta R.T.: -0.004 min
Response: 26671749
Conc: 479.70 ng/ml




#4 AR-1016-2

R.T.: 5.184 min
Delta R.T.: -0.004 min
Response: 47416937
Conc: 468.22 ng/ml



#4 AR-1016-2


R.T.: 4.327 min
Delta R.T.: -0.004 min
Response: 37542752
Conc: 484.53 ng/ml



#5 AR-1016-3


R.T.: 5.250 min
Delta R.T.: -0.004 min
Response: 28674321
Conc: 457.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.511 min
Delta R.T.: -0.005 min
Response: 20795709
Conc: 484.79 ng/ml




#6 AR-1016-4

R.T.: 5.356 min
Delta R.T.: -0.004 min
Response: 23412895
Conc: 451.00 ng/ml



#6 AR-1016-4


R.T.: 4.560 min
Delta R.T.: -0.005 min
Response: 16481990
Conc: 489.00 ng/ml



#7 AR-1016-5


R.T.: 5.677 min
Delta R.T.: -0.004 min
Response: 23606125
Conc: 468.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 4.785 min
Delta R.T.: -0.005 min
Response: 20224620
Conc: 486.22 ng/ml




#31 AR-1260-1

R.T.: 6.906 min
Delta R.T.: -0.005 min
Response: 38778272
Conc: 441.26 ng/ml



#31 AR-1260-1


R.T.: 5.894 min
Delta R.T.: -0.006 min
Response: 38526153
Conc: 480.27 ng/ml



#32 AR-1260-2


R.T.: 7.189 min
Delta R.T.: -0.005 min
Response: 43157491
Conc: 449.28 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.101 min
Delta R.T.: -0.006 min
Response: 46684558
Conc: 473.93 ng/ml




#33 AR-1260-3

R.T.: 7.584 min
Delta R.T.: -0.006 min
Response: 29111599
Conc: 448.70 ng/ml



#33 AR-1260-3


R.T.: 6.263 min
Delta R.T.: -0.006 min
Response: 43983523
Conc: 474.22 ng/ml



#34 AR-1260-4


R.T.: 7.831 min
Delta R.T.: -0.005 min
Response: 35116361
Conc: 450.23 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 6.775 min
Delta R.T.: -0.007 min
Response: 34316987
Conc: 481.28 ng/ml



#35 AR-1260-5

R.T.: 8.174 min
Delta R.T.: -0.006 min
Response: 63406988
Conc: 452.40 ng/ml



#35 AR-1260-5

R.T.: 7.043 min
Delta R.T.: -0.007 min
Response: 77577248
Conc: 469.83 ng/ml