

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031023\
 Data File : PP056360.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Mar 2023 15:49
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 09 20:30:32 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 08 02:03:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.320	3.562	84045780	68511319	54.128	48.553
2) SA Decachlor...	10.061	8.558	91011384	54713608	45.818	48.945
Target Compounds						
3) L1 AR-1016-1	5.490	4.642	29875920	20788784	516.056	493.443
4) L1 AR-1016-2	5.512	4.660	43325164	29260404	505.735	482.773
5) L1 AR-1016-3	5.574	4.836	27456968	15970056	503.405	493.407
6) L1 AR-1016-4	5.673	4.878	22057010	13644167	507.971	486.156
7) L1 AR-1016-5	5.968	5.091	23007898	17858893	501.821	499.169
31) L7 AR-1260-1	7.100	6.124	46040519	31696328	467.187	464.953
32) L7 AR-1260-2	7.358	6.314	52956321	36101520	456.541	471.928
33) L7 AR-1260-3	7.718	6.467	40671180	34060782	475.122	471.039
34) L7 AR-1260-4	7.944	6.939	46088178	28633711	466.171	477.489
35) L7 AR-1260-5	8.260	7.183	92325849	59994879	461.986	469.517



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031023\
Data File : PP056360.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Mar 2023 15:49
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 09 20:30:32 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030823.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 08 02:03:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.320 min
Delta R.T.: 0.002 min
Response: 84045780
Conc: 54.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.562 min
Delta R.T.: 0.002 min
Response: 68511319
Conc: 48.55 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.061 min
Delta R.T.: -0.002 min
Response: 91011384
Conc: 45.82 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.558 min
Delta R.T.: -0.002 min
Response: 54713608
Conc: 48.94 ng/ml



#3 AR-1016-1


R.T.: 5.490 min
Delta R.T.: 0.000 min
Response: 29875920
Conc: 516.06 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.642 min
Delta R.T.: 0.000 min
Response: 20788784
Conc: 493.44 ng/ml




#4 AR-1016-2

R.T.: 5.512 min
Delta R.T.: 0.000 min
Response: 43325164
Conc: 505.73 ng/ml



#4 AR-1016-2


R.T.: 4.660 min
Delta R.T.: 0.000 min
Response: 29260404
Conc: 482.77 ng/ml



#5 AR-1016-3


R.T.: 5.574 min
Delta R.T.: 0.000 min
Response: 27456968
Conc: 503.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.836 min
Delta R.T.: 0.000 min
Response: 15970056
Conc: 493.41 ng/ml




#6 AR-1016-4

R.T.: 5.673 min
Delta R.T.: 0.000 min
Response: 22057010
Conc: 507.97 ng/ml



#6 AR-1016-4


R.T.: 4.878 min
Delta R.T.: 0.000 min
Response: 13644167
Conc: 486.16 ng/ml



#7 AR-1016-5


R.T.: 5.968 min
Delta R.T.: 0.000 min
Response: 23007898
Conc: 501.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.091 min
Delta R.T.: 0.000 min
Response: 17858893
Conc: 499.17 ng/ml




#31 AR-1260-1

R.T.: 7.100 min
Delta R.T.: 0.000 min
Response: 46040519
Conc: 467.19 ng/ml



#31 AR-1260-1


R.T.: 6.124 min
Delta R.T.: -0.001 min
Response: 31696328
Conc: 464.95 ng/ml



#32 AR-1260-2


R.T.: 7.358 min
Delta R.T.: 0.000 min
Response: 52956321
Conc: 456.54 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.314 min
Delta R.T.: -0.001 min
Response: 36101520
Conc: 471.93 ng/ml




#33 AR-1260-3

R.T.: 7.718 min
Delta R.T.: -0.002 min
Response: 40671180
Conc: 475.12 ng/ml



#33 AR-1260-3


R.T.: 6.467 min
Delta R.T.: 0.000 min
Response: 34060782
Conc: 471.04 ng/ml



#34 AR-1260-4


R.T.: 7.944 min
Delta R.T.: -0.002 min
Response: 46088178
Conc: 466.17 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 6.939 min
Delta R.T.: -0.001 min
Response: 28633711
Conc: 477.49 ng/ml



#35 AR-1260-5

R.T.: 8.260 min
Delta R.T.: -0.002 min
Response: 92325849
Conc: 461.99 ng/ml



#35 AR-1260-5

R.T.: 7.183 min
Delta R.T.: 0.000 min
Response: 59994879
Conc: 469.52 ng/ml