

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031121\
 Data File : PP033906.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 11 Mar 2021 19:38
 Operator : DD\AJ
 Sample : M1625-01MS
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 12 00:49:56 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030921.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 10 03:01:34 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.857	4.268	1973409	944260	22.807	23.243
2)	SA Decachlor...	10.785	9.594	1776929	609096	21.040	22.825
Target Compounds							
3)	L1 AR-1016-1	6.170	5.525	1429876	642787	536.712	564.764
4)	L1 AR-1016-2	6.193	5.546	2114119	928283	523.059	561.561
5)	L1 AR-1016-3	6.259	5.739	1310701	525584	529.666	573.586
6)	L1 AR-1016-4	6.365	5.788	1108040	402999	543.566	558.085
7)	L1 AR-1016-5	6.678	6.018	1037178	531355	511.676	574.689
8)	L2 AR-1221-1	5.106	4.524	363286	73951	393.680	173.669 #
9)	L2 AR-1221-2	5.204	4.625	199494	97405	283.906	301.023
10)	L2 AR-1221-3	5.291	4.714	804554	413756	382.244	422.108
11)	L3 AR-1232-1	5.291	4.714	804554	413756	465.032	509.295
12)	L3 AR-1232-2	5.876	5.546	1106986	928283	1208.712	1280.937
13)	L3 AR-1232-3	6.193	5.739	2114119	525584	1250.767	1366.717
14)	L3 AR-1232-4	6.365	5.833	1108040	504332	1297.093	1512.111
15)	L3 AR-1232-5	6.463	6.018	833088	531355	1404.316	1466.213
16)	L4 AR-1242-1	6.170	5.525	1429876	642787	704.700	727.949
17)	L4 AR-1242-2	6.193	5.546	2114119	928283	696.275	740.122
18)	L4 AR-1242-3	6.259	5.739	1310701	525584	686.005	744.314
19)	L4 AR-1242-4	6.365	5.833	1108040	504332	712.407	736.599
20)	L4 AR-1242-5	7.144	6.396	147507	402959	91.891	504.085 #
21)	L5 AR-1248-1	6.170	5.525	1429876	642787	916.610	944.295
22)	L5 AR-1248-2	6.463	5.788	833088	402999	391.367	429.512
23)	L5 AR-1248-3	6.678	5.833	1037178	504332	402.383	500.857
24)	L5 AR-1248-4	7.106	6.018	171905	531355	62.136	453.837 #
25)	L5 AR-1248-5	7.144	6.439	147507	57802	54.085	53.664
26)	L6 AR-1254-1	7.076	6.396	769782	402959	269.976	244.939
27)	L6 AR-1254-2	7.304	6.553	837046	372365	190.683	259.889 #
28)	L6 AR-1254-3	7.685	6.991	479902	666499	101.136	288.725 #
29)	L6 AR-1254-4	7.978	7.210	284477	68744	84.299	53.149 #
30)	L6 AR-1254-5	8.407	7.638	2940915	1184163	785.003	609.705
31)	L7 AR-1260-1	7.851	7.110	1851567	903907	539.173	591.457
32)	L7 AR-1260-2	8.117	7.303	2185333	1039626	530.776	574.774
33)	L7 AR-1260-3	8.481	7.460	1444422	964327	446.254	569.400 #
34)	L7 AR-1260-4	8.710	7.941	1853103	695748	500.275	473.073
35)	L7 AR-1260-5	9.029	8.184	3461110	1596647	447.391	491.079
36)	L8 AR-1262-1	8.481	7.638	1444422	1184163	303.569	1113.971 #
37)	L8 AR-1262-2	9.029	8.184	3461110	1596647	400.433	479.159
38)	L8 AR-1262-3	9.351f	8.470	2261841	318394	372.696	222.948 #
39)	L8 AR-1262-4	9.404f	8.532	1348430	1147418	272.041	455.111 #
40)	L8 AR-1262-5	10.067	9.031	955446	365401	290.291	320.561
41)	L9 AR-1268-1	9.351f	8.470	2261841	318394	217.227	83.243 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031121\
 Data File : PP033906.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 11 Mar 2021 19:38
 Operator : DD\AJ
 Sample : M1625-01MS
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 12 00:49:56 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030921.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 10 03:01:34 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.404f	8.532	1348430	1147418	133.418	321.142 #
43) L9	AR-1268-3	9.644	8.743	68110	34285	7.669	11.037 #
44) L9	AR-1268-4	10.067	9.031	955446	365401	274.354	301.064
45) L9	AR-1268-5	10.463	9.330	364527	125024	12.329	13.675


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031121\
Data File : PP033906.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 11 Mar 2021 19:38
Operator : DD\AJ
Sample : M1625-01MS
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 12 00:49:56 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030921.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 10 03:01:34 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.857 min
Delta R.T.: -0.001 min
Response: 1973409
Conc: 22.81 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.268 min
Delta R.T.: -0.003 min
Response: 944260
Conc: 23.24 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.785 min
Delta R.T.: -0.002 min
Response: 1776929
Conc: 21.04 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.594 min
Delta R.T.: -0.010 min
Response: 609096
Conc: 22.83 ng/ml



#3 AR-1016-1


R.T.: 6.170 min
Delta R.T.: -0.002 min
Response: 1429876
Conc: 536.71 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 5.525 min
Delta R.T.: -0.005 min
Response: 642787
Conc: 564.76 ng/ml




#4 AR-1016-2

R.T.: 6.193 min
Delta R.T.: -0.002 min
Response: 2114119
Conc: 523.06 ng/ml



#4 AR-1016-2


R.T.: 5.546 min
Delta R.T.: -0.005 min
Response: 928283
Conc: 561.56 ng/ml



#5 AR-1016-3


R.T.: 6.259 min
Delta R.T.: -0.002 min
Response: 1310701
Conc: 529.67 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 5.739 min
Delta R.T.: -0.005 min
Response: 525584
Conc: 573.59 ng/ml




#6 AR-1016-4

R.T.: 6.365 min
Delta R.T.: -0.001 min
Response: 1108040
Conc: 543.57 ng/ml



#6 AR-1016-4


R.T.: 5.788 min
Delta R.T.: -0.005 min
Response: 402999
Conc: 558.09 ng/ml



#7 AR-1016-5


R.T.: 6.678 min
Delta R.T.: -0.002 min
Response: 1037178
Conc: 511.68 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 6.018 min
Delta R.T.: -0.006 min
Response: 531355
Conc: 574.69 ng/ml




#8 AR-1221-1

R.T.: 5.106 min
Delta R.T.: -0.003 min
Response: 363286
Conc: 393.68 ng/ml



#8 AR-1221-1


R.T.: 4.524 min
Delta R.T.: -0.006 min
Response: 73951
Conc: 173.67 ng/ml



#9 AR-1221-2


R.T.: 5.204 min
Delta R.T.: -0.003 min
Response: 199494
Conc: 283.91 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 4.625 min
Delta R.T.: -0.004 min
Response: 97405
Conc: 301.02 ng/ml




#10 AR-1221-3

R.T.: 5.291 min
Delta R.T.: -0.002 min
Response: 804554
Conc: 382.24 ng/ml



#10 AR-1221-3


R.T.: 4.714 min
Delta R.T.: -0.005 min
Response: 413756
Conc: 422.11 ng/ml



#11 AR-1232-1


R.T.: 5.291 min
Delta R.T.: -0.001 min
Response: 804554
Conc: 465.03 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.714 min
Delta R.T.: -0.004 min
Response: 413756
Conc: 509.29 ng/ml




#12 AR-1232-2

R.T.: 5.876 min
Delta R.T.: -0.002 min
Response: 1106986
Conc: 1208.71 ng/ml



#12 AR-1232-2


R.T.: 5.546 min
Delta R.T.: -0.005 min
Response: 928283
Conc: 1280.94 ng/ml



#13 AR-1232-3


R.T.: 6.193 min
Delta R.T.: -0.003 min
Response: 2114119
Conc: 1250.77 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 5.739 min
Delta R.T.: -0.005 min
Response: 525584
Conc: 1366.72 ng/ml




#14 AR-1232-4

R.T.: 6.365 min
Delta R.T.: -0.002 min
Response: 1108040
Conc: 1297.09 ng/ml



#14 AR-1232-4


R.T.: 5.833 min
Delta R.T.: -0.005 min
Response: 504332
Conc: 1512.11 ng/ml



#15 AR-1232-5


R.T.: 6.463 min
Delta R.T.: -0.003 min
Response: 833088
Conc: 1404.32 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 6.018 min
Delta R.T.: -0.005 min
Response: 531355
Conc: 1466.21 ng/ml




#16 AR-1242-1

R.T.: 6.170 min
Delta R.T.: -0.002 min
Response: 1429876
Conc: 704.70 ng/ml



#16 AR-1242-1


R.T.: 5.525 min
Delta R.T.: -0.005 min
Response: 642787
Conc: 727.95 ng/ml



#17 AR-1242-2


R.T.: 6.193 min
Delta R.T.: -0.002 min
Response: 2114119
Conc: 696.27 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 5.546 min
Delta R.T.: -0.005 min
Response: 928283
Conc: 740.12 ng/ml




#18 AR-1242-3

R.T.: 6.259 min
Delta R.T.: -0.002 min
Response: 1310701
Conc: 686.01 ng/ml



#18 AR-1242-3


R.T.: 5.739 min
Delta R.T.: -0.005 min
Response: 525584
Conc: 744.31 ng/ml



#19 AR-1242-4


R.T.: 6.365 min
Delta R.T.: -0.002 min
Response: 1108040
Conc: 712.41 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.833 min
Delta R.T.: -0.005 min
Response: 504332
Conc: 736.60 ng/ml




#20 AR-1242-5

R.T.: 7.144 min
Delta R.T.: -0.002 min
Response: 147507
Conc: 91.89 ng/ml



#20 AR-1242-5


R.T.: 6.396 min
Delta R.T.: -0.005 min
Response: 402959
Conc: 504.08 ng/ml



#21 AR-1248-1


R.T.: 6.170 min
Delta R.T.: -0.002 min
Response: 1429876
Conc: 916.61 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 5.525 min
Delta R.T.: -0.005 min
Response: 642787
Conc: 944.29 ng/ml




#22 AR-1248-2

R.T.: 6.463 min
Delta R.T.: -0.002 min
Response: 833088
Conc: 391.37 ng/ml



#22 AR-1248-2


R.T.: 5.788 min
Delta R.T.: -0.006 min
Response: 402999
Conc: 429.51 ng/ml



#23 AR-1248-3


R.T.: 6.678 min
Delta R.T.: -0.001 min
Response: 1037178
Conc: 402.38 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.833 min
Delta R.T.: -0.006 min
Response: 504332
Conc: 500.86 ng/ml




#24 AR-1248-4

R.T.: 7.106 min
Delta R.T.: -0.001 min
Response: 171905
Conc: 62.14 ng/ml



#24 AR-1248-4


R.T.: 6.018 min
Delta R.T.: -0.006 min
Response: 531355
Conc: 453.84 ng/ml



#25 AR-1248-5


R.T.: 7.144 min
Delta R.T.: -0.002 min
Response: 147507
Conc: 54.08 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 6.439 min
Delta R.T.: -0.007 min
Response: 57802
Conc: 53.66 ng/ml




#26 AR-1254-1

R.T.: 7.076 min
Delta R.T.: -0.001 min
Response: 769782
Conc: 269.98 ng/ml



#26 AR-1254-1


R.T.: 6.396 min
Delta R.T.: -0.007 min
Response: 402959
Conc: 244.94 ng/ml



#27 AR-1254-2


R.T.: 7.304 min
Delta R.T.: -0.003 min
Response: 837046
Conc: 190.68 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 6.553 min
Delta R.T.: -0.007 min
Response: 372365
Conc: 259.89 ng/ml




#28 AR-1254-3

R.T.: 7.685 min
Delta R.T.: -0.001 min
Response: 479902
Conc: 101.14 ng/ml



#28 AR-1254-3


R.T.: 6.991 min
Delta R.T.: 0.011 min
Response: 666499
Conc: 288.72 ng/ml



#29 AR-1254-4


R.T.: 7.978 min
Delta R.T.: -0.004 min
Response: 284477
Conc: 84.30 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 7.210 min
Delta R.T.: -0.007 min
Response: 68744
Conc: 53.15 ng/ml




#30 AR-1254-5

R.T.: 8.407 min
Delta R.T.: -0.001 min
Response: 2940915
Conc: 785.00 ng/ml



#30 AR-1254-5


R.T.: 7.638 min
Delta R.T.: -0.008 min
Response: 1184163
Conc: 609.70 ng/ml



#31 AR-1260-1


R.T.: 7.851 min
Delta R.T.: 0.000 min
Response: 1851567
Conc: 539.17 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 7.110 min
Delta R.T.: -0.006 min
Response: 903907
Conc: 591.46 ng/ml




#32 AR-1260-2

R.T.: 8.117 min
Delta R.T.: 0.000 min
Response: 2185333
Conc: 530.78 ng/ml



#32 AR-1260-2


R.T.: 7.303 min
Delta R.T.: -0.006 min
Response: 1039626
Conc: 574.77 ng/ml



#33 AR-1260-3


R.T.: 8.481 min
Delta R.T.: -0.001 min
Response: 1444422
Conc: 446.25 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 7.460 min
Delta R.T.: -0.007 min
Response: 964327
Conc: 569.40 ng/ml




#34 AR-1260-4

R.T.: 8.710 min
Delta R.T.: -0.001 min
Response: 1853103
Conc: 500.27 ng/ml



#34 AR-1260-4


R.T.: 7.941 min
Delta R.T.: -0.007 min
Response: 695748
Conc: 473.07 ng/ml



#35 AR-1260-5


R.T.: 9.029 min
Delta R.T.: -0.002 min
Response: 3461110
Conc: 447.39 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 8.184 min
Delta R.T.: -0.007 min
Response: 1596647
Conc: 491.08 ng/ml




#36 AR-1262-1

R.T.: 8.481 min
Delta R.T.: -0.002 min
Response: 1444422
Conc: 303.57 ng/ml



#36 AR-1262-1


R.T.: 7.638 min
Delta R.T.: -0.006 min
Response: 1184163
Conc: 1113.97 ng/ml



#37 AR-1262-2


R.T.: 9.029 min
Delta R.T.: -0.002 min
Response: 3461110
Conc: 400.43 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 8.184 min
Delta R.T.: -0.006 min
Response: 1596647
Conc: 479.16 ng/ml




#38 AR-1262-3

R.T.: 9.351 min
Delta R.T.: 0.019 min
Response: 2261841
Conc: 372.70 ng/ml



#38 AR-1262-3


R.T.: 8.470 min
Delta R.T.: -0.006 min
Response: 318394
Conc: 222.95 ng/ml



#39 AR-1262-4


R.T.: 9.404 min
Delta R.T.: -0.020 min
Response: 1348430
Conc: 272.04 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 8.532 min
Delta R.T.: -0.008 min
Response: 1147418
Conc: 455.11 ng/ml




#40 AR-1262-5

R.T.: 10.067 min
Delta R.T.: 0.000 min
Response: 955446
Conc: 290.29 ng/ml



#40 AR-1262-5


R.T.: 9.031 min
Delta R.T.: -0.008 min
Response: 365401
Conc: 320.56 ng/ml



#41 AR-1268-1


R.T.: 9.351 min
Delta R.T.: 0.019 min
Response: 2261841
Conc: 217.23 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 8.470 min
Delta R.T.: -0.007 min
Response: 318394
Conc: 83.24 ng/ml




#42 AR-1268-2

R.T.: 9.404 min
Delta R.T.: -0.022 min
Response: 1348430
Conc: 133.42 ng/ml



#42 AR-1268-2


R.T.: 8.532 min
Delta R.T.: -0.009 min
Response: 1147418
Conc: 321.14 ng/ml



#43 AR-1268-3


R.T.: 9.644 min
Delta R.T.: 0.000 min
Response: 68110
Conc: 7.67 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 8.743 min
Delta R.T.: -0.008 min
Response: 34285
Conc: 11.04 ng/ml




#44 AR-1268-4

R.T.: 10.067 min
Delta R.T.: 0.000 min
Response: 955446
Conc: 274.35 ng/ml



#44 AR-1268-4


R.T.: 9.031 min
Delta R.T.: -0.007 min
Response: 365401
Conc: 301.06 ng/ml



#45 AR-1268-5

R.T.: 10.463 min
Delta R.T.: -0.003 min
Response: 364527
Conc: 12.33 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 9.330 min
Delta R.T.: -0.009 min
Response: 125024
Conc: 13.67 ng/ml