

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031121\
 Data File : PP033884.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 11 Mar 2021 13:23
 Operator : DD\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 12 00:44:10 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030921.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 10 03:01:34 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.857	4.267	4487645	2113790	51.865	52.031
2)	SA Decachlor...	10.787	9.595	4084138	1282393	48.359	48.056
Target Compounds							
3)	L1 AR-1016-1	6.170	5.525	1334190	553310	500.796	486.149
4)	L1 AR-1016-2	6.193	5.546	1946253	779997	481.527	471.856
5)	L1 AR-1016-3	6.259	5.739	1255763	450321	507.465	491.450
6)	L1 AR-1016-4	6.365	5.788	1036057	370505	508.254	513.086
7)	L1 AR-1016-5	6.680	6.018	1024568	474759	505.455	513.478
31)	L7 AR-1260-1	7.852	7.110	1724430	755751	502.151	494.513
32)	L7 AR-1260-2	8.117	7.304	1974298	874402	479.520	483.428
33)	L7 AR-1260-3	8.481	7.461	1620874	814066	500.769	480.676
34)	L7 AR-1260-4	8.711	7.942	1837712	686003	496.120	466.447
35)	L7 AR-1260-5	9.030	8.184	3551929	1501291	459.130	461.751



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031121\
Data File : PP033884.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 11 Mar 2021 13:23
Operator : DD\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 12 00:44:10 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030921.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 10 03:01:34 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.857 min
Delta R.T.: 0.000 min
Response: 4487645
Conc: 51.87 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.267 min
Delta R.T.: -0.004 min
Response: 2113790
Conc: 52.03 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.787 min
Delta R.T.: 0.000 min
Response: 4084138
Conc: 48.36 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.595 min
Delta R.T.: -0.009 min
Response: 1282393
Conc: 48.06 ng/ml



#3 AR-1016-1


R.T.: 6.170 min
Delta R.T.: -0.001 min
Response: 1334190
Conc: 500.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 5.525 min
Delta R.T.: -0.005 min
Response: 553310
Conc: 486.15 ng/ml




#4 AR-1016-2

R.T.: 6.193 min
Delta R.T.: -0.002 min
Response: 1946253
Conc: 481.53 ng/ml



#4 AR-1016-2


R.T.: 5.546 min
Delta R.T.: -0.006 min
Response: 779997
Conc: 471.86 ng/ml



#5 AR-1016-3


R.T.: 6.259 min
Delta R.T.: -0.002 min
Response: 1255763
Conc: 507.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.739 min
Delta R.T.: -0.006 min
Response: 450321
Conc: 491.45 ng/ml




#6 AR-1016-4

R.T.: 6.365 min
Delta R.T.: -0.001 min
Response: 1036057
Conc: 508.25 ng/ml



#6 AR-1016-4


R.T.: 5.788 min
Delta R.T.: -0.005 min
Response: 370505
Conc: 513.09 ng/ml



#7 AR-1016-5


R.T.: 6.680 min
Delta R.T.: 0.000 min
Response: 1024568
Conc: 505.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 6.018 min
Delta R.T.: -0.006 min
Response: 474759
Conc: 513.48 ng/ml




#31 AR-1260-1

R.T.: 7.852 min
Delta R.T.: 0.000 min
Response: 1724430
Conc: 502.15 ng/ml



#31 AR-1260-1


R.T.: 7.110 min
Delta R.T.: -0.006 min
Response: 755751
Conc: 494.51 ng/ml



#32 AR-1260-2


R.T.: 8.117 min
Delta R.T.: 0.000 min
Response: 1974298
Conc: 479.52 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 7.304 min
Delta R.T.: -0.006 min
Response: 874402
Conc: 483.43 ng/ml




#33 AR-1260-3

R.T.: 8.481 min
Delta R.T.: -0.001 min
Response: 1620874
Conc: 500.77 ng/ml



#33 AR-1260-3


R.T.: 7.461 min
Delta R.T.: -0.007 min
Response: 814066
Conc: 480.68 ng/ml



#34 AR-1260-4


R.T.: 8.711 min
Delta R.T.: 0.000 min
Response: 1837712
Conc: 496.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.942 min
Delta R.T.: -0.006 min
Response: 686003
Conc: 466.45 ng/ml



#35 AR-1260-5

R.T.: 9.030 min
Delta R.T.: 0.000 min
Response: 3551929
Conc: 459.13 ng/ml



#35 AR-1260-5

R.T.: 8.184 min
Delta R.T.: -0.006 min
Response: 1501291
Conc: 461.75 ng/ml