

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031122\
 Data File : PP044551.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Mar 2022 11:21
 Operator : YP\AJ
 Sample : N1645-03
 Misc : AR1262 MDL 25 PPB
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 12 05:54:46 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 03 02:21:53 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.908	3.164	34580381	30503116	16.690	17.436
2)	SA Decachlor...	10.041	8.527	23420204	26857784	17.016	17.509
Target Compounds							
3)	L1 AR-1016-1	5.161	4.307	94516	71184	1.392	1.280
4)	L1 AR-1016-2	5.187	4.325	55346	105475	0.547	1.361 #
5)	L1 AR-1016-3	5.253	4.512	165027	153585	2.631	3.580 #
6)	L1 AR-1016-4	5.354	4.566	28025	52463	0.540	1.557 #
7)	L1 AR-1016-5	5.688	4.789	117571	163024	2.334	3.919 #
8)	L2 AR-1221-1	4.127	3.391	872368	13577	35.305	0.591 #
9)	L2 AR-1221-2	4.230	3.475	614200	618418	32.736	35.804
10)	L2 AR-1221-3	4.315	3.569	110313	21039	1.971	0.423 #
11)	L3 AR-1232-1	4.315	3.569	110313	21039	2.404	0.506 #
12)	L3 AR-1232-2	4.874	4.325	70607	105475	3.138	3.054
13)	L3 AR-1232-3	5.187	4.512	55346	153585	1.279	8.232 #
14)	L3 AR-1232-4	5.366	4.601	22380	121522	1.115	7.370 #
15)	L3 AR-1232-5	5.463	4.789	56356	163024	3.715	9.006 #
16)	L4 AR-1242-1	5.161	4.307	94516	71184	1.736	1.552
17)	L4 AR-1242-2	5.187	4.325	55346	105475	0.695	1.669 #
18)	L4 AR-1242-3	5.253	4.512	165027	153585	3.495	4.300
19)	L4 AR-1242-4	5.366	4.601	22380	121522	0.578	3.513 #
20)	L4 AR-1242-5	6.160	5.161	198671	470201	4.660	11.132 #
21)	L5 AR-1248-1	5.161	4.307	94516	71184	2.286	2.007
22)	L5 AR-1248-2	5.463	4.566	56356	52463	1.005	1.115
23)	L5 AR-1248-3	5.688	4.601	117571	121522	1.721	2.466 #
24)	L5 AR-1248-4	6.123	4.789	93215	163024	1.278	2.884 #
25)	L5 AR-1248-5	6.160	5.235f	198671	120559	2.705	2.149
26)	L6 AR-1254-1	6.086	5.161	141484	470201	1.886	5.687 #
27)	L6 AR-1254-2	6.322	5.319	426478	644349	3.796	8.957 #
28)	L6 AR-1254-3	6.726	5.768	486329	1276208	4.156	11.077 #
29)	L6 AR-1254-4	7.042	6.001	57126	122033	0.711	1.621 #
30)	L6 AR-1254-5	7.495	6.451	1664451	1347627	19.122	12.366 #
31)	L7 AR-1260-1	6.904	5.891	1551900	1799136	17.659	22.428 #
32)	L7 AR-1260-2	7.188	6.098	1552637	2192493	16.163	22.258 #
33)	L7 AR-1260-3	7.583	6.262	2308027	1867450	35.574	20.134 #
34)	L7 AR-1260-4	7.832	6.773	2666304	2325569	34.185	32.615
35)	L7 AR-1260-5	8.173	7.041	4458528	4497136	31.811	27.236
36)	L8 AR-1262-1	7.583	6.494	2308027	2707594	22.836	26.090
37)	L8 AR-1262-2	8.173	6.773	4458528	2325569	26.852	24.298
38)	L8 AR-1262-3	8.504	7.348	2359037	1876105	21.078	23.782
39)	L8 AR-1262-4	8.596	7.418	854066	3434245	17.205	23.779 #
40)	L8 AR-1262-5	9.280	7.960	1040821	1749537	18.236	24.437 #
41)	L9 AR-1268-1	8.504	7.348	2359037	1876105	12.078	8.249 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031122\
 Data File : PP044551.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Mar 2022 11:21
 Operator : YP\AJ
 Sample : N1645-03
 Misc : AR1262 MDL 25 PPB
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 12 05:54:46 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 03 02:21:53 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.596	7.418	854066	3434245	4.689	16.764 #
43) L9	AR-1268-3	8.851	7.638	43055	183757	0.279	1.051 #
44) L9	AR-1268-4	9.280	7.960	1040821	1749537	16.635	21.560 #
45) L9	AR-1268-5	9.702	8.265	680376	737101	1.449	1.368



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031122\
Data File : PP044551.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Mar 2022 11:21
Operator : YP\AJ
Sample : N1645-03
Misc : AR1262 MDL 25 PPB
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 12 05:54:46 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 03 02:21:53 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.908 min
Delta R.T.: -0.004 min
Response: 34580381
Conc: 16.69 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.164 min
Delta R.T.: -0.002 min
Response: 30503116
Conc: 17.44 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.041 min
Delta R.T.: -0.011 min
Response: 23420204
Conc: 17.02 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.527 min
Delta R.T.: -0.010 min
Response: 26857784
Conc: 17.51 ng/ml



#3 AR-1016-1


R.T.: 5.161 min
Delta R.T.: -0.004 min
Response: 94516
Conc: 1.39 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.307 min
Delta R.T.: -0.006 min
Response: 71184
Conc: 1.28 ng/ml




#4 AR-1016-2

R.T.: 5.187 min
Delta R.T.: 0.000 min
Response: 55346
Conc: 0.55 ng/ml



#4 AR-1016-2


R.T.: 4.325 min
Delta R.T.: -0.006 min
Response: 105475
Conc: 1.36 ng/ml



#5 AR-1016-3


R.T.: 5.253 min
Delta R.T.: -0.001 min
Response: 165027
Conc: 2.63 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.512 min
Delta R.T.: -0.004 min
Response: 153585
Conc: 3.58 ng/ml




#6 AR-1016-4

R.T.: 5.354 min
Delta R.T.: -0.006 min
Response: 28025
Conc: 0.54 ng/ml



#6 AR-1016-4


R.T.: 4.566 min
Delta R.T.: 0.000 min
Response: 52463
Conc: 1.56 ng/ml



#7 AR-1016-5


R.T.: 5.688 min
Delta R.T.: 0.007 min
Response: 117571
Conc: 2.33 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 4.789 min
Delta R.T.: 0.000 min
Response: 163024
Conc: 3.92 ng/ml




#8 AR-1221-1

R.T.: 4.127 min
Delta R.T.: -0.011 min
Response: 872368
Conc: 35.30 ng/ml



#8 AR-1221-1


R.T.: 3.391 min
Delta R.T.: -0.004 min
Response: 13577
Conc: 0.59 ng/ml



#9 AR-1221-2


R.T.: 4.230 min
Delta R.T.: 0.000 min
Response: 614200
Conc: 32.74 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.475 min
Delta R.T.: -0.009 min
Response: 618418
Conc: 35.80 ng/ml




#10 AR-1221-3

R.T.: 4.315 min
Delta R.T.: 0.004 min
Response: 110313
Conc: 1.97 ng/ml



#10 AR-1221-3


R.T.: 3.569 min
Delta R.T.: 0.006 min
Response: 21039
Conc: 0.42 ng/ml



#11 AR-1232-1


R.T.: 4.315 min
Delta R.T.: 0.003 min
Response: 110313
Conc: 2.40 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 3.569 min
Delta R.T.: 0.007 min
Response: 21039
Conc: 0.51 ng/ml




#12 AR-1232-2

R.T.: 4.874 min
Delta R.T.: -0.001 min
Response: 70607
Conc: 3.14 ng/ml



#12 AR-1232-2


R.T.: 4.325 min
Delta R.T.: -0.006 min
Response: 105475
Conc: 3.05 ng/ml



#13 AR-1232-3


R.T.: 5.187 min
Delta R.T.: -0.001 min
Response: 55346
Conc: 1.28 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.512 min
Delta R.T.: -0.003 min
Response: 153585
Conc: 8.23 ng/ml




#14 AR-1232-4

R.T.: 5.366 min
Delta R.T.: 0.005 min
Response: 22380
Conc: 1.11 ng/ml



#14 AR-1232-4


R.T.: 4.601 min
Delta R.T.: -0.007 min
Response: 121522
Conc: 7.37 ng/ml



#15 AR-1232-5


R.T.: 5.463 min
Delta R.T.: 0.000 min
Response: 56356
Conc: 3.71 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 4.789 min
Delta R.T.: 0.000 min
Response: 163024
Conc: 9.01 ng/ml




#16 AR-1242-1

R.T.: 5.161 min
Delta R.T.: -0.005 min
Response: 94516
Conc: 1.74 ng/ml



#16 AR-1242-1


R.T.: 4.307 min
Delta R.T.: -0.005 min
Response: 71184
Conc: 1.55 ng/ml



#17 AR-1242-2


R.T.: 5.187 min
Delta R.T.: -0.001 min
Response: 55346
Conc: 0.69 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.325 min
Delta R.T.: -0.006 min
Response: 105475
Conc: 1.67 ng/ml




#18 AR-1242-3

R.T.: 5.253 min
Delta R.T.: -0.002 min
Response: 165027
Conc: 3.49 ng/ml



#18 AR-1242-3


R.T.: 4.512 min
Delta R.T.: -0.003 min
Response: 153585
Conc: 4.30 ng/ml



#19 AR-1242-4


R.T.: 5.366 min
Delta R.T.: 0.006 min
Response: 22380
Conc: 0.58 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.601 min
Delta R.T.: -0.007 min
Response: 121522
Conc: 3.51 ng/ml




#20 AR-1242-5

R.T.: 6.160 min
Delta R.T.: -0.002 min
Response: 198671
Conc: 4.66 ng/ml



#20 AR-1242-5


R.T.: 5.161 min
Delta R.T.: -0.006 min
Response: 470201
Conc: 11.13 ng/ml



#21 AR-1248-1


R.T.: 5.161 min
Delta R.T.: -0.005 min
Response: 94516
Conc: 2.29 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.307 min
Delta R.T.: -0.006 min
Response: 71184
Conc: 2.01 ng/ml




#22 AR-1248-2

R.T.: 5.463 min
Delta R.T.: 0.001 min
Response: 56356
Conc: 1.00 ng/ml



#22 AR-1248-2


R.T.: 4.566 min
Delta R.T.: 0.000 min
Response: 52463
Conc: 1.11 ng/ml



#23 AR-1248-3


R.T.: 5.688 min
Delta R.T.: 0.007 min
Response: 117571
Conc: 1.72 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.601 min
Delta R.T.: -0.007 min
Response: 121522
Conc: 2.47 ng/ml




#24 AR-1248-4

R.T.: 6.123 min
Delta R.T.: 0.002 min
Response: 93215
Conc: 1.28 ng/ml



#24 AR-1248-4


R.T.: 4.789 min
Delta R.T.: 0.000 min
Response: 163024
Conc: 2.88 ng/ml



#25 AR-1248-5


R.T.: 6.160 min
Delta R.T.: -0.002 min
Response: 198671
Conc: 2.70 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.235 min
Delta R.T.: 0.024 min
Response: 120559
Conc: 2.15 ng/ml




#26 AR-1254-1

R.T.: 6.086 min
Delta R.T.: -0.005 min
Response: 141484
Conc: 1.89 ng/ml



#26 AR-1254-1


R.T.: 5.161 min
Delta R.T.: -0.006 min
Response: 470201
Conc: 5.69 ng/ml



#27 AR-1254-2


R.T.: 6.322 min
Delta R.T.: -0.009 min
Response: 426478
Conc: 3.80 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.319 min
Delta R.T.: -0.008 min
Response: 644349
Conc: 8.96 ng/ml




#28 AR-1254-3

R.T.: 6.726 min
Delta R.T.: -0.004 min
Response: 486329
Conc: 4.16 ng/ml



#28 AR-1254-3


R.T.: 5.768 min
Delta R.T.: 0.010 min
Response: 1276208
Conc: 11.08 ng/ml



#29 AR-1254-4


R.T.: 7.042 min
Delta R.T.: -0.002 min
Response: 57126
Conc: 0.71 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.001 min
Delta R.T.: -0.008 min
Response: 122033
Conc: 1.62 ng/ml




#30 AR-1254-5

R.T.: 7.495 min
Delta R.T.: -0.011 min
Response: 1664451
Conc: 19.12 ng/ml



#30 AR-1254-5


R.T.: 6.451 min
Delta R.T.: -0.010 min
Response: 1347627
Conc: 12.37 ng/ml



#31 AR-1260-1


R.T.: 6.904 min
Delta R.T.: -0.007 min
Response: 1551900
Conc: 17.66 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 5.891 min
Delta R.T.: -0.009 min
Response: 1799136
Conc: 22.43 ng/ml




#32 AR-1260-2

R.T.: 7.188 min
Delta R.T.: -0.007 min
Response: 1552637
Conc: 16.16 ng/ml



#32 AR-1260-2


R.T.: 6.098 min
Delta R.T.: -0.009 min
Response: 2192493
Conc: 22.26 ng/ml



#33 AR-1260-3


R.T.: 7.583 min
Delta R.T.: -0.006 min
Response: 2308027
Conc: 35.57 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.262 min
Delta R.T.: -0.007 min
Response: 1867450
Conc: 20.13 ng/ml




#34 AR-1260-4

R.T.: 7.832 min
Delta R.T.: -0.005 min
Response: 2666304
Conc: 34.19 ng/ml



#34 AR-1260-4


R.T.: 6.773 min
Delta R.T.: -0.009 min
Response: 2325569
Conc: 32.62 ng/ml



#35 AR-1260-5


R.T.: 8.173 min
Delta R.T.: -0.008 min
Response: 4458528
Conc: 31.81 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.041 min
Delta R.T.: -0.009 min
Response: 4497136
Conc: 27.24 ng/ml




#36 AR-1262-1

R.T.: 7.583 min
Delta R.T.: -0.007 min
Response: 2308027
Conc: 22.84 ng/ml



#36 AR-1262-1


R.T.: 6.494 min
Delta R.T.: -0.008 min
Response: 2707594
Conc: 26.09 ng/ml



#37 AR-1262-2


R.T.: 8.173 min
Delta R.T.: -0.007 min
Response: 4458528
Conc: 26.85 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 6.773 min
Delta R.T.: -0.008 min
Response: 2325569
Conc: 24.30 ng/ml




#38 AR-1262-3

R.T.: 8.504 min
Delta R.T.: -0.008 min
Response: 2359037
Conc: 21.08 ng/ml



#38 AR-1262-3


R.T.: 7.348 min
Delta R.T.: -0.008 min
Response: 1876105
Conc: 23.78 ng/ml



#39 AR-1262-4


R.T.: 8.596 min
Delta R.T.: -0.009 min
Response: 854066
Conc: 17.21 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.418 min
Delta R.T.: -0.009 min
Response: 3434245
Conc: 23.78 ng/ml




#40 AR-1262-5

R.T.: 9.280 min
Delta R.T.: -0.007 min
Response: 1040821
Conc: 18.24 ng/ml



#40 AR-1262-5


R.T.: 7.960 min
Delta R.T.: -0.009 min
Response: 1749537
Conc: 24.44 ng/ml



#41 AR-1268-1


R.T.: 8.504 min
Delta R.T.: -0.005 min
Response: 2359037
Conc: 12.08 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.348 min
Delta R.T.: -0.009 min
Response: 1876105
Conc: 8.25 ng/ml




#42 AR-1268-2

R.T.: 8.596 min
Delta R.T.: -0.013 min
Response: 854066
Conc: 4.69 ng/ml



#42 AR-1268-2


R.T.: 7.418 min
Delta R.T.: -0.009 min
Response: 3434245
Conc: 16.76 ng/ml



#43 AR-1268-3


R.T.: 8.851 min
Delta R.T.: 0.007 min
Response: 43055
Conc: 0.28 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.638 min
Delta R.T.: -0.011 min
Response: 183757
Conc: 1.05 ng/ml




#44 AR-1268-4

R.T.: 9.280 min
Delta R.T.: -0.007 min
Response: 1040821
Conc: 16.64 ng/ml



#44 AR-1268-4


R.T.: 7.960 min
Delta R.T.: -0.010 min
Response: 1749537
Conc: 21.56 ng/ml



#45 AR-1268-5

R.T.: 9.702 min
Delta R.T.: -0.008 min
Response: 680376
Conc: 1.45 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.265 min
Delta R.T.: -0.009 min
Response: 737101
Conc: 1.37 ng/ml