

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031122\
 Data File : PP044632.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Mar 2022 11:48
 Operator : YP\AJ
 Sample : N1893-02 10X
 Misc :
 ALS Vial : 65 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 13 22:12:44 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 03 02:21:53 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.908	3.163	2869132	2832444	1.385	1.619
2)	SA Decachlor...	10.041	8.527	3407813	3364490	2.476	2.193
Target Compounds							
3)	L1 AR-1016-1	5.159	4.314	80634	29740	1.187	0.535 #
4)	L1 AR-1016-2	5.184	4.329	95937	48000	0.947	0.619 #
5)	L1 AR-1016-3	5.257	4.515	36882	12003	0.588	0.280 #
6)	L1 AR-1016-4	5.362	4.569	21208	29175	0.409	0.866 #
7)	L1 AR-1016-5	5.671	4.796	77421	24278	1.537	0.584 #
8)	L2 AR-1221-1	4.139	3.397	84236	4042	3.409	0.176 #
9)	L2 AR-1221-2	4.229	3.484	54962	29831	2.929	1.727 #
10)	L2 AR-1221-3	4.311	3.563	89021	51246	1.591	1.031 #
11)	L3 AR-1232-1	4.311	3.563	89021	51246	1.940	1.233 #
12)	L3 AR-1232-2	4.870	4.329	52898	48000	2.351	1.390 #
13)	L3 AR-1232-3	5.184	4.515	95937	12003	2.217	0.643 #
14)	L3 AR-1232-4	5.362	4.595	21208	146252	1.056	8.870 #
15)	L3 AR-1232-5	5.466	4.783	16140	13149	1.064	0.726 #
16)	L4 AR-1242-1	5.159	4.314	80634	29740	1.481	0.649 #
17)	L4 AR-1242-2	5.184	4.329	95937	48000	1.205	0.760 #
18)	L4 AR-1242-3	5.257	4.515	36882	12003	0.781	0.336 #
19)	L4 AR-1242-4	5.362	4.595	21208	146252	0.548	4.227 #
20)	L4 AR-1242-5	6.156	5.157	132081	22687	3.098	0.537 #
21)	L5 AR-1248-1	5.159	4.314	80634	29740	1.951	0.838 #
22)	L5 AR-1248-2	5.466	4.569	16140	29175	0.288	0.620 #
23)	L5 AR-1248-3	5.671	4.595	77421	146252	1.134	2.968 #
24)	L5 AR-1248-4	6.124	4.796	280528	24278	3.847	0.430 #
25)	L5 AR-1248-5	6.156	5.207	132081	32933	1.798	0.587 #
26)	L6 AR-1254-1	6.089	5.157	226810	22687	3.024	0.274 #
27)	L6 AR-1254-2	6.331	5.333	64530	5044	0.574	0.070 #
28)	L6 AR-1254-3	6.727	5.753	325210	35458	2.779	0.308 #
29)	L6 AR-1254-4	7.040	6.012	79478	15241	0.989	0.202 #
30)	L6 AR-1254-5	7.514	6.458	111124	104863	1.277	0.962
31)	L7 AR-1260-1	6.911	5.888	42651	655884	0.485	8.176 #
32)	L7 AR-1260-2	7.177f	6.106	44667	68360	0.465	0.694 #
33)	L7 AR-1260-3	7.599	6.271	179366	640835	2.765	6.909 #
34)	L7 AR-1260-4	7.837	6.780	170379	6536	2.184	0.092 #
35)	L7 AR-1260-5	8.179	7.049	2640030	55899	18.836	0.339 #
36)	L8 AR-1262-1	7.599	6.496	179366	38826	1.775	0.374 #
37)	L8 AR-1262-2	8.179	6.780	2640030	6536	15.900	0.068 #
38)	L8 AR-1262-3	8.514	7.342	41343	6682	0.369	0.085 #
39)	L8 AR-1262-4	8.600	7.447f	246562	616955	4.967	4.272
40)	L8 AR-1262-5	9.290	7.970	44062	86367	0.772	1.206 #
41)	L9 AR-1268-1	8.506	7.342	79163	6682	0.405	0.029 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031122\
 Data File : PP044632.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Mar 2022 11:48
 Operator : YP\AJ
 Sample : N1893-02 10X
 Misc :
 ALS Vial : 65 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 13 22:12:44 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 03 02:21:53 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.614	7.447f	43070	616955	0.236	3.012 #
43) L9	AR-1268-3	8.841	7.645	43572	5775	0.282	0.033 #
44) L9	AR-1268-4	9.290	7.970	44062	86367	0.704	1.064 #
45) L9	AR-1268-5	9.706	8.276	61212	23406	0.130	0.043 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031122\
Data File : PP044632.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Mar 2022 11:48
Operator : YP\AJ
Sample : N1893-02 10X
Misc :
ALS Vial : 65 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 13 22:12:44 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 03 02:21:53 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.908 min
Delta R.T.: -0.004 min
Response: 2869132
Conc: 1.38 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.163 min
Delta R.T.: -0.003 min
Response: 2832444
Conc: 1.62 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.041 min
Delta R.T.: -0.011 min
Response: 3407813
Conc: 2.48 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.527 min
Delta R.T.: -0.009 min
Response: 3364490
Conc: 2.19 ng/ml



#3 AR-1016-1


R.T.: 5.159 min
Delta R.T.: -0.006 min
Response: 80634
Conc: 1.19 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.314 min
Delta R.T.: 0.001 min
Response: 29740
Conc: 0.53 ng/ml




#4 AR-1016-2

R.T.: 5.184 min
Delta R.T.: -0.005 min
Response: 95937
Conc: 0.95 ng/ml



#4 AR-1016-2


R.T.: 4.329 min
Delta R.T.: -0.002 min
Response: 48000
Conc: 0.62 ng/ml



#5 AR-1016-3


R.T.: 5.257 min
Delta R.T.: 0.002 min
Response: 36882
Conc: 0.59 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.515 min
Delta R.T.: 0.000 min
Response: 12003
Conc: 0.28 ng/ml




#6 AR-1016-4

R.T.: 5.362 min
Delta R.T.: 0.002 min
Response: 21208
Conc: 0.41 ng/ml



#6 AR-1016-4


R.T.: 4.569 min
Delta R.T.: 0.003 min
Response: 29175
Conc: 0.87 ng/ml



#7 AR-1016-5


R.T.: 5.671 min
Delta R.T.: -0.010 min
Response: 77421
Conc: 1.54 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 4.796 min
Delta R.T.: 0.007 min
Response: 24278
Conc: 0.58 ng/ml




#8 AR-1221-1

R.T.: 4.139 min
Delta R.T.: 0.001 min
Response: 84236
Conc: 3.41 ng/ml



#8 AR-1221-1


R.T.: 3.397 min
Delta R.T.: 0.003 min
Response: 4042
Conc: 0.18 ng/ml



#9 AR-1221-2


R.T.: 4.229 min
Delta R.T.: -0.001 min
Response: 54962
Conc: 2.93 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.484 min
Delta R.T.: 0.000 min
Response: 29831
Conc: 1.73 ng/ml




#10 AR-1221-3

R.T.: 4.311 min
Delta R.T.: 0.000 min
Response: 89021
Conc: 1.59 ng/ml



#10 AR-1221-3


R.T.: 3.563 min
Delta R.T.: 0.000 min
Response: 51246
Conc: 1.03 ng/ml



#11 AR-1232-1


R.T.: 4.311 min
Delta R.T.: 0.000 min
Response: 89021
Conc: 1.94 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 3.563 min
Delta R.T.: 0.000 min
Response: 51246
Conc: 1.23 ng/ml




#12 AR-1232-2

R.T.: 4.870 min
Delta R.T.: -0.005 min
Response: 52898
Conc: 2.35 ng/ml



#12 AR-1232-2


R.T.: 4.329 min
Delta R.T.: -0.002 min
Response: 48000
Conc: 1.39 ng/ml



#13 AR-1232-3


R.T.: 5.184 min
Delta R.T.: -0.005 min
Response: 95937
Conc: 2.22 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.515 min
Delta R.T.: 0.000 min
Response: 12003
Conc: 0.64 ng/ml




#14 AR-1232-4

R.T.: 5.362 min
Delta R.T.: 0.000 min
Response: 21208
Conc: 1.06 ng/ml



#14 AR-1232-4


R.T.: 4.595 min
Delta R.T.: -0.013 min
Response: 146252
Conc: 8.87 ng/ml



#15 AR-1232-5


R.T.: 5.466 min
Delta R.T.: 0.003 min
Response: 16140
Conc: 1.06 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 4.783 min
Delta R.T.: -0.007 min
Response: 13149
Conc: 0.73 ng/ml




#16 AR-1242-1

R.T.: 5.159 min
Delta R.T.: -0.007 min
Response: 80634
Conc: 1.48 ng/ml



#16 AR-1242-1


R.T.: 4.314 min
Delta R.T.: 0.002 min
Response: 29740
Conc: 0.65 ng/ml



#17 AR-1242-2


R.T.: 5.184 min
Delta R.T.: -0.005 min
Response: 95937
Conc: 1.20 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.329 min
Delta R.T.: -0.002 min
Response: 48000
Conc: 0.76 ng/ml




#18 AR-1242-3

R.T.: 5.257 min
Delta R.T.: 0.001 min
Response: 36882
Conc: 0.78 ng/ml



#18 AR-1242-3


R.T.: 4.515 min
Delta R.T.: 0.000 min
Response: 12003
Conc: 0.34 ng/ml



#19 AR-1242-4


R.T.: 5.362 min
Delta R.T.: 0.001 min
Response: 21208
Conc: 0.55 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.595 min
Delta R.T.: -0.013 min
Response: 146252
Conc: 4.23 ng/ml




#20 AR-1242-5

R.T.: 6.156 min
Delta R.T.: -0.005 min
Response: 132081
Conc: 3.10 ng/ml



#20 AR-1242-5


R.T.: 5.157 min
Delta R.T.: -0.010 min
Response: 22687
Conc: 0.54 ng/ml



#21 AR-1248-1


R.T.: 5.159 min
Delta R.T.: -0.007 min
Response: 80634
Conc: 1.95 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.314 min
Delta R.T.: 0.001 min
Response: 29740
Conc: 0.84 ng/ml




#22 AR-1248-2

R.T.: 5.466 min
Delta R.T.: 0.004 min
Response: 16140
Conc: 0.29 ng/ml



#22 AR-1248-2


R.T.: 4.569 min
Delta R.T.: 0.004 min
Response: 29175
Conc: 0.62 ng/ml



#23 AR-1248-3


R.T.: 5.671 min
Delta R.T.: -0.010 min
Response: 77421
Conc: 1.13 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.595 min
Delta R.T.: -0.013 min
Response: 146252
Conc: 2.97 ng/ml




#24 AR-1248-4

R.T.: 6.124 min
Delta R.T.: 0.004 min
Response: 280528
Conc: 3.85 ng/ml



#24 AR-1248-4


R.T.: 4.796 min
Delta R.T.: 0.007 min
Response: 24278
Conc: 0.43 ng/ml



#25 AR-1248-5


R.T.: 6.156 min
Delta R.T.: -0.005 min
Response: 132081
Conc: 1.80 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.207 min
Delta R.T.: -0.003 min
Response: 32933
Conc: 0.59 ng/ml




#26 AR-1254-1

R.T.: 6.089 min
Delta R.T.: -0.002 min
Response: 226810
Conc: 3.02 ng/ml



#26 AR-1254-1


R.T.: 5.157 min
Delta R.T.: -0.010 min
Response: 22687
Conc: 0.27 ng/ml



#27 AR-1254-2


R.T.: 6.331 min
Delta R.T.: 0.000 min
Response: 64530
Conc: 0.57 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.333 min
Delta R.T.: 0.006 min
Response: 5044
Conc: 0.07 ng/ml




#28 AR-1254-3

R.T.: 6.727 min
Delta R.T.: -0.004 min
Response: 325210
Conc: 2.78 ng/ml



#28 AR-1254-3


R.T.: 5.753 min
Delta R.T.: -0.006 min
Response: 35458
Conc: 0.31 ng/ml



#29 AR-1254-4


R.T.: 7.040 min
Delta R.T.: -0.004 min
Response: 79478
Conc: 0.99 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.012 min
Delta R.T.: 0.003 min
Response: 15241
Conc: 0.20 ng/ml




#30 AR-1254-5

R.T.: 7.514 min
Delta R.T.: 0.008 min
Response: 111124
Conc: 1.28 ng/ml



#30 AR-1254-5


R.T.: 6.458 min
Delta R.T.: -0.003 min
Response: 104863
Conc: 0.96 ng/ml



#31 AR-1260-1


R.T.: 6.911 min
Delta R.T.: 0.000 min
Response: 42651
Conc: 0.49 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 5.888 min
Delta R.T.: -0.012 min
Response: 655884
Conc: 8.18 ng/ml




#32 AR-1260-2

R.T.: 7.177 min
Delta R.T.: -0.017 min
Response: 44667
Conc: 0.46 ng/ml



#32 AR-1260-2


R.T.: 6.106 min
Delta R.T.: -0.001 min
Response: 68360
Conc: 0.69 ng/ml



#33 AR-1260-3


R.T.: 7.599 min
Delta R.T.: 0.009 min
Response: 179366
Conc: 2.76 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.271 min
Delta R.T.: 0.002 min
Response: 640835
Conc: 6.91 ng/ml




#34 AR-1260-4

R.T.: 7.837 min
Delta R.T.: 0.000 min
Response: 170379
Conc: 2.18 ng/ml



#34 AR-1260-4


R.T.: 6.780 min
Delta R.T.: -0.003 min
Response: 6536
Conc: 0.09 ng/ml



#35 AR-1260-5


R.T.: 8.179 min
Delta R.T.: 0.000 min
Response: 2640030
Conc: 18.84 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.049 min
Delta R.T.: -0.001 min
Response: 55899
Conc: 0.34 ng/ml




#36 AR-1262-1

R.T.: 7.599 min
Delta R.T.: 0.009 min
Response: 179366
Conc: 1.77 ng/ml



#36 AR-1262-1


R.T.: 6.496 min
Delta R.T.: -0.006 min
Response: 38826
Conc: 0.37 ng/ml



#37 AR-1262-2


R.T.: 8.179 min
Delta R.T.: 0.000 min
Response: 2640030
Conc: 15.90 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 6.780 min
Delta R.T.: -0.002 min
Response: 6536
Conc: 0.07 ng/ml




#38 AR-1262-3

R.T.: 8.514 min
Delta R.T.: 0.002 min
Response: 41343
Conc: 0.37 ng/ml



#38 AR-1262-3


R.T.: 7.342 min
Delta R.T.: -0.014 min
Response: 6682
Conc: 0.08 ng/ml



#39 AR-1262-4


R.T.: 8.600 min
Delta R.T.: -0.004 min
Response: 246562
Conc: 4.97 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.447 min
Delta R.T.: 0.020 min
Response: 616955
Conc: 4.27 ng/ml




#40 AR-1262-5

R.T.: 9.290 min
Delta R.T.: 0.003 min
Response: 44062
Conc: 0.77 ng/ml



#40 AR-1262-5


R.T.: 7.970 min
Delta R.T.: 0.002 min
Response: 86367
Conc: 1.21 ng/ml



#41 AR-1268-1


R.T.: 8.506 min
Delta R.T.: -0.003 min
Response: 79163
Conc: 0.41 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.342 min
Delta R.T.: -0.014 min
Response: 6682
Conc: 0.03 ng/ml




#42 AR-1268-2

R.T.: 8.614 min
Delta R.T.: 0.006 min
Response: 43070
Conc: 0.24 ng/ml



#42 AR-1268-2


R.T.: 7.447 min
Delta R.T.: 0.019 min
Response: 616955
Conc: 3.01 ng/ml



#43 AR-1268-3


R.T.: 8.841 min
Delta R.T.: -0.003 min
Response: 43572
Conc: 0.28 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.645 min
Delta R.T.: -0.004 min
Response: 5775
Conc: 0.03 ng/ml




#44 AR-1268-4

R.T.: 9.290 min
Delta R.T.: 0.003 min
Response: 44062
Conc: 0.70 ng/ml



#44 AR-1268-4


R.T.: 7.970 min
Delta R.T.: 0.000 min
Response: 86367
Conc: 1.06 ng/ml



#45 AR-1268-5

R.T.: 9.706 min
Delta R.T.: -0.005 min
Response: 61212
Conc: 0.13 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.276 min
Delta R.T.: 0.002 min
Response: 23406
Conc: 0.04 ng/ml