

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031122\
 Data File : PP044606.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Mar 2022 03:16
 Operator : YP\AJ
 Sample : N1886-04
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 30970

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/14/2022
 Supervised By :Ankita Jodhani 03/14/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 12 12:19:21 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 03 02:21:53 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.908	3.166	28359164	27701673	13.687m	15.835
2) SA Decachlor...	10.041	8.526	19360259	20899008	14.066	13.624

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031122\
Data File : PP044606.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Mar 2022 03:16
Operator : YP\AJ
Sample : N1886-04
Misc :
ALS Vial : 52 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
30970


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/14/2022
Supervised By :Ankita Jodhani 03/14/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 12 12:19:21 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 03 02:21:53 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.908 min
Delta R.T.: -0.004 min
Response: 28359164
Conc: 13.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
30970


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/14/2022
Supervised By :Ankita Jodhani 03/14/2022




#1 Tetrachloro-m-xylene

R.T.: 3.166 min
Delta R.T.: 0.000 min
Response: 27701673
Conc: 15.83 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.041 min
Delta R.T.: -0.010 min
Response: 19360259
Conc: 14.07 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.526 min
Delta R.T.: -0.011 min
Response: 20899008
Conc: 13.62 ng/ml