

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031122\
 Data File : PP044611.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Mar 2022 05:30
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/14/2022
 Supervised By :Ankita Jodhani 03/14/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 12 06:13:14 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 03 02:21:53 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.908	3.164	106.6E6	90889117	51.450	51.953
2) SA Decachlor...	10.042	8.526	64469524	72754573	46.841	47.429m
Target Compounds						
21) L5 AR-1248-1	5.159	4.306	20229640	17533288	489.357	494.328
22) L5 AR-1248-2	5.456	4.558	25984342	22074858	463.196	469.065
23) L5 AR-1248-3	5.674	4.601	30700891	22709401	449.511	460.794
24) L5 AR-1248-4	6.114	4.782	32344283	26295087	443.532	465.242
25) L5 AR-1248-5	6.155	5.202	31507842	25510560	428.981	454.807

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031122\
Data File : PP044611.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Mar 2022 05:30
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1248CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/14/2022
Supervised By :Ankita Jodhani 03/14/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 12 06:13:14 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 03 02:21:53 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.908 min
Delta R.T.: -0.004 min
Response: 106603178
Conc: 51.45 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/14/2022
Supervised By :Ankita Jodhani 03/14/2022




#1 Tetrachloro-m-xylene

R.T.: 3.164 min
Delta R.T.: -0.003 min
Response: 90889117
Conc: 51.95 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.042 min
Delta R.T.: -0.010 min
Response: 64469524
Conc: 46.84 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.526 min
Delta R.T.: -0.011 min
Response: 72754573
Conc: 47.43 ng/ml m




#21 AR-1248-1

R.T.: 5.159 min
Delta R.T.: -0.006 min
Response: 20229640
Conc: 489.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/14/2022
Supervised By :Ankita Jodhani 03/14/2022




#21 AR-1248-1

R.T.: 4.306 min
Delta R.T.: -0.007 min
Response: 17533288
Conc: 494.33 ng/ml




#22 AR-1248-2

R.T.: 5.456 min
Delta R.T.: -0.006 min
Response: 25984342
Conc: 463.20 ng/ml



#22 AR-1248-2

R.T.: 4.558 min
Delta R.T.: -0.007 min
Response: 22074858
Conc: 469.06 ng/ml




#23 AR-1248-3

R.T.: 5.674 min
Delta R.T.: -0.006 min
Response: 30700891
Conc: 449.51 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/14/2022
Supervised By :Ankita Jodhani 03/14/2022




#23 AR-1248-3

R.T.: 4.601 min
Delta R.T.: -0.007 min
Response: 22709401
Conc: 460.79 ng/ml




#24 AR-1248-4

R.T.: 6.114 min
Delta R.T.: -0.006 min
Response: 32344283
Conc: 443.53 ng/ml



#24 AR-1248-4

R.T.: 4.782 min
Delta R.T.: -0.007 min
Response: 26295087
Conc: 465.24 ng/ml




#25 AR-1248-5

R.T.: 6.155 min
Delta R.T.: -0.006 min
Response: 31507842
Conc: 428.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/14/2022
Supervised By :Ankita Jodhani 03/14/2022



#25 AR-1248-5

R.T.: 5.202 min
Delta R.T.: -0.008 min
Response: 25510560
Conc: 454.81 ng/ml