

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031122\
 Data File : PP044617.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Mar 2022 07:02
 Operator : YP\AJ
 Sample : N1892-02 10X
 Misc :
 ALS Vial : 58 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 MOO-2022-19A

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/14/2022
 Supervised By :Ankita Jodhani 03/14/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 13 23:19:55 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 03 02:21:53 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.907	3.163	2686846	2856458	1.297m	1.633 #
2) SA Decachlor...	10.043	8.526	2768162	3083429	2.011m	2.010

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031122\
Data File : PP044617.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Mar 2022 07:02
Operator : YP\AJ
Sample : N1892-02 10X
Misc :
ALS Vial : 58 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
MOO-2022-19A


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/14/2022
Supervised By :Ankita Jodhani 03/14/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 13 23:19:55 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 03 02:21:53 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.907 min
Delta R.T.: -0.005 min
Response: 2686846
Conc: 1.30 ng/ml

Instrument :
ECD_P
ClientSampleId :
MOO-2022-19A


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/14/2022
Supervised By :Ankita Jodhani 03/14/2022




#1 Tetrachloro-m-xylene

R.T.: 3.163 min
Delta R.T.: -0.004 min
Response: 2856458
Conc: 1.63 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.043 min
Delta R.T.: -0.009 min
Response: 2768162
Conc: 2.01 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.526 min
Delta R.T.: -0.011 min
Response: 3083429
Conc: 2.01 ng/ml