

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031123\
 Data File : PP056407.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Mar 2023 20:46
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/13/2023
 Supervised By :Ankita Jodhani 03/13/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 11 00:01:31 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 08 02:03:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.320	3.561	87547551	70439233	56.383	49.919
2)	SA Decachlor...	10.063	8.557	98249143	53862210	49.462	48.183
Target Compounds							
3)	L1 AR-1016-1	5.489	4.640	26894669	24367531	464.560m	578.388
4)	L1 AR-1016-2	5.512	4.659	37505795	32405638	437.805m	534.667
5)	L1 AR-1016-3	5.575	4.835	21979132	15148216	402.972m	468.016
6)	L1 AR-1016-4	5.673	4.877	26801898	12300013	617.246m	438.263 #
7)	L1 AR-1016-5	5.970	5.090	23286851	21975053	507.905m	614.219
31)	L7 AR-1260-1	7.101	6.124	59567622	36587626	604.451m	536.704m
32)	L7 AR-1260-2	7.358	6.313	57733613	35277057	497.726m	461.151m
33)	L7 AR-1260-3	7.720	6.466	42244623	35721246	493.503	494.002m
34)	L7 AR-1260-4	7.945	6.939	49245014	40522729	498.102m	675.747 #
35)	L7 AR-1260-5	8.263	7.183	114.8E6	63219895	574.516m	494.755m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031123\
Data File : PP056407.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Mar 2023 20:46
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/13/2023
Supervised By :Ankita Jodhani 03/13/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 11 00:01:31 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030823.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 08 02:03:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

