

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031125\
 Data File : PP070431.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Mar 2025 18:28
 Operator : YP\AJ
 Sample : AR1242ICC750
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 11 22:16:47 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 11 22:15:44 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.525	3.828	115.4E6	72710939	72.677	72.698
2) SA Decachlor...	10.252	8.880	79754897	74745771	72.592	73.377
Target Compounds						
16) L4 AR-1242-1	5.678	4.916	31278275	21743405	719.336	720.598
17) L4 AR-1242-2	5.700	4.935	45085458	30497830	724.511	727.171
18) L4 AR-1242-3	5.762	5.112	27285376	16877928	705.908	709.226
19) L4 AR-1242-4	5.859	5.196	23125338	15945991	710.396	712.863
20) L4 AR-1242-5	6.590	5.721	23553432	20212182	735.787	718.220



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031125\
Data File : PP070431.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Mar 2025 18:28
Operator : YP\AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 11 22:16:47 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 11 22:15:44 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.525 min
Delta R.T.: 0.000 min
Response: 115389928
Conc: 72.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.828 min
Delta R.T.: 0.000 min
Response: 72710939
Conc: 72.70 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.252 min
Delta R.T.: -0.002 min
Response: 79754897
Conc: 72.59 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.880 min
Delta R.T.: -0.001 min
Response: 74745771
Conc: 73.38 ng/ml



#16 AR-1242-1


R.T.: 5.678 min
Delta R.T.: 0.000 min
Response: 31278275
Conc: 719.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750




#16 AR-1242-1

R.T.: 4.916 min
Delta R.T.: 0.000 min
Response: 21743405
Conc: 720.60 ng/ml




#17 AR-1242-2

R.T.: 5.700 min
Delta R.T.: 0.000 min
Response: 45085458
Conc: 724.51 ng/ml



#17 AR-1242-2


R.T.: 4.935 min
Delta R.T.: 0.000 min
Response: 30497830
Conc: 727.17 ng/ml



#18 AR-1242-3


R.T.: 5.762 min
Delta R.T.: 0.000 min
Response: 27285376
Conc: 705.91 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750




#18 AR-1242-3

R.T.: 5.112 min
Delta R.T.: 0.000 min
Response: 16877928
Conc: 709.23 ng/ml




#19 AR-1242-4

R.T.: 5.859 min
Delta R.T.: 0.000 min
Response: 23125338
Conc: 710.40 ng/ml



#19 AR-1242-4


R.T.: 5.196 min
Delta R.T.: 0.000 min
Response: 15945991
Conc: 712.86 ng/ml



#20 AR-1242-5

R.T.: 6.590 min
Delta R.T.: 0.000 min
Response: 23553432
Conc: 735.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750



#20 AR-1242-5

R.T.: 5.721 min
Delta R.T.: 0.000 min
Response: 20212182
Conc: 718.22 ng/ml