

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031221\
 Data File : PP033951.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 13 Mar 2021 00:07
 Operator : DD\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 13 03:02:52 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030921.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Mar 12 12:21:17 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.856	4.266	5064130	2472208	58.528	60.853
2)	SA Decachlor...	10.786	9.591	4832331	1550121	57.218	58.089
Target Compounds							
3)	L1 AR-1016-1	6.169	5.523	1584958	609285	594.923	535.329
4)	L1 AR-1016-2	6.191	5.544	2336457	901520	578.068	545.371
5)	L1 AR-1016-3	6.257	5.737	1561374	489849	630.965	534.589
6)	L1 AR-1016-4	6.363	5.786	1263063	361323	619.616	500.371
7)	L1 AR-1016-5	6.677	6.016	1124129	447727	554.572	484.241
8)	L2 AR-1221-1	5.098	4.520	149811	93222	162.345	218.925 #
9)	L2 AR-1221-2	5.203	4.622	241741	123401	344.028	381.363
10)	L2 AR-1221-3	5.288	4.711	894058	392754	424.767	400.682
11)	L3 AR-1232-1	5.288	4.711	894058	392754	516.766	483.444
12)	L3 AR-1232-2	5.874	5.544	1405121	901520	1534.245	1244.007
13)	L3 AR-1232-3	6.191	5.737	2336457	489849	1382.308	1273.795
14)	L3 AR-1232-4	6.363	5.831	1263063	463453	1478.566	1389.545
15)	L3 AR-1232-5	6.462	6.016	912339	447727	1537.908	1235.450
16)	L4 AR-1242-1	6.169	5.523	1584958	609285	781.131	690.009
17)	L4 AR-1242-2	6.191	5.544	2336457	901520	769.501	718.784
18)	L4 AR-1242-3	6.257	5.737	1561374	489849	817.204	693.708
19)	L4 AR-1242-4	6.363	5.831	1263063	463453	812.079	676.892
20)	L4 AR-1242-5	7.144	6.394	466288	345838	290.478	432.628 #
21)	L5 AR-1248-1	6.169	5.523	1584958	609285	1016.024	895.079
22)	L5 AR-1248-2	6.462	5.786	912339	361323	428.598	385.094
23)	L5 AR-1248-3	6.677	5.831	1124129	463453	436.116	460.259
24)	L5 AR-1248-4	7.104	6.016	305047	447727	110.261	382.409 #
25)	L5 AR-1248-5	7.144	6.437	466288	66150	170.968	61.415 #
26)	L6 AR-1254-1	7.074	6.394	810709	345838	277.517	212.129
27)	L6 AR-1254-2	7.302	6.551	1057522	364156	232.570	254.974
28)	L6 AR-1254-3	7.683	6.989f	746296	717918	150.943	313.664 #
29)	L6 AR-1254-4	7.977	7.208	590924	111142	172.279	87.443 #
30)	L6 AR-1254-5	8.405	7.635	2759102	1114630	700.963	591.780
31)	L7 AR-1260-1	7.850	7.108	2370328	941617	690.236	616.132
32)	L7 AR-1260-2	8.116	7.301	2033740	1104283	493.957	610.521
33)	L7 AR-1260-3	8.480	7.458	1784941	948706	551.457	560.176
34)	L7 AR-1260-4	8.710	7.939	2115459	801221	571.102	544.789
35)	L7 AR-1260-5	9.029	8.181	4422893	1870502	571.713	575.309
36)	L8 AR-1262-1	8.480	7.635	1784941	1114630	375.135	1048.560 #
37)	L8 AR-1262-2	9.029	8.181	4422893	1870502	511.707	561.344
38)	L8 AR-1262-3	9.352f	8.468	2860470	435525	471.336	304.967 #
39)	L8 AR-1262-4	9.406f	8.531	2042946	1387506	412.157	550.339 #
40)	L8 AR-1262-5	10.067	9.030	1432173	511838	435.134	449.029
41)	L9 AR-1268-1	9.352f	8.468	2860470	435525	274.719	113.867 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031221\
 Data File : PP033951.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 13 Mar 2021 00:07
 Operator : DD\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 13 03:02:52 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030921.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Mar 12 12:21:17 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.406f	8.531	2042946	1387506	202.136	388.338 #
43) L9	AR-1268-3	9.643	8.741	105280	25274	11.855	8.136 #
44) L9	AR-1268-4	10.067	9.030	1432173	511838	411.245	421.718
45) L9	AR-1268-5	10.464	9.329	413353	138513	13.980	15.150



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031221\
Data File : PP033951.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 Mar 2021 00:07
Operator : DD\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 13 03:02:52 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030921.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 12 12:21:17 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.856 min
Delta R.T.: -0.002 min
Response: 5064130
Conc: 58.53 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.266 min
Delta R.T.: -0.005 min
Response: 2472208
Conc: 60.85 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.786 min
Delta R.T.: -0.002 min
Response: 4832331
Conc: 57.22 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.591 min
Delta R.T.: -0.013 min
Response: 1550121
Conc: 58.09 ng/ml



#3 AR-1016-1


R.T.: 6.169 min
Delta R.T.: -0.002 min
Response: 1584958
Conc: 594.92 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 5.523 min
Delta R.T.: -0.008 min
Response: 609285
Conc: 535.33 ng/ml




#4 AR-1016-2

R.T.: 6.191 min
Delta R.T.: -0.003 min
Response: 2336457
Conc: 578.07 ng/ml



#4 AR-1016-2


R.T.: 5.544 min
Delta R.T.: -0.008 min
Response: 901520
Conc: 545.37 ng/ml



#5 AR-1016-3


R.T.: 6.257 min
Delta R.T.: -0.002 min
Response: 1561374
Conc: 630.96 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 5.737 min
Delta R.T.: -0.007 min
Response: 489849
Conc: 534.59 ng/ml




#6 AR-1016-4

R.T.: 6.363 min
Delta R.T.: -0.002 min
Response: 1263063
Conc: 619.62 ng/ml



#6 AR-1016-4


R.T.: 5.786 min
Delta R.T.: -0.007 min
Response: 361323
Conc: 500.37 ng/ml



#7 AR-1016-5


R.T.: 6.677 min
Delta R.T.: -0.002 min
Response: 1124129
Conc: 554.57 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 6.016 min
Delta R.T.: -0.007 min
Response: 447727
Conc: 484.24 ng/ml




#8 AR-1221-1

R.T.: 5.098 min
Delta R.T.: -0.011 min
Response: 149811
Conc: 162.34 ng/ml



#8 AR-1221-1


R.T.: 4.520 min
Delta R.T.: -0.009 min
Response: 93222
Conc: 218.93 ng/ml



#9 AR-1221-2


R.T.: 5.203 min
Delta R.T.: -0.004 min
Response: 241741
Conc: 344.03 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 4.622 min
Delta R.T.: -0.007 min
Response: 123401
Conc: 381.36 ng/ml




#10 AR-1221-3

R.T.: 5.288 min
Delta R.T.: -0.005 min
Response: 894058
Conc: 424.77 ng/ml



#10 AR-1221-3


R.T.: 4.711 min
Delta R.T.: -0.007 min
Response: 392754
Conc: 400.68 ng/ml



#11 AR-1232-1


R.T.: 5.288 min
Delta R.T.: -0.004 min
Response: 894058
Conc: 516.77 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.711 min
Delta R.T.: -0.007 min
Response: 392754
Conc: 483.44 ng/ml




#12 AR-1232-2

R.T.: 5.874 min
Delta R.T.: -0.003 min
Response: 1405121
Conc: 1534.24 ng/ml



#12 AR-1232-2


R.T.: 5.544 min
Delta R.T.: -0.008 min
Response: 901520
Conc: 1244.01 ng/ml



#13 AR-1232-3


R.T.: 6.191 min
Delta R.T.: -0.004 min
Response: 2336457
Conc: 1382.31 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 5.737 min
Delta R.T.: -0.007 min
Response: 489849
Conc: 1273.79 ng/ml




#14 AR-1232-4

R.T.: 6.363 min
Delta R.T.: -0.003 min
Response: 1263063
Conc: 1478.57 ng/ml



#14 AR-1232-4


R.T.: 5.831 min
Delta R.T.: -0.008 min
Response: 463453
Conc: 1389.54 ng/ml



#15 AR-1232-5


R.T.: 6.462 min
Delta R.T.: -0.003 min
Response: 912339
Conc: 1537.91 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 6.016 min
Delta R.T.: -0.007 min
Response: 447727
Conc: 1235.45 ng/ml




#16 AR-1242-1

R.T.: 6.169 min
Delta R.T.: -0.002 min
Response: 1584958
Conc: 781.13 ng/ml



#16 AR-1242-1


R.T.: 5.523 min
Delta R.T.: -0.008 min
Response: 609285
Conc: 690.01 ng/ml



#17 AR-1242-2


R.T.: 6.191 min
Delta R.T.: -0.004 min
Response: 2336457
Conc: 769.50 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 5.544 min
Delta R.T.: -0.008 min
Response: 901520
Conc: 718.78 ng/ml




#18 AR-1242-3

R.T.: 6.257 min
Delta R.T.: -0.003 min
Response: 1561374
Conc: 817.20 ng/ml



#18 AR-1242-3


R.T.: 5.737 min
Delta R.T.: -0.007 min
Response: 489849
Conc: 693.71 ng/ml



#19 AR-1242-4


R.T.: 6.363 min
Delta R.T.: -0.003 min
Response: 1263063
Conc: 812.08 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.831 min
Delta R.T.: -0.007 min
Response: 463453
Conc: 676.89 ng/ml




#20 AR-1242-5

R.T.: 7.144 min
Delta R.T.: -0.002 min
Response: 466288
Conc: 290.48 ng/ml



#20 AR-1242-5


R.T.: 6.394 min
Delta R.T.: -0.007 min
Response: 345838
Conc: 432.63 ng/ml



#21 AR-1248-1


R.T.: 6.169 min
Delta R.T.: -0.002 min
Response: 1584958
Conc: 1016.02 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 5.523 min
Delta R.T.: -0.008 min
Response: 609285
Conc: 895.08 ng/ml




#22 AR-1248-2

R.T.: 6.462 min
Delta R.T.: -0.003 min
Response: 912339
Conc: 428.60 ng/ml



#22 AR-1248-2


R.T.: 5.786 min
Delta R.T.: -0.008 min
Response: 361323
Conc: 385.09 ng/ml



#23 AR-1248-3


R.T.: 6.677 min
Delta R.T.: -0.003 min
Response: 1124129
Conc: 436.12 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.831 min
Delta R.T.: -0.008 min
Response: 463453
Conc: 460.26 ng/ml




#24 AR-1248-4

R.T.: 7.104 min
Delta R.T.: -0.003 min
Response: 305047
Conc: 110.26 ng/ml



#24 AR-1248-4


R.T.: 6.016 min
Delta R.T.: -0.008 min
Response: 447727
Conc: 382.41 ng/ml



#25 AR-1248-5


R.T.: 7.144 min
Delta R.T.: -0.002 min
Response: 466288
Conc: 170.97 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 6.437 min
Delta R.T.: -0.009 min
Response: 66150
Conc: 61.41 ng/ml




#26 AR-1254-1

R.T.: 7.074 min
Delta R.T.: 0.000 min
Response: 810709
Conc: 277.52 ng/ml



#26 AR-1254-1


R.T.: 6.394 min
Delta R.T.: 0.000 min
Response: 345838
Conc: 212.13 ng/ml



#27 AR-1254-2


R.T.: 7.302 min
Delta R.T.: -0.002 min
Response: 1057522
Conc: 232.57 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 6.551 min
Delta R.T.: -0.001 min
Response: 364156
Conc: 254.97 ng/ml




#28 AR-1254-3

R.T.: 7.683 min
Delta R.T.: 0.000 min
Response: 746296
Conc: 150.94 ng/ml



#28 AR-1254-3


R.T.: 6.989 min
Delta R.T.: 0.016 min
Response: 717918
Conc: 313.66 ng/ml



#29 AR-1254-4


R.T.: 7.977 min
Delta R.T.: -0.002 min
Response: 590924
Conc: 172.28 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 7.208 min
Delta R.T.: 0.000 min
Response: 111142
Conc: 87.44 ng/ml




#30 AR-1254-5

R.T.: 8.405 min
Delta R.T.: -0.001 min
Response: 2759102
Conc: 700.96 ng/ml



#30 AR-1254-5


R.T.: 7.635 min
Delta R.T.: -0.001 min
Response: 1114630
Conc: 591.78 ng/ml



#31 AR-1260-1


R.T.: 7.850 min
Delta R.T.: -0.002 min
Response: 2370328
Conc: 690.24 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 7.108 min
Delta R.T.: -0.008 min
Response: 941617
Conc: 616.13 ng/ml




#32 AR-1260-2

R.T.: 8.116 min
Delta R.T.: -0.002 min
Response: 2033740
Conc: 493.96 ng/ml



#32 AR-1260-2


R.T.: 7.301 min
Delta R.T.: -0.009 min
Response: 1104283
Conc: 610.52 ng/ml



#33 AR-1260-3


R.T.: 8.480 min
Delta R.T.: -0.001 min
Response: 1784941
Conc: 551.46 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 7.458 min
Delta R.T.: -0.009 min
Response: 948706
Conc: 560.18 ng/ml




#34 AR-1260-4

R.T.: 8.710 min
Delta R.T.: -0.002 min
Response: 2115459
Conc: 571.10 ng/ml



#34 AR-1260-4


R.T.: 7.939 min
Delta R.T.: -0.009 min
Response: 801221
Conc: 544.79 ng/ml



#35 AR-1260-5


R.T.: 9.029 min
Delta R.T.: -0.002 min
Response: 4422893
Conc: 571.71 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 8.181 min
Delta R.T.: -0.010 min
Response: 1870502
Conc: 575.31 ng/ml




#36 AR-1262-1

R.T.: 8.480 min
Delta R.T.: -0.002 min
Response: 1784941
Conc: 375.14 ng/ml



#36 AR-1262-1


R.T.: 7.635 min
Delta R.T.: -0.008 min
Response: 1114630
Conc: 1048.56 ng/ml



#37 AR-1262-2


R.T.: 9.029 min
Delta R.T.: -0.002 min
Response: 4422893
Conc: 511.71 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 8.181 min
Delta R.T.: -0.009 min
Response: 1870502
Conc: 561.34 ng/ml




#38 AR-1262-3

R.T.: 9.352 min
Delta R.T.: 0.019 min
Response: 2860470
Conc: 471.34 ng/ml



#38 AR-1262-3


R.T.: 8.468 min
Delta R.T.: -0.009 min
Response: 435525
Conc: 304.97 ng/ml



#39 AR-1262-4


R.T.: 9.406 min
Delta R.T.: -0.018 min
Response: 2042946
Conc: 412.16 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 8.531 min
Delta R.T.: -0.010 min
Response: 1387506
Conc: 550.34 ng/ml




#40 AR-1262-5

R.T.: 10.067 min
Delta R.T.: 0.000 min
Response: 1432173
Conc: 435.13 ng/ml



#40 AR-1262-5


R.T.: 9.030 min
Delta R.T.: -0.009 min
Response: 511838
Conc: 449.03 ng/ml



#41 AR-1268-1


R.T.: 9.352 min
Delta R.T.: 0.020 min
Response: 2860470
Conc: 274.72 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 8.468 min
Delta R.T.: -0.009 min
Response: 435525
Conc: 113.87 ng/ml




#42 AR-1268-2

R.T.: 9.406 min
Delta R.T.: -0.020 min
Response: 2042946
Conc: 202.14 ng/ml



#42 AR-1268-2


R.T.: 8.531 min
Delta R.T.: -0.010 min
Response: 1387506
Conc: 388.34 ng/ml



#43 AR-1268-3


R.T.: 9.643 min
Delta R.T.: 0.000 min
Response: 105280
Conc: 11.85 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 8.741 min
Delta R.T.: -0.010 min
Response: 25274
Conc: 8.14 ng/ml




#44 AR-1268-4

R.T.: 10.067 min
Delta R.T.: -0.001 min
Response: 1432173
Conc: 411.24 ng/ml



#44 AR-1268-4


R.T.: 9.030 min
Delta R.T.: -0.008 min
Response: 511838
Conc: 421.72 ng/ml



#45 AR-1268-5

R.T.: 10.464 min
Delta R.T.: -0.002 min
Response: 413353
Conc: 13.98 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 9.329 min
Delta R.T.: -0.009 min
Response: 138513
Conc: 15.15 ng/ml