

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031221\
 Data File : PP033979.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 13 Mar 2021 9:24
 Operator : DD\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 15 07:41:15 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030921.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Mar 12 12:21:17 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.853	4.265	2154816	1087937	24.904	26.779
2) SA Decachlor...	10.781	9.587f	2133864	698155	25.267	26.163
Target Compounds						
9) L2 AR-1221-2	5.202	4.622	57395	27575	81.680	85.220
24) L5 AR-1248-4	7.082f	6.052f	1342231	-38913	485.157	N.D. #
26) L6 AR-1254-1	7.082	0.000	1342231	0	459.465	N.D. #
28) L6 AR-1254-3	7.695	6.947f	856162	3818	173.164	1.668 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031221\
Data File : PP033979.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 Mar 2021 9:24
Operator : DD\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 15 07:41:15 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030921.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 12 12:21:17 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.853 min
Delta R.T.: -0.005 min
Response: 2154816
Conc: 24.90 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.265 min
Delta R.T.: -0.006 min
Response: 1087937
Conc: 26.78 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.781 min
Delta R.T.: -0.006 min
Response: 2133864
Conc: 25.27 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.587 min
Delta R.T.: -0.015 min
Response: 698155
Conc: 26.16 ng/ml



#9 AR-1221-2


R.T.: 5.202 min
Delta R.T.: -0.005 min
Response: 57395
Conc: 81.68 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 4.622 min
Delta R.T.: -0.006 min
Response: 27575
Conc: 85.22 ng/ml




#24 AR-1248-4

R.T.: 7.082 min
Delta R.T.: -0.025 min
Response: 1342231
Conc: 485.16 ng/ml



#24 AR-1248-4


R.T.: 6.052 min
Delta R.T.: 0.028 min
Response: -38913
Conc: N.D.



#26 AR-1254-1


R.T.: 7.082 min
Delta R.T.: 0.008 min
Response: 1342231
Conc: 459.46 ng/ml

Instrument :
ECD_P
ClientSampleId :




#26 AR-1254-1

R.T.: 0.000 min
Exp R.T. : 6.395 min
Response: 0
Conc: N.D.



#28 AR-1254-3

R.T.: 7.695 min
Delta R.T.: 0.012 min
Response: 856162
Conc: 173.16 ng/ml



#28 AR-1254-3

R.T.: 6.947 min
Delta R.T.: -0.026 min
Response: 3818
Conc: 1.67 ng/ml