

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031323\
 Data File : PP056485.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Mar 2023 05:57
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 14 06:33:36 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 08 02:03:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.317	3.557	90341913	75970006	58.183	53.838
2) SA Decachlor...	10.065	8.557	94330851	56917623	47.489	50.917
Target Compounds						
16) L4 AR-1242-1	5.487	4.638	25563560	18288889	524.551	517.175
17) L4 AR-1242-2	5.511	4.656	37245755	26107149	522.024	513.835
18) L4 AR-1242-3	5.572	4.832	23704715	14141826	520.705	524.069
19) L4 AR-1242-4	5.671	4.916	19308679	15645511	540.218	533.987
20) L4 AR-1242-5	6.411	5.440	20981807	17097829	490.800	529.940



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031323\
Data File : PP056485.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Mar 2023 05:57
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 14 06:33:36 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030823.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 08 02:03:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.317 min
Delta R.T.: -0.001 min
Response: 90341913
Conc: 58.18 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.557 min
Delta R.T.: -0.003 min
Response: 75970006
Conc: 53.84 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.065 min
Delta R.T.: 0.003 min
Response: 94330851
Conc: 47.49 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.557 min
Delta R.T.: -0.004 min
Response: 56917623
Conc: 50.92 ng/ml



#16 AR-1242-1


R.T.: 5.487 min
Delta R.T.: -0.001 min
Response: 25563560
Conc: 524.55 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.638 min
Delta R.T.: -0.003 min
Response: 18288889
Conc: 517.18 ng/ml




#17 AR-1242-2

R.T.: 5.511 min
Delta R.T.: 0.000 min
Response: 37245755
Conc: 522.02 ng/ml



#17 AR-1242-2


R.T.: 4.656 min
Delta R.T.: -0.002 min
Response: 26107149
Conc: 513.83 ng/ml



#18 AR-1242-3


R.T.: 5.572 min
Delta R.T.: 0.000 min
Response: 23704715
Conc: 520.71 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.832 min
Delta R.T.: -0.002 min
Response: 14141826
Conc: 524.07 ng/ml




#19 AR-1242-4

R.T.: 5.671 min
Delta R.T.: 0.000 min
Response: 19308679
Conc: 540.22 ng/ml



#19 AR-1242-4


R.T.: 4.916 min
Delta R.T.: -0.002 min
Response: 15645511
Conc: 533.99 ng/ml



#20 AR-1242-5

R.T.: 6.411 min
Delta R.T.: 0.000 min
Response: 20981807
Conc: 490.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.440 min
Delta R.T.: -0.002 min
Response: 17097829
Conc: 529.94 ng/ml