

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031325\
 Data File : PP070496.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Mar 2025 16:42
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 13 17:09:01 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 12 02:42:43 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.526	3.828	76553526	48625390	49.218	48.515
2) SA Decachlor...	10.248	8.875	54048181	47925847	50.876	48.986
Target Compounds						
21) L5 AR-1248-1	5.677	4.915	16022534	11346929	469.099	483.351
22) L5 AR-1248-2	5.949	5.152	21303081	14555691	501.872	464.987
23) L5 AR-1248-3	6.152	5.195	23339701	15249984	505.751	480.745
24) L5 AR-1248-4	6.551	5.367	28146223	18989569	485.353	487.247
25) L5 AR-1248-5	6.590	5.761	27022408	18432629	503.329	483.965



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031325\
Data File : PP070496.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Mar 2025 16:42
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 13 17:09:01 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031125.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 12 02:42:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.526 min
Delta R.T.: 0.002 min
Response: 76553526
Conc: 49.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.828 min
Delta R.T.: 0.000 min
Response: 48625390
Conc: 48.51 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.248 min
Delta R.T.: -0.004 min
Response: 54048181
Conc: 50.88 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.875 min
Delta R.T.: -0.006 min
Response: 47925847
Conc: 48.99 ng/ml



#21 AR-1248-1


R.T.: 5.677 min
Delta R.T.: 0.002 min
Response: 16022534
Conc: 469.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.915 min
Delta R.T.: 0.000 min
Response: 11346929
Conc: 483.35 ng/ml




#22 AR-1248-2

R.T.: 5.949 min
Delta R.T.: 0.000 min
Response: 21303081
Conc: 501.87 ng/ml



#22 AR-1248-2


R.T.: 5.152 min
Delta R.T.: 0.000 min
Response: 14555691
Conc: 464.99 ng/ml



#23 AR-1248-3


R.T.: 6.152 min
Delta R.T.: 0.001 min
Response: 23339701
Conc: 505.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.195 min
Delta R.T.: 0.000 min
Response: 15249984
Conc: 480.74 ng/ml




#24 AR-1248-4

R.T.: 6.551 min
Delta R.T.: 0.001 min
Response: 28146223
Conc: 485.35 ng/ml



#24 AR-1248-4


R.T.: 5.367 min
Delta R.T.: -0.001 min
Response: 18989569
Conc: 487.25 ng/ml



#25 AR-1248-5

R.T.: 6.590 min
Delta R.T.: 0.000 min
Response: 27022408
Conc: 503.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.761 min
Delta R.T.: -0.001 min
Response: 18432629
Conc: 483.97 ng/ml