

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031325\
 Data File : PP070497.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Mar 2025 16:58
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 13 17:09:48 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 12 02:42:43 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.521	3.827	79320178	51340762	50.997	51.224
2) SA Decachlor...	10.243	8.875	55834026	51610988	52.557	52.752
Target Compounds						
26) L6 AR-1254-1	6.522	5.719	28333115	27580341	494.817	489.683
27) L6 AR-1254-2	6.738	5.867	42291085	23674980	521.776	484.087
28) L6 AR-1254-3	7.101	6.271	42006921	36993526	499.062	490.513
29) L6 AR-1254-4	7.383	6.500	35386263	25922303	504.529	491.400
30) L6 AR-1254-5	7.800	6.918	34432034	33250921	492.891	481.516



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031325\
Data File : PP070497.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Mar 2025 16:58
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 13 17:09:48 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031125.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 12 02:42:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.521 min
Delta R.T.: -0.003 min
Response: 79320178
Conc: 51.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.827 min
Delta R.T.: -0.002 min
Response: 51340762
Conc: 51.22 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.243 min
Delta R.T.: -0.009 min
Response: 55834026
Conc: 52.56 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.875 min
Delta R.T.: -0.007 min
Response: 51610988
Conc: 52.75 ng/ml



#26 AR-1254-1


R.T.: 6.522 min
Delta R.T.: -0.008 min
Response: 28333115
Conc: 494.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.719 min
Delta R.T.: -0.001 min
Response: 27580341
Conc: 489.68 ng/ml




#27 AR-1254-2

R.T.: 6.738 min
Delta R.T.: -0.007 min
Response: 42291085
Conc: 521.78 ng/ml



#27 AR-1254-2


R.T.: 5.867 min
Delta R.T.: -0.002 min
Response: 23674980
Conc: 484.09 ng/ml



#28 AR-1254-3


R.T.: 7.101 min
Delta R.T.: -0.007 min
Response: 42006921
Conc: 499.06 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.271 min
Delta R.T.: -0.003 min
Response: 36993526
Conc: 490.51 ng/ml




#29 AR-1254-4

R.T.: 7.383 min
Delta R.T.: -0.008 min
Response: 35386263
Conc: 504.53 ng/ml



#29 AR-1254-4


R.T.: 6.500 min
Delta R.T.: -0.002 min
Response: 25922303
Conc: 491.40 ng/ml



#30 AR-1254-5

R.T.: 7.800 min
Delta R.T.: -0.008 min
Response: 34432034
Conc: 492.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.918 min
Delta R.T.: -0.003 min
Response: 33250921
Conc: 481.52 ng/ml