

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031419\
 Data File : PP029301.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 14 Mar 2019 15:32
 Operator : EI\MA
 Sample : K1824-02
 Misc : hexane
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 TP-24

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 15 02:32:04 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022119.M
 Quant Title : 8080.M
 QLast Update : Fri Feb 22 02:24:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.930	12.125	19963863	20969966	664.214	552.532
Target Compounds							
3) T	4-Nitroph...	0.000	11.334	0	9278839	N.D.	387.461 #
8) T	DICHLORPROP	0.000	13.551	0	860509	N.D.	25.452 #
9) T	2,4-D	13.814	14.035	468977	2459310	11.961	54.237 #
10) T	Pentachlo...	0.000	14.644	0	3899269	N.D.	7.576 #
11) T	2,4,5-TP ...	15.139	15.199	789705	3624966	3.329	17.821 #
12) T	2,4,5-T	15.443	15.852	8059103	2682991	38.174	16.204 #
13) T	2,4-DB	16.243	16.438	22963584	6598350	1202.135	292.914 #
14) T	DINOSEB	17.696	16.816	2016948	3006293	13.850	22.936 #
15) T	Picloram	17.484	18.228	7336875	8710015	26.387	38.707 #
16) T	DCPA	0.000	18.150	0	11218369	N.D.	48.273 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031419\
Data File : PP029301.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 14 Mar 2019 15:32
Operator : EI\MA
Sample : K1824-02
Misc : hexane
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :
TP-24

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 15 02:32:04 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022119.M
Quant Title : 8080.M
QLast Update : Fri Feb 22 02:24:04 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#3 4-Nitrophenol


R.T.: 0.000 min
Exp R.T.: 11.504 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
TP-24




#3 4-Nitrophenol

R.T.: 11.334 min
Delta R.T.: 0.061 min
Response: 9278839
Conc: 387.46 ng/ml




#4 2,4-DCAA

R.T.: 11.930 min
Delta R.T.: -0.011 min
Response: 19963863
Conc: 664.21 ng/ml



#4 2,4-DCAA


R.T.: 12.125 min
Delta R.T.: -0.004 min
Response: 20969966
Conc: 552.53 ng/ml



#8 DICHLORPROP


R.T.: 0.000 min
Exp R.T.: 13.452 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
TP-24




#8 DICHLORPROP

R.T.: 13.551 min
Delta R.T.: 0.000 min
Response: 860509
Conc: 25.45 ng/ml




#9 2,4-D

R.T.: 13.814 min
Delta R.T.: 0.014 min
Response: 468977
Conc: 11.96 ng/ml



#9 2,4-D


R.T.: 14.035 min
Delta R.T.: -0.003 min
Response: 2459310
Conc: 54.24 ng/ml



#10 Pentachlorophenol


R.T.: 0.000 min
Exp R.T.: 14.216 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
TP-24




#10 Pentachlorophenol

R.T.: 14.644 min
Delta R.T.: -0.001 min
Response: 3899269
Conc: 7.58 ng/ml




#11 2,4,5-TP (SILVEX)

R.T.: 15.139 min
Delta R.T.: 0.064 min
Response: 789705
Conc: 3.33 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 15.199 min
Delta R.T.: -0.017 min
Response: 3624966
Conc: 17.82 ng/ml



#12 2,4,5-T


R.T.: 15.443 min
Delta R.T.: -0.022 min
Response: 8059103
Conc: 38.17 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-24




#12 2,4,5-T

R.T.: 15.852 min
Delta R.T.: -0.013 min
Response: 2682991
Conc: 16.20 ng/ml




#13 2,4-DB

R.T.: 16.243 min
Delta R.T.: 0.020 min
Response: 22963584
Conc: 1202.14 ng/ml



#13 2,4-DB


R.T.: 16.438 min
Delta R.T.: -0.022 min
Response: 6598350
Conc: 292.91 ng/ml



#14 DINOSEB


R.T.: 17.696 min
Delta R.T.: 0.010 min
Response: 2016948
Conc: 13.85 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-24




#14 DINOSEB

R.T.: 16.816 min
Delta R.T.: -0.034 min
Response: 3006293
Conc: 22.94 ng/ml




#15 Picloram

R.T.: 17.484 min
Delta R.T.: 0.008 min
Response: 7336875
Conc: 26.39 ng/ml



#15 Picloram


R.T.: 18.228 min
Delta R.T.: -0.017 min
Response: 8710015
Conc: 38.71 ng/ml



#16 DCPA

R.T.: 0.000 min
Exp R.T.: 18.069 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
TP-24



#16 DCPA

R.T.: 18.150 min
Delta R.T.: 0.031 min
Response: 11218369
Conc: 48.27 ng/ml