

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031419\
 Data File : PP029307.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 14 Mar 2019 18:05
 Operator : EI\MA
 Sample : K1824-14
 Misc : hexane
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 P-1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 15 02:33:12 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022119.M
 Quant Title : 8080.M
 QLast Update : Fri Feb 22 02:24:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.930	12.125	20907122	18203250	695.597	479.632 #
Target Compounds							
1) T	Dalapon	0.000	2.977	0	1433082	N.D.	27.693 #
3) T	4-Nitroph...	0.000	11.339	0	5110337	N.D.	213.395 #
8) T	DICHLORPROP	0.000	13.552	0	420593	N.D.	12.440 #
9) T	2,4-D	13.815	14.032	960797	2348532	24.506	51.794 #
10) T	Pentachlo...	14.162	14.601	861770	2379838	1.322	4.624 #
11) T	2,4,5-TP ...	0.000	15.200	0	4774531	N.D.	23.473 #
12) T	2,4,5-T	15.440	15.844	7594245	12939250	35.972	78.148 #
13) T	2,4-DB	16.237	16.456	24051473	9902714	1259.086	439.601 #
14) T	DINOSEB	17.709	16.867	2360180	6704087	16.207	51.148 #
15) T	Picloram	17.476	18.224	5869746	20610834	21.110	91.594 #
16) T	DCPA	0.000	18.105	0	6968244	N.D.	29.984 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031419\
Data File : PP029307.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 14 Mar 2019 18:05
Operator : EI\MA
Sample : K1824-14
Misc : hexane
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
P-1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 15 02:33:12 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022119.M
Quant Title : 8080.M
QLast Update : Fri Feb 22 02:24:04 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 0.000 min
Exp R.T.: 3.246 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
P-1




#1 Dalapon

R.T.: 2.977 min
Delta R.T.: 0.019 min
Response: 1433082
Conc: 27.69 ng/ml




#3 4-Nitrophenol

R.T.: 0.000 min
Exp R.T.: 11.504 min
Response: 0
Conc: N.D.



#3 4-Nitrophenol


R.T.: 11.339 min
Delta R.T.: 0.066 min
Response: 5110337
Conc: 213.39 ng/ml



#4 2,4-DCAA


R.T.: 11.930 min
Delta R.T.: -0.011 min
Response: 20907122
Conc: 695.60 ng/ml

Instrument :
ECD_P
ClientSampleId :
P-1




#4 2,4-DCAA

R.T.: 12.125 min
Delta R.T.: -0.004 min
Response: 18203250
Conc: 479.63 ng/ml




#8 DICHLORPROP

R.T.: 0.000 min
Exp R.T. : 13.452 min
Response: 0
Conc: N.D.



#8 DICHLORPROP


R.T.: 13.552 min
Delta R.T.: 0.000 min
Response: 420593
Conc: 12.44 ng/ml



#9 2,4-D


R.T.: 13.815 min
Delta R.T.: 0.015 min
Response: 960797
Conc: 24.51 ng/ml

Instrument :
ECD_P
ClientSampleId :
P-1




#9 2,4-D

R.T.: 14.032 min
Delta R.T.: -0.007 min
Response: 2348532
Conc: 51.79 ng/ml




#10 Pentachlorophenol

R.T.: 14.162 min
Delta R.T.: -0.054 min
Response: 861770
Conc: 1.32 ng/ml



#10 Pentachlorophenol


R.T.: 14.601 min
Delta R.T.: -0.044 min
Response: 2379838
Conc: 4.62 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 0.000 min
Exp R.T.: 15.075 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
P-1




#11 2,4,5-TP (SILVEX)

R.T.: 15.200 min
Delta R.T.: -0.016 min
Response: 4774531
Conc: 23.47 ng/ml




#12 2,4,5-T

R.T.: 15.440 min
Delta R.T.: -0.025 min
Response: 7594245
Conc: 35.97 ng/ml



#12 2,4,5-T


R.T.: 15.844 min
Delta R.T.: -0.021 min
Response: 12939250
Conc: 78.15 ng/ml



#13 2,4-DB


R.T.: 16.237 min
Delta R.T.: 0.014 min
Response: 24051473
Conc: 1259.09 ng/ml

Instrument :
ECD_P
ClientSampleId :
P-1




#13 2,4-DB

R.T.: 16.456 min
Delta R.T.: -0.004 min
Response: 9902714
Conc: 439.60 ng/ml




#14 DINOSEB

R.T.: 17.709 min
Delta R.T.: 0.024 min
Response: 2360180
Conc: 16.21 ng/ml



#14 DINOSEB


R.T.: 16.867 min
Delta R.T.: 0.018 min
Response: 6704087
Conc: 51.15 ng/ml



#15 Picloram


R.T.: 17.476 min
Delta R.T.: 0.000 min
Response: 5869746
Conc: 21.11 ng/ml

Instrument :
ECD_P
ClientSampleId :
P-1




#15 Picloram

R.T.: 18.224 min
Delta R.T.: -0.020 min
Response: 20610834
Conc: 91.59 ng/ml



#16 DCPA

R.T.: 0.000 min
Exp R.T. : 18.069 min
Response: 0
Conc: N.D.



#16 DCPA

R.T.: 18.105 min
Delta R.T.: -0.013 min
Response: 6968244
Conc: 29.98 ng/ml