

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031423\
 Data File : PP056499.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Mar 2023 13:31
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 14 15:19:39 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 08 02:03:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.317	3.557	88343023	70319091	56.895	49.834
2) SA Decachlor...	10.064	8.556	92634525	55285529	46.635	49.457
Target Compounds						
16) L4 AR-1242-1	5.487	4.637	25028816	16888350	513.579	477.571
17) L4 AR-1242-2	5.510	4.655	36253816	23890856	508.122	470.214
18) L4 AR-1242-3	5.572	4.831	23236218	12931096	510.414	479.201
19) L4 AR-1242-4	5.670	4.915	18843717	13790039	527.209	470.659
20) L4 AR-1242-5	6.410	5.439	20957206	15239912	490.225	472.355



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031423\
Data File : PP056499.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Mar 2023 13:31
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 14 15:19:39 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030823.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 08 02:03:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.317 min
Delta R.T.: -0.001 min
Response: 88343023
Conc: 56.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.557 min
Delta R.T.: -0.003 min
Response: 70319091
Conc: 49.83 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.064 min
Delta R.T.: 0.001 min
Response: 92634525
Conc: 46.64 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.556 min
Delta R.T.: -0.005 min
Response: 55285529
Conc: 49.46 ng/ml



#16 AR-1242-1


R.T.: 5.487 min
Delta R.T.: -0.001 min
Response: 25028816
Conc: 513.58 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.637 min
Delta R.T.: -0.003 min
Response: 16888350
Conc: 477.57 ng/ml




#17 AR-1242-2

R.T.: 5.510 min
Delta R.T.: -0.001 min
Response: 36253816
Conc: 508.12 ng/ml



#17 AR-1242-2


R.T.: 4.655 min
Delta R.T.: -0.003 min
Response: 23890856
Conc: 470.21 ng/ml



#18 AR-1242-3


R.T.: 5.572 min
Delta R.T.: 0.000 min
Response: 23236218
Conc: 510.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.831 min
Delta R.T.: -0.003 min
Response: 12931096
Conc: 479.20 ng/ml




#19 AR-1242-4

R.T.: 5.670 min
Delta R.T.: 0.000 min
Response: 18843717
Conc: 527.21 ng/ml



#19 AR-1242-4


R.T.: 4.915 min
Delta R.T.: -0.004 min
Response: 13790039
Conc: 470.66 ng/ml



#20 AR-1242-5

R.T.: 6.410 min
Delta R.T.: -0.001 min
Response: 20957206
Conc: 490.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.439 min
Delta R.T.: -0.002 min
Response: 15239912
Conc: 472.35 ng/ml